

CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: P5279 SAS No.: P5279 SDG NO.: P5279

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL01 Date Analyzed: 12/16/2024

Lab Sample No.: AR1660CCC500 Data File : PO108540.D Time Analyzed: 09:35

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.802	4.709	4.909	470.220	500.000	-6.0
Aroclor-1016-2	4.823	4.729	4.929	471.680	500.000	-5.7
Aroclor-1016-3	4.878	4.785	4.985	468.530	500.000	-6.3
Aroclor-1016-4	5.000	4.907	5.107	477.440	500.000	-4.5
Aroclor-1016-5	5.257	5.165	5.365	473.620	500.000	-5.3
Aroclor-1260-1	6.301	6.210	6.410	432.940	500.000	-13.4
Aroclor-1260-2	6.490	6.398	6.598	446.600	500.000	-10.7
Aroclor-1260-3	6.859	6.769	6.969	434.560	500.000	-13.1
Aroclor-1260-4	7.120	7.029	7.229	424.670	500.000	-15.1
Aroclor-1260-5	7.360	7.270	7.470	436.990	500.000	-12.6
Decachlorobiphenyl	8.775	8.691	8.891	42.560	50.000	-14.9
Tetrachloro-m-xylene	3.706	3.610	3.810	50.520	50.000	1.0

CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: P5279 SAS No.: P5279 SDG NO.: P5279

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL01 Date Analyzed: 12/16/2024

Lab Sample No.: AR1660CCC500 Data File : PO108540.D Time Analyzed: 09:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.789	4.698	4.898	500.620	500.000	0.1
Aroclor-1016-2	4.808	4.718	4.918	503.680	500.000	0.7
Aroclor-1016-3	4.984	4.894	5.094	493.000	500.000	-1.4
Aroclor-1016-4	5.025	4.935	5.135	496.360	500.000	-0.7
Aroclor-1016-5	5.239	5.150	5.350	493.480	500.000	-1.3
Aroclor-1260-1	6.273	6.186	6.386	480.580	500.000	-3.9
Aroclor-1260-2	6.460	6.373	6.573	483.310	500.000	-3.3
Aroclor-1260-3	6.615	6.527	6.727	472.940	500.000	-5.4
Aroclor-1260-4	7.087	7.000	7.200	486.560	500.000	-2.7
Aroclor-1260-5	7.326	7.239	7.439	497.490	500.000	-0.5
Decachlorobiphenyl	8.723	8.641	8.841	48.130	50.000	-3.7
Tetrachloro-m-xylene	3.702	3.608	3.808	51.300	50.000	2.6