

CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: P5279 SAS No.: P5279 SDG NO.: P5279

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL02 Date Analyzed: 12/16/2024

Lab Sample No.: AR1660CCC500 Data File : PO108555.D Time Analyzed: 16:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.804	4.709	4.909	456.760	500.000	-8.6
Aroclor-1016-2	4.824	4.729	4.929	463.380	500.000	-7.3
Aroclor-1016-3	4.880	4.785	4.985	461.510	500.000	-7.7
Aroclor-1016-4	5.001	4.907	5.107	460.680	500.000	-7.9
Aroclor-1016-5	5.258	5.165	5.365	455.330	500.000	-8.9
Aroclor-1260-1	6.301	6.210	6.410	439.970	500.000	-12.0
Aroclor-1260-2	6.490	6.398	6.598	445.440	500.000	-10.9
Aroclor-1260-3	6.859	6.769	6.969	439.030	500.000	-12.2
Aroclor-1260-4	7.120	7.029	7.229	424.930	500.000	-15.0
Aroclor-1260-5	7.360	7.270	7.470	426.740	500.000	-14.7
Decachlorobiphenyl	8.776	8.691	8.891	41.110	50.000	-17.8
Tetrachloro-m-xylene	3.707	3.610	3.810	46.360	50.000	-7.3

CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: P5279 SAS No.: P5279 SDG NO.: P5279

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL02 Date Analyzed: 12/16/2024

Lab Sample No.: AR1660CCC500 Data File : PO108555.D Time Analyzed: 16:27

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.791	4.698	4.898	522.450	500.000	4.5
Aroclor-1016-2	4.811	4.718	4.918	521.590	500.000	4.3
Aroclor-1016-3	4.987	4.894	5.094	513.030	500.000	2.6
Aroclor-1016-4	5.028	4.935	5.135	512.530	500.000	2.5
Aroclor-1016-5	5.242	5.150	5.350	517.620	500.000	3.5
Aroclor-1260-1	6.276	6.186	6.386	510.870	500.000	2.2
Aroclor-1260-2	6.463	6.373	6.573	513.750	500.000	2.8
Aroclor-1260-3	6.617	6.527	6.727	507.620	500.000	1.5
Aroclor-1260-4	7.089	7.000	7.200	516.260	500.000	3.3
Aroclor-1260-5	7.328	7.239	7.439	522.020	500.000	4.4
Decachlorobiphenyl	8.725	8.641	8.841	49.500	50.000	-1.0
Tetrachloro-m-xylene	3.704	3.608	3.808	51.760	50.000	3.5