

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P5279

Client: Tully Construction Co., Inc.

Analytical Method: 8082A

DataFile : PO108549.D

Lab Sample ID:	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	
			Result	Result			Qual	RPD	Qual	Low	High	RPD
Client Sample ID: P5268-01MSD	OR-02-12122024MSD											
	AR1016	190.7	0	190	ug/kg	100		2		55	146	20
	AR1260	190.7	0	186	ug/kg	98		2		45	144	20