

TPH GC CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Chemtech Contract: TULL02
ProjectID: MTA Rockaway Park
Lab Code: CHEM Case No.: P5279 SAS No.: P5279 SDG No.: P5279
DataFile: FE051708.D Analyst Name: YP\AJ Analyst Date: 12-19-2024

Conc. (PPM)	Area Count	RF	Average RF	%D
850	101277419	119150	123220	3.303