

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P5279

Client: Tully Construction Co., Inc.

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	P5277-01MS	Client Sample ID:	COMP-1AMS					DataFile:	BF140909.D		
Naphthalene	1900	0	1800	ug/Kg	95				72	110	
Acenaphthylene	1900	0	1900	ug/Kg	100				79	118	
Acenaphthene	1900	0	1900	ug/Kg	100				70	121	
Fluorene	1900	0	1800	ug/Kg	95				68	116	
Phenanthrene	1900	0	1900	ug/Kg	100				52	128	
Anthracene	1900	0	2000	ug/Kg	105				62	124	
Fluoranthene	1900	0	1800	ug/Kg	95				44	125	
Pyrene	1900	0	2100	ug/Kg	111				26	142	
Benzo(a)anthracene	1900	0	1800	ug/Kg	95				71	114	
Chrysene	1900	0	1900	ug/Kg	100				57	121	
Benzo(b)fluoranthene	1900	0	1800	ug/Kg	95				67	121	
Benzo(k)fluoranthene	1900	0	1700	ug/Kg	89				57	134	
Benzo(a)pyrene	1900	0	2000	ug/Kg	105				70	142	
Indeno(1,2,3-cd)pyrene	1900	0	2000	ug/Kg	105				40	129	
Dibenz(a,h)anthracene	1900	0	2000	ug/Kg	105				43	123	
Benzo(g,h,i)perylene	1900	0	1800	ug/Kg	95				24	125	