

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P5279 SAS No.: P5279 SDG NO.: P5279

EPA Sample No.: SSTDCCC040 Date Analyzed: 12/17/2024

Lab File ID: BF140872.D Time Analyzed: 09:24

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	80900	6.851	293024	8.13	171346	9.89
UPPER LIMIT	161800	7.351	586048	8.628	342692	10.386
LOWER LIMIT	40450	6.351	146512	7.628	85673	9.386
EPA SAMPLE NO.						
01 ROCKAWAY-PARK	57445	6.85	220060	8.13	113565	9.88

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.