

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TULL02

Lab Code: CHEM

SAS No.: P5279

SDG NO.: P5279

Lab File ID: BF140871.D

DFTPP Injection Date: 12/17/2024

Instrument ID: BNA_F

DFTPP Injection Time: 08:58

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	33.9
68	Less than 2.0% of mass 69	0.6 (1.6) 1
69	Mass 69 relative abundance	36
70	Less than 2.0% of mass 69	0.3 (0.8) 1
127	10.0 - 80.0% of mass 198	48.4
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7
275	10.0 - 60.0% of mass 198	27.5
365	Greater than 1% of mass 198	3.5
441	Present, but less than mass 443	12.5
442	Greater than 50% of mass 198	80.7
443	15.0 - 24.0% of mass 442	15.3 (19) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF140872.D	12/17/2024	09:24
ROCKAWAY-PARK	P5279-01	BF140890.D	12/17/2024	17:22