

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: TULL02
Lab Code: CHEM Case No.: P5279 SAS No.: P5279 SDG No.: P5279
Instrument ID: MSVOA_Y Calibration Date(s): 12/17/2024 12/17/2024
Heated Purge: (Y/N) Y Calibration Time(s): 10:01 14:51
GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF010 = VY020613.D		RRF020 = VY020614.D		RRF050 = VY020615.D		
		RRF100 = VY020616.D		RRF150 = VY020617.D		RRF005 = VY020619.D		
COMPOUND	RRF010	RRF020	RRF050	RRF100	RRF150	RRF005	RRF	% RSD
Dichlorodifluoromethane	0.678	0.633	0.617	0.601	0.552	0.665	0.625	7.3
Chloromethane	0.490	0.506	0.428	0.426	0.372	0.555	0.463	14.3
Vinyl Chloride	0.443	0.469	0.371	0.440	0.374	0.487	0.431	11.3
Bromomethane	0.291	0.279	0.229	0.282	0.254	0.327	0.277	12.1
Chloroethane	0.273	0.287	0.212	0.272	0.251	0.302	0.266	11.8
Trichlorofluoromethane	0.981	0.954	0.791	0.878	0.810	0.993	0.901	9.7
1,1,2-Trichlorotrifluoroethane	0.715	0.659	0.603	0.649	0.537	0.712	0.646	10.5
1,1-Dichloroethene	0.701	0.637	0.589	0.638	0.528	0.667	0.627	9.7
Acetone	0.141	0.107	0.098	0.110		0.168	0.125	23.2
Carbon Disulfide	2.315	2.083	1.939	2.092	1.669	2.142	2.040	10.7
Methyl tert-butyl Ether	1.799	1.584	1.663	1.652	1.466	1.542	1.618	7.1
Methyl Acetate	0.474	0.351	0.350	0.391	0.292	0.378	0.373	16.1
Methylene Chloride	0.758	0.683	0.617	0.665	0.596	0.766	0.681	10.3
trans-1,2-Dichloroethene	0.771	0.703	0.710	0.692	0.636	0.707	0.703	6.2
1,1-Dichloroethane	1.401	1.307	1.323	1.300	1.173	1.333	1.306	5.7
Cyclohexane	1.388	1.247	1.198	1.160	1.056	1.443	1.249	11.6
2-Butanone	0.245	0.191	0.197	0.195	0.166	0.228	0.204	13.9
Carbon Tetrachloride	0.768	0.715	0.701	0.697	0.656	0.730	0.711	5.2
cis-1,2-Dichloroethene	0.875	0.815	0.817	0.797	0.724	0.827	0.809	6.1
Bromochloromethane	0.436	0.353	0.495	0.501	0.432	0.427	0.441	12.3
Chloroform	1.768	1.465	1.378	1.302	1.186	2.069	1.528	21.6
1,1,1-Trichloroethane	1.298	1.215	1.198	1.183	1.086	1.219	1.200	5.7
Methylcyclohexane	0.883	0.848	0.832	0.812	0.766	0.883	0.837	5.3
Benzene	2.055	1.900	1.859	1.847	1.713	1.927	1.883	5.9
1,2-Dichloroethane	0.546	0.499	0.500	0.498	0.452	0.502	0.500	6
Trichloroethene	0.501	0.462	0.458	0.453	0.422	0.481	0.463	5.8
1,2-Dichloropropane	0.473	0.450	0.434	0.440	0.401	0.442	0.440	5.3
Bromodichloromethane	0.672	0.632	0.632	0.627	0.581	0.626	0.628	4.6
4-Methyl-2-Pentanone	0.364	0.290	0.299	0.301	0.261	0.297	0.302	11.1
Toluene	1.247	1.185	1.167	1.165	1.082	1.176	1.170	4.5

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: TULL02
Lab Code: CHEM Case No.: P5279 SAS No.: P5279 SDG No.: P5279
Instrument ID: MSVOA_Y Calibration Date(s): 12/17/2024 12/17/2024
Heated Purge: (Y/N) Y Calibration Time(s): 10:01 14:51
GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:	RRF010 = VY020613.D			RRF020 = VY020614.D		RRF050 = VY020615.D		RRF100 = VY020616.D		RRF150 = VY020617.D		RRF005 = VY020619.D	
COMPOUND	RRF010	RRF020	RRF050	RRF100	RRF150	RRF005	RRF	% RSD					
t-1,3-Dichloropropene	0.638	0.572	0.602	0.611	0.556	0.534	0.586	6.5					
cis-1,3-Dichloropropene	0.742	0.694	0.706	0.714	0.646	0.661	0.694	5.1					
1,1,2-Trichloroethane	0.344	0.306	0.303	0.309	0.276	0.308	0.308	7.1					
2-Hexanone	0.225	0.183	0.199	0.202	0.175	0.180	0.194	9.5					
Dibromochloromethane	0.447	0.402	0.420	0.421	0.381	0.396	0.411	5.7					
1,2-Dibromoethane	0.319	0.281	0.288	0.290	0.261	0.277	0.286	6.7					
Tetrachloroethene	0.505	0.464	0.452	0.446	0.422	0.533	0.470	8.7					
Chlorobenzene	1.501	1.397	1.374	1.381	1.293	1.441	1.398	5					
Ethyl Benzene	2.747	2.583	2.556	2.543	2.414	2.633	2.579	4.2					
m/p-Xylenes	1.031	0.951	0.953	0.942	0.892	0.984	0.959	4.8					
o-Xylene	0.956	0.892	0.898	0.887	0.842	0.894	0.895	4.1					
Styrene	1.586	1.496	1.502	1.501	1.406	1.478	1.495	3.8					
Bromoform	0.302	0.256	0.270	0.272	0.250	0.262	0.269	6.9					
Isopropylbenzene	4.636	4.451	4.356	4.284	4.051	4.401	4.363	4.4					
1,1,2,2-Tetrachloroethane	0.804	0.702	0.710	0.710	0.635	0.689	0.708	7.7					
1,3-Dichlorobenzene	1.982	1.912	1.864	1.832	1.702	1.966	1.876	5.5					
1,4-Dichlorobenzene	1.980	1.852	1.832	1.797	1.680	2.008	1.858	6.5					
1,2-Dichlorobenzene	1.735	1.614	1.606	1.602	1.501	1.694	1.626	5					
1,2-Dibromo-3-Chloropropane	0.133	0.107	0.113	0.114	0.107	0.116	0.115	8.4					
1,2,4-Trichlorobenzene	0.975	0.927	1.014	1.030	0.990	0.964	0.983	3.8					
1,2,3-Trichlorobenzene	0.798	0.791	0.864	0.874	0.846	0.813	0.831	4.2					
1,2-Dichloroethane-d4	0.619	0.487	0.569	0.507	0.467	0.636	0.548	13					
Dibromofluoromethane	0.385	0.328	0.360	0.325	0.313	0.404	0.352	10.4					
Toluene-d8	1.314	1.127	1.271	1.143	1.100	1.445	1.233	10.9					
4-Bromofluorobenzene	0.538	0.448	0.480	0.433	0.412	0.585	0.482	13.9					

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.