



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122024\
Data File : VY020663.D
Acq On : 20 Dec 2024 08:37
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121724S.M
Title : SW846 8260
Last Update : Wed Dec 18 05:27:13 2024



AutoFind: Scans 1753, 1754, 1755; Background Corrected with Scan 1746

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.3	3225	PASS
75	95	30	60	50.6	8037	PASS
95	95	100	100	100.0	15882	PASS
96	95	5	9	6.3	996	PASS
173	174	0.00	2	1.1	146	PASS
174	95	50	100	85.2	13528	PASS
175	174	5	9	8.0	1087	PASS
176	174	95	101	96.0	12988	PASS
177	176	5	9	6.7	869	PASS