

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5279

Client: Tully Construction Co., Inc.

Analytical Method: SW8260D

Datafile : VY020667.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY1220SBSD01	Dichlorodifluoromethane	20	23.9	ug/Kg	119	1		64	136	20
	Chloromethane	20	18.3	ug/Kg	92	9		70	130	20
	Vinyl chloride	20	15.5	ug/Kg	78	1		72	129	20
	Bromomethane	20	15.1	ug/Kg	76	1		58	141	20
	Chloroethane	20	14.5	ug/Kg	73	7		69	130	20
	Trichlorofluoromethane	20	16.7	ug/Kg	84	1		69	134	20
	1,1,2-Trichlorotrifluoroethane	20	14.2	ug/Kg	71	12	*	81	123	20
	1,1-Dichloroethene	20	14.0	ug/Kg	70	15	*	79	121	20
	Acetone	100	56.1	ug/Kg	56	12	*	60	131	20
	Carbon disulfide	20	14.0	ug/Kg	70	22	*	45	154	20
	Methyl tert-butyl Ether	20	14.3	ug/Kg	72	33	*	77	129	20
	Methyl Acetate	20	11.9	ug/Kg	59	25	*	69	149	20
	Methylene Chloride	20	13.3	ug/Kg	67	28	*	56	174	20
	trans-1,2-Dichloroethene	20	14.3	ug/Kg	72	40	*	80	123	20
	1,1-Dichloroethane	20	13.9	ug/Kg	70	42	*	82	123	20
	Cyclohexane	20	21.4	ug/Kg	107	1		76	122	20
	2-Butanone	100	91.5	ug/Kg	92	0		69	131	20
	Carbon Tetrachloride	20	22.2	ug/Kg	111	1		76	129	20
	cis-1,2-Dichloroethene	20	19.0	ug/Kg	95	9		82	123	20
	Bromochloromethane	20	15.9	ug/Kg	79	7	*	80	127	20
	Chloroform	20	18.0	ug/Kg	90	6		82	125	20
	1,1,1-Trichloroethane	20	21.2	ug/Kg	106	3		80	126	20
	Methylcyclohexane	20	21.4	ug/Kg	107	3		77	123	20
	Benzene	20	22.1	ug/Kg	111	0		84	121	20
	1,2-Dichloroethane	20	22.1	ug/Kg	111	6		81	126	20
	Trichloroethene	20	22.5	ug/Kg	113	3		83	122	20
	1,2-Dichloropropane	20	22.4	ug/Kg	112	2		83	122	20
	Bromodichloromethane	20	22.0	ug/Kg	110	3		82	123	20
	4-Methyl-2-Pentanone	100	110	ug/Kg	110	15		70	135	20
	Toluene	20	22.1	ug/Kg	111	0		83	122	20
	t-1,3-Dichloropropene	20	21.6	ug/Kg	108	40	*	78	124	20
	cis-1,3-Dichloropropene	20	21.8	ug/Kg	109	2		81	122	20
	1,1,2-Trichloroethane	20	22.4	ug/Kg	112	7		82	125	20
	2-Hexanone	100	110	ug/Kg	110	11		66	138	20
	Dibromochloromethane	20	22.2	ug/Kg	111	5		79	125	20
	1,2-Dibromoethane	20	22.8	ug/Kg	114	8		80	125	20
	Tetrachloroethene	20	22.1	ug/Kg	111	1		83	125	20
	Chlorobenzene	20	22.4	ug/Kg	112	0		84	122	20
	Ethyl Benzene	20	22.1	ug/Kg	111	0		82	124	20
	m/p-Xylenes	40	44.3	ug/Kg	111	1		83	124	20
	o-Xylene	20	22.1	ug/Kg	111	0		83	123	20
	Styrene	20	22.3	ug/Kg	112	2		82	124	20
	Bromoform	20	22.6	ug/Kg	113	8		75	127	20
	Isopropylbenzene	20	21.5	ug/Kg	108	1		82	124	20
	1,1,2,2-Tetrachloroethane	20	22.1	ug/Kg	111	7		77	127	20
	1,3-Dichlorobenzene	20	21.4	ug/Kg	107	1		83	122	20
	1,4-Dichlorobenzene	20	21.5	ug/Kg	108	0		84	121	20
	1,2-Dichlorobenzene	20	21.5	ug/Kg	108	0		83	124	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5279

Client: Tully Construction Co., Inc.

Analytical Method: SW8260D

Datafile : VY020667.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY1220SBSD01	1,2-Dibromo-3-Chloropropane	20	21.2	ug/Kg	106	49	*	66	134	20
	1,2,4-Trichlorobenzene	20	20.6	ug/Kg	103	4		78	127	20
	1,2,3-Trichlorobenzene	20	20.8	ug/Kg	104	2		70	137	20