

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TULL02
Lab Code: CHEM Case No.: P5279 SAS No.: P5279 SDG NO.: P5279
Lab File ID: VY020664.D Date Analyzed: 12/20/2024
Instrument ID: MSVOA_Y Time Analyzed: 09:09
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	162547	7.71	240673	8.62	212145	11.42
UPPER LIMIT	325094	8.213	481346	9.116	424290	11.92
LOWER LIMIT	81273.5	7.213	120337	8.116	106073	10.92
EPA SAMPLE NO.						
ROCKAWAY-PARK	177087	7.71	313093	8.62	249829	11.42
VY1220SBL01	159056	7.71	246351	8.62	208495	11.41
VY1220SBS01	157442	7.71	237485	8.62	207752	11.42
VY1220SBSD01	154597	7.71	238791	8.62	207851	11.41

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.