

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: P5280 SAS No.: P5280 SDG No.: P5280

Instrument ID: MSVOA_X Calibration Date/Time: 12/20/2024 19:49

Lab File ID: VX044464.D Init. Calib. Date(s): 12/11/2024 12/11/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:41 13:00

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.773	0.744		-3.75	50
Chloromethane	0.739	0.696	0.1	-5.82	50
Vinyl Chloride	0.770	0.730		-5.2	50
Bromomethane	0.546	0.498		-8.79	50
Chloroethane	0.480	0.507		5.63	50
Trichlorofluoromethane	1.508	1.202		-20.29	50
1,1,2-Trichlorotrifluoroethane	0.610	0.590		-3.28	50
1,1-Dichloroethene	0.560	0.560		0	50
Acetone	0.357	0.270		-24.37	50
Carbon Disulfide	1.065	1.089		2.25	50
Methyl tert-butyl Ether	2.187	1.972		-9.83	50
Methyl Acetate	0.922	0.860		-6.72	50
Methylene Chloride	0.671	0.636		-5.22	50
trans-1,2-Dichloroethene	0.596	0.578		-3.02	50
1,1-Dichloroethane	1.172	1.125	0.1	-4.01	50
Cyclohexane	0.946	0.871		-7.93	50
2-Butanone	0.508	0.433		-14.76	50
Carbon Tetrachloride	0.482	0.495		2.7	50
cis-1,2-Dichloroethene	0.756	0.719		-4.89	50
Bromochloromethane	0.575	0.510		-11.3	50
Chloroform	1.316	1.190		-9.57	50
1,1,1-Trichloroethane	1.093	1.029		-5.86	50
Methylcyclohexane	0.535	0.550		2.8	50
Benzene	1.359	1.393		2.5	50
1,2-Dichloroethane	0.560	0.534		-4.64	50
Trichloroethene	0.339	0.360		6.2	50
1,2-Dichloropropane	0.347	0.345		-0.58	50
Bromodichloromethane	0.465	0.491		5.59	50
4-Methyl-2-Pentanone	0.556	0.524		-5.76	50
Toluene	0.853	0.855		0.23	50
t-1,3-Dichloropropene	0.455	0.482		5.93	50
cis-1,3-Dichloropropene	0.507	0.531		4.73	50
1,1,2-Trichloroethane	0.339	0.345		1.77	50
2-Hexanone	0.403	0.381		-5.46	50
Dibromochloromethane	0.326	0.361		10.74	50
1,2-Dibromoethane	0.345	0.366		6.09	50
Tetrachloroethene	0.332	0.343		3.31	50
Chlorobenzene	1.071	1.094	0.3	2.15	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: P5280 SAS No.: P5280 SDG No.: P5280

Instrument ID: MSVOA_X Calibration Date/Time: 12/20/2024 19:49

Lab File ID: VX044464.D Init. Calib. Date(s): 12/11/2024 12/11/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:41 13:00

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.851	1.868		0.92	50
m/p-Xylenes	0.679	0.713		5.01	50
o-Xylene	0.687	0.708		3.06	50
Styrene	1.102	1.163		5.53	50
Bromoform	0.219	0.251	0.1	14.61	50
Isopropylbenzene	3.828	3.829		0.03	50
1,1,2,2-Tetrachloroethane	1.329	1.278	0.3	-3.84	50
1,3-Dichlorobenzene	1.621	1.669		2.96	50
1,4-Dichlorobenzene	1.667	1.653		-0.84	50
1,2-Dichlorobenzene	1.684	1.709		1.49	50
1,2-Dibromo-3-Chloropropane	0.258	0.238		-7.75	50
1,2,4-Trichlorobenzene	0.923	1.000		8.34	50
1,2,3-Trichlorobenzene	0.979	1.036		5.82	50
1,2-Dichloroethane-d4	0.877	0.778		-11.29	50
Dibromofluoromethane	0.346	0.363		4.91	50
Toluene-d8	1.168	1.213		3.85	50
4-Bromofluorobenzene	0.408	0.418		2.45	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.