

LAB CHRONICLE

OrderID:	P5280	OrderDate:	12/13/2024 1:04:00 PM
Client:	AECOM Technical Services, Inc.	Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258
Contact:	Eleanor Vivadou	Location:	L51,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
P5280-01	BPOW6-7-20241212	Water	VOCMS Group1	8260-Low	12/12/24		12/16/24	12/13/24
P5280-02	BPOW6-8-20241212	Water	VOCMS Group1	8260-Low	12/12/24		12/18/24	12/13/24
P5280-03	BPOW6-11-20241212	Water	VOCMS Group1	8260-Low	12/12/24		12/20/24	12/13/24
P5280-06	BPOW6-9-20241212	Water	VOCMS Group1	8260-Low	12/12/24		12/18/24	12/13/24
P5280-07	BPOW6-10-20241212	Water	VOCMS Group1	8260-Low	12/12/24		12/18/24	12/13/24
P5280-08	DUP-03-20241212	Water	VOCMS Group1	8260-Low	12/12/24		12/18/24	12/13/24

Hit Summary Sheet
SW-846

SDG No.: P5280

Client: AECOM Technical Services, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID: P5280-01	BPOW6-7-20241212 BPOW6-7-2024121	Water	Trichloroethene	3.80		0.32	0.75	1.00	ug/L
			Total Voc :	3.80					
			Total Concentration:	3.80					
Client ID: P5280-02	BPOW6-8-20241212 BPOW6-8-2024121	Water	Acetone	1.80	J	1.40	3.80	5.00	ug/L
			Total Voc :	1.80					
			Total Concentration:	1.80					
Client ID: P5280-03	BPOW6-11-20241212 BPOW6-11-202412	Water	Acetone	1.60	J	1.40	3.80	5.00	ug/L
P5280-03	BPOW6-11-202412	Water	Methyl Acetate	2.00		0.60	0.75	1.00	ug/L
			Total Voc :	3.60					
			Total Concentration:	3.60					
Client ID: P5280-07	BPOW6-10-20241212 BPOW6-10-202412	Water	Chloromethane	0.63	J	0.35	0.50	1.00	ug/L
P5280-07	BPOW6-10-202412	Water	Acetone	2.00	J	1.40	3.80	5.00	ug/L
			Total Voc :	2.63					
			Total Concentration:	2.63					



SAMPLE DATA

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/12/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	BPOW6-7-20241212		SDG No.:	P5280	
Lab Sample ID:	P5280-01		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044332.D	1		12/16/24 18:26	VX121624

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	3.80		0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.			Date Collected:	12/12/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258			Date Received:	12/13/24	
Client Sample ID:	BPOW6-7-20241212			SDG No.:	P5280	
Lab Sample ID:	P5280-01			Matrix:	Water	
Analytical Method:	SW8260			% Solid:	0	
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID :	0.18	Level :	LOW	
Prep Method :						

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044332.D	1		12/16/24 18:26	VX121624

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	46.5		81 - 118		93%	SPK: 50
1868-53-7	Dibromofluoromethane	47.2		80 - 119		94%	SPK: 50
2037-26-5	Toluene-d8	51.9		89 - 112		104%	SPK: 50
460-00-4	4-Bromofluorobenzene	47.4		85 - 114		95%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	113000	5.55				
540-36-3	1,4-Difluorobenzene	211000	6.757				
3114-55-4	Chlorobenzene-d5	182000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	72000	12.024				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/12/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	BPOW6-7-20241212		SDG No.:	P5280	
Lab Sample ID:	P5280-01		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044332.D	1		12/16/24 18:26	VX121624

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/12/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	BPOW6-8-20241212		SDG No.:	P5280	
Lab Sample ID:	P5280-02		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044384.D	1		12/18/24 15:34	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	1.80	J	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.			Date Collected:	12/12/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258			Date Received:	12/13/24	
Client Sample ID:	BPOW6-8-20241212			SDG No.:	P5280	
Lab Sample ID:	P5280-02			Matrix:	Water	
Analytical Method:	SW8260			% Solid:	0	
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID :	0.18	Level :	LOW	
Prep Method :						

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044384.D	1		12/18/24 15:34	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	44.0		81 - 118		88%	SPK: 50
1868-53-7	Dibromofluoromethane	46.5		80 - 119		93%	SPK: 50
2037-26-5	Toluene-d8	51.9		89 - 112		104%	SPK: 50
460-00-4	4-Bromofluorobenzene	47.0		85 - 114		94%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	119000	5.55				
540-36-3	1,4-Difluorobenzene	214000	6.757				
3114-55-4	Chlorobenzene-d5	187000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	77400	12.024				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/12/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	BPOW6-8-20241212		SDG No.:	P5280	
Lab Sample ID:	P5280-02		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044384.D	1		12/18/24 15:34	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/12/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	BPOW6-11-20241212		SDG No.:	P5280	
Lab Sample ID:	P5280-03		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044458.D	1		12/20/24 17:30	VX122024

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	1.60	J	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	2.00		0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/12/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	BPOW6-11-20241212		SDG No.:	P5280	
Lab Sample ID:	P5280-03		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044458.D	1		12/20/24 17:30	VX122024

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	42.6		81 - 118		85%	SPK: 50
1868-53-7	Dibromofluoromethane	48.5		80 - 119		97%	SPK: 50
2037-26-5	Toluene-d8	51.3		89 - 112		103%	SPK: 50
460-00-4	4-Bromofluorobenzene	44.2		85 - 114		88%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	114000	5.544				
540-36-3	1,4-Difluorobenzene	195000	6.757				
3114-55-4	Chlorobenzene-d5	167000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	65400	12.024				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/12/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	BPOW6-11-20241212		SDG No.:	P5280	
Lab Sample ID:	P5280-03		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044458.D	1		12/20/24 17:30	VX122024

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/12/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	BPOW6-9-20241212		SDG No.:	P5280	
Lab Sample ID:	P5280-06		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044385.D	1		12/18/24 15:58	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.			Date Collected:	12/12/24		
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258			Date Received:	12/13/24		
Client Sample ID:	BPOW6-9-20241212			SDG No.:	P5280		
Lab Sample ID:	P5280-06			Matrix:	Water		
Analytical Method:	SW8260			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	DB-624UI	ID :	0.18	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044385.D	1		12/18/24 15:58	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	44.8		81 - 118		90%	SPK: 50
1868-53-7	Dibromofluoromethane	48.0		80 - 119		96%	SPK: 50
2037-26-5	Toluene-d8	50.8		89 - 112		102%	SPK: 50
460-00-4	4-Bromofluorobenzene	47.2		85 - 114		94%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	124000	5.55				
540-36-3	1,4-Difluorobenzene	223000	6.757				
3114-55-4	Chlorobenzene-d5	195000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	82100	12.024				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/12/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	BPOW6-9-20241212		SDG No.:	P5280	
Lab Sample ID:	P5280-06		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044385.D	1		12/18/24 15:58	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.			Date Collected:	12/12/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258			Date Received:	12/13/24	
Client Sample ID:	BPOW6-10-20241212			SDG No.:	P5280	
Lab Sample ID:	P5280-07			Matrix:	Water	
Analytical Method:	SW8260			% Solid:	0	
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID :	0.18	Level :	LOW	
Prep Method :						

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044386.D	1		12/18/24 16:21	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	0.63	J	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	2.00	J	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/12/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	BPOW6-10-20241212		SDG No.:	P5280	
Lab Sample ID:	P5280-07		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044386.D	1		12/18/24 16:21	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	44.1		81 - 118		88%	SPK: 50
1868-53-7	Dibromofluoromethane	47.6		80 - 119		95%	SPK: 50
2037-26-5	Toluene-d8	51.0		89 - 112		102%	SPK: 50
460-00-4	4-Bromofluorobenzene	46.6		85 - 114		93%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	120000	5.549				
540-36-3	1,4-Difluorobenzene	215000	6.757				
3114-55-4	Chlorobenzene-d5	188000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	79300	12.024				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/12/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	BPOW6-10-20241212		SDG No.:	P5280	
Lab Sample ID:	P5280-07		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044386.D	1		12/18/24 16:21	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/12/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	DUP-03-20241212		SDG No.:	P5280	
Lab Sample ID:	P5280-08		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044387.D	1		12/18/24 16:45	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.			Date Collected:	12/12/24		
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258			Date Received:	12/13/24		
Client Sample ID:	DUP-03-20241212			SDG No.:	P5280		
Lab Sample ID:	P5280-08			Matrix:	Water		
Analytical Method:	SW8260			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	DB-624UI	ID :	0.18	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044387.D	1		12/18/24 16:45	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	45.1		81 - 118		90%	SPK: 50
1868-53-7	Dibromofluoromethane	46.9		80 - 119		94%	SPK: 50
2037-26-5	Toluene-d8	51.5		89 - 112		103%	SPK: 50
460-00-4	4-Bromofluorobenzene	47.0		85 - 114		94%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	111000	5.544				
540-36-3	1,4-Difluorobenzene	204000	6.757				
3114-55-4	Chlorobenzene-d5	177000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	70000	12.024				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/12/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	DUP-03-20241212		SDG No.:	P5280	
Lab Sample ID:	P5280-08		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044387.D	1		12/18/24 16:45	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SDG No.: **P5280**

Client: **AECOM Technical Services, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P5280-01	BPOW6-7-20241212	1,2-Dichloroethane-d4	50	46.5	93	81	118
		Dibromofluoromethane	50	47.2	94	80	119
		Toluene-d8	50	52.0	104	89	112
		4-Bromofluorobenzene	50	47.4	95	85	114
P5280-02	BPOW6-8-20241212	1,2-Dichloroethane-d4	50	44.0	88	81	118
		Dibromofluoromethane	50	46.5	93	80	119
		Toluene-d8	50	51.9	104	89	112
		4-Bromofluorobenzene	50	47.0	94	85	114
P5280-03	BPOW6-11-20241212	1,2-Dichloroethane-d4	50	42.6	85	81	118
		Dibromofluoromethane	50	48.5	97	80	119
		Toluene-d8	50	51.3	103	89	112
		4-Bromofluorobenzene	50	44.3	88	85	114
P5280-04MS	BPOW6-11-20241212MS	1,2-Dichloroethane-d4	50	44.6	89	81	118
		Dibromofluoromethane	50	51.9	104	80	119
		Toluene-d8	50	52.9	106	89	112
		4-Bromofluorobenzene	50	52.8	106	85	114
P5280-05MSD	BPOW6-11-20241212MSD	1,2-Dichloroethane-d4	50	43.9	88	81	118
		Dibromofluoromethane	50	50.3	101	80	119
		Toluene-d8	50	50.8	102	89	112
		4-Bromofluorobenzene	50	50.1	100	85	114
P5280-06	BPOW6-9-20241212	1,2-Dichloroethane-d4	50	44.8	90	81	118
		Dibromofluoromethane	50	48.0	96	80	119
		Toluene-d8	50	50.8	102	89	112
		4-Bromofluorobenzene	50	47.2	94	85	114
P5280-07	BPOW6-10-20241212	1,2-Dichloroethane-d4	50	44.1	88	81	118
		Dibromofluoromethane	50	47.6	95	80	119
		Toluene-d8	50	51.0	102	89	112
		4-Bromofluorobenzene	50	46.6	93	85	114
P5280-08	DUP-03-20241212	1,2-Dichloroethane-d4	50	45.1	90	81	118
		Dibromofluoromethane	50	46.9	94	80	119
		Toluene-d8	50	51.5	103	89	112
		4-Bromofluorobenzene	50	47.0	94	85	114
VX1216WBL01	VX1216WBL01	1,2-Dichloroethane-d4	50	47.5	95	81	118
		Dibromofluoromethane	50	46.8	94	80	119
		Toluene-d8	50	51.5	103	89	112
		4-Bromofluorobenzene	50	49.4	99	85	114
VX1216WBS01	VX1216WBS01	1,2-Dichloroethane-d4	50	47.2	94	81	118
		Dibromofluoromethane	50	48.5	97	80	119
		Toluene-d8	50	50.9	102	89	112
		4-Bromofluorobenzene	50	48.6	97	85	114
VX1218WBL01	VX1218WBL01	1,2-Dichloroethane-d4	50	45.6	91	81	118
		Dibromofluoromethane	50	47.1	94	80	119
		Toluene-d8	50	51.2	102	89	112
		4-Bromofluorobenzene	50	47.2	94	85	114
VX1218WBS01	VX1218WBS01	1,2-Dichloroethane-d4	50	47.2	94	81	118
		Dibromofluoromethane	50	50.5	101	80	119
		Toluene-d8	50	50.8	102	89	112
		4-Bromofluorobenzene	50	48.2	96	85	114
VX1220WBL01	VX1220WBL01	1,2-Dichloroethane-d4	50	44.1	88	81	118
		Dibromofluoromethane	50	47.5	95	80	119

Surrogate Summary

SDG No.: P5280

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
VX1220WBL01	VX1220WBL01	Toluene-d8	50	52.0	104	89	112
		4-Bromofluorobenzene	50	48.4	97	85	114
VX1220WBS01	VX1220WBS01	1,2-Dichloroethane-d4	50	42.7	85	81	118
		Dibromofluoromethane	50	48.5	97	80	119
		Toluene-d8	50	48.7	97	89	112
		4-Bromofluorobenzene	50	46.4	93	85	114

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P5280

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

		Sample		Rec			RPD		Limits		
Parameter	Spike	Result	Result	Units	Rec	Qual	RPD	Qual	Low	High	RPD
Lab Sample ID :	P5280-04MS	Client Sample ID :	BPOW6-11-20241212MS					Datafile :	VX044459.D		
Dichlorodifluoromethane	50	0	47.5	ug/L	95				32	152	
Chloromethane	50	0	49.4	ug/L	99				50	139	
Vinyl chloride	50	0	48.5	ug/L	97				58	137	
Bromomethane	50	0	44.4	ug/L	89				53	141	
Chloroethane	50	0	49.5	ug/L	99				60	138	
Trichlorofluoromethane	50	0	43.4	ug/L	87				65	141	
1,1,2-Trichlorotrifluoroethane	50	0	47.2	ug/L	94				70	136	
1,1-Dichloroethene	50	0	50.7	ug/L	101				71	131	
Acetone	250	1.60	210	ug/L	83				39	160	
Carbon disulfide	50	0	47.0	ug/L	94				64	133	
Methyl tert-butyl Ether	50	0	45.4	ug/L	91				71	124	
Methyl Acetate	50	2.00	44.2	ug/L	84				56	136	
Methylene Chloride	50	0	48.0	ug/L	96				74	124	
trans-1,2-Dichloroethene	50	0	48.3	ug/L	97				75	124	
1,1-Dichloroethane	50	0	47.8	ug/L	96				77	125	
Cyclohexane	50	0	46.0	ug/L	92				71	130	
2-Butanone	250	0	230	ug/L	92				56	143	
Carbon Tetrachloride	50	0	50.1	ug/L	100				72	136	
cis-1,2-Dichloroethene	50	0	48.4	ug/L	97				78	123	
Bromochloromethane	50	0	41.8	ug/L	84				78	123	
Chloroform	50	0	45.9	ug/L	92				79	124	
1,1,1-Trichloroethane	50	0	47.3	ug/L	95				74	131	
Methylcyclohexane	50	0	49.7	ug/L	99				72	132	
Benzene	50	0	52.4	ug/L	105				79	120	
1,2-Dichloroethane	50	0	48.0	ug/L	96				73	128	
Trichloroethene	50	0	52.2	ug/L	104				79	123	
1,2-Dichloropropane	50	0	49.5	ug/L	99				78	122	
Bromodichloromethane	50	0	52.7	ug/L	105				79	125	
4-Methyl-2-Pentanone	250	0	250	ug/L	100				67	130	
Toluene	50	0	50.2	ug/L	100				80	121	
t-1,3-Dichloropropene	50	0	48.0	ug/L	96				73	127	
cis-1,3-Dichloropropene	50	0	51.3	ug/L	103				75	124	
1,1,2-Trichloroethane	50	0	52.8	ug/L	106				80	119	
2-Hexanone	250	0	250	ug/L	100				57	139	
Dibromochloromethane	50	0	49.9	ug/L	100				74	126	
1,2-Dibromoethane	50	0	53.0	ug/L	106				77	121	
Tetrachloroethene	50	0	51.9	ug/L	104				74	129	
Chlorobenzene	50	0	51.4	ug/L	103				82	118	
Ethyl Benzene	50	0	50.5	ug/L	101				79	121	
m/p-Xylenes	100	0	100	ug/L	100				80	121	
o-Xylene	50	0	51.3	ug/L	103				78	122	
Styrene	50	0	54.0	ug/L	108				78	123	
Bromoform	50	0	51.8	ug/L	104				66	130	
Isopropylbenzene	50	0	49.7	ug/L	99				72	131	
1,1,2,2-Tetrachloroethane	50	0	48.9	ug/L	98				71	121	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P5280

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Parameter	Spike	Sample	Result	Units	Rec			RPD		Limits	
		Result			Rec	Qual	RPD	Qual	Low	High	RPD
1,3-Dichlorobenzene	50	0	52.3	ug/L	105					80	119
1,4-Dichlorobenzene	50	0	49.4	ug/L	99					79	118
1,2-Dichlorobenzene	50	0	50.4	ug/L	101					80	119
1,2-Dibromo-3-Chloropropane	50	0	47.1	ug/L	94					62	128
1,2,4-Trichlorobenzene	50	0	52.0	ug/L	104					69	130
1,2,3-Trichlorobenzene	50	0	52.6	ug/L	105					69	129

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P5280

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

											Limits	
Parameter	Spike	Sample	Result	Units	Rec		RPD	RPD		High	RPD	
		Result			Qual	Qual		Low				
Lab Sample ID :	P5280-05MSD	Client Sample ID :	BPOW6-11-20241212MSD					Datafile :	VX044460.D			
Dichlorodifluoromethane	50	0	48.3	ug/L	97		2		32	152	20	
Chloromethane	50	0	48.6	ug/L	97		2		50	139	20	
Vinyl chloride	50	0	48.6	ug/L	97		0		58	137	20	
Bromomethane	50	0	47.0	ug/L	94		6		53	141	20	
Chloroethane	50	0	50.3	ug/L	101		2		60	138	20	
Trichlorofluoromethane	50	0	44.6	ug/L	89		3		65	141	20	
1,1,2-Trichlorotrifluoroethane	50	0	49.6	ug/L	99		5		70	136	20	
1,1-Dichloroethene	50	0	52.4	ug/L	105		3		71	131	20	
Acetone	250	1.60	210	ug/L	83		0		39	160	20	
Carbon disulfide	50	0	50.3	ug/L	101		7		64	133	20	
Methyl tert-butyl Ether	50	0	46.2	ug/L	92		2		71	124	20	
Methyl Acetate	50	2.00	44.3	ug/L	85		1		56	136	20	
Methylene Chloride	50	0	49.9	ug/L	100		4		74	124	20	
trans-1,2-Dichloroethene	50	0	50.9	ug/L	102		5		75	124	20	
1,1-Dichloroethane	50	0	49.6	ug/L	99		4		77	125	20	
Cyclohexane	50	0	46.4	ug/L	93		1		71	130	20	
2-Butanone	250	0	230	ug/L	92		0		56	143	20	
Carbon Tetrachloride	50	0	51.2	ug/L	102		2		72	136	20	
cis-1,2-Dichloroethene	50	0	48.5	ug/L	97		0		78	123	20	
Bromochloromethane	50	0	44.8	ug/L	90		7		78	123	20	
Chloroform	50	0	47.0	ug/L	94		2		79	124	20	
1,1,1-Trichloroethane	50	0	49.1	ug/L	98		4		74	131	20	
Methylcyclohexane	50	0	51.4	ug/L	103		3		72	132	20	
Benzene	50	0	51.6	ug/L	103		2		79	120	20	
1,2-Dichloroethane	50	0	47.9	ug/L	96		0		73	128	20	
Trichloroethene	50	0	53.5	ug/L	107		2		79	123	20	
1,2-Dichloropropane	50	0	51.2	ug/L	102		3		78	122	20	
Bromodichloromethane	50	0	53.2	ug/L	106		1		79	125	20	
4-Methyl-2-Pentanone	250	0	240	ug/L	96		4		67	130	20	
Toluene	50	0	51.3	ug/L	103		2		80	121	20	
t-1,3-Dichloropropene	50	0	48.2	ug/L	96		0		73	127	20	
cis-1,3-Dichloropropene	50	0	51.7	ug/L	103		1		75	124	20	
1,1,2-Trichloroethane	50	0	52.7	ug/L	105		0		80	119	20	
2-Hexanone	250	0	250	ug/L	100		0		57	139	20	
Dibromochloromethane	50	0	50.5	ug/L	101		1		74	126	20	
1,2-Dibromoethane	50	0	52.3	ug/L	105		1		77	121	20	
Tetrachloroethene	50	0	52.5	ug/L	105		1		74	129	20	
Chlorobenzene	50	0	51.9	ug/L	104		1		82	118	20	
Ethyl Benzene	50	0	50.7	ug/L	101		0		79	121	20	
m/p-Xylenes	100	0	110	ug/L	110		10		80	121	20	
o-Xylene	50	0	52.0	ug/L	104		1		78	122	20	
Styrene	50	0	53.9	ug/L	108		0		78	123	20	
Bromoform	50	0	52.9	ug/L	106		2		66	130	20	
Isopropylbenzene	50	0	50.9	ug/L	102		2		72	131	20	
1,1,2,2-Tetrachloroethane	50	0	49.5	ug/L	99		1		71	121	20	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P5280

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Parameter	Spike	Sample Result	Result	Units	Rec			RPD	Limits		
					Rec	Qual	RPD	Qual	Low	High	RPD
1,3-Dichlorobenzene	50	0	52.9	ug/L	106		1		80	119	20
1,4-Dichlorobenzene	50	0	50.8	ug/L	102		3		79	118	20
1,2-Dichlorobenzene	50	0	50.9	ug/L	102		1		80	119	20
1,2-Dibromo-3-Chloropropane	50	0	47.9	ug/L	96		2		62	128	20
1,2,4-Trichlorobenzene	50	0	54.3	ug/L	109		4		69	130	20
1,2,3-Trichlorobenzene	50	0	53.6	ug/L	107		2		69	129	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5280

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX044311.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VX1216WBS01	Dichlorodifluoromethane	20	19.9	ug/L	100			32	152	
	Chloromethane	20	20.8	ug/L	104			50	139	
	Vinyl chloride	20	19.7	ug/L	99			58	137	
	Bromomethane	20	19.0	ug/L	95			53	141	
	Chloroethane	20	20.3	ug/L	102			60	138	
	Trichlorofluoromethane	20	16.0	ug/L	80			65	141	
	1,1,2-Trichlorotrifluoroethane	20	20.1	ug/L	101			70	136	
	1,1-Dichloroethene	20	20.3	ug/L	102			71	131	
	Acetone	100	110	ug/L	110			39	160	
	Carbon disulfide	20	19.3	ug/L	97			64	133	
	Methyl tert-butyl Ether	20	19.6	ug/L	98			71	124	
	Methyl Acetate	20	18.9	ug/L	95			56	136	
	Methylene Chloride	20	20.6	ug/L	103			74	124	
	trans-1,2-Dichloroethene	20	20.3	ug/L	102			75	124	
	1,1-Dichloroethane	20	20.4	ug/L	102			77	125	
	Cyclohexane	20	19.4	ug/L	97			71	130	
	2-Butanone	100	98.7	ug/L	99			56	143	
	Carbon Tetrachloride	20	20.4	ug/L	102			72	136	
	cis-1,2-Dichloroethene	20	19.3	ug/L	97			78	123	
	Bromochloromethane	20	16.6	ug/L	83			78	123	
	Chloroform	20	19.6	ug/L	98			79	124	
	1,1,1-Trichloroethane	20	19.8	ug/L	99			74	131	
	Methylcyclohexane	20	20.6	ug/L	103			72	132	
	Benzene	20	20.5	ug/L	103			79	120	
	1,2-Dichloroethane	20	20.0	ug/L	100			73	128	
	Trichloroethene	20	20.4	ug/L	102			79	123	
	1,2-Dichloropropane	20	20.8	ug/L	104			78	122	
	Bromodichloromethane	20	20.8	ug/L	104			79	125	
	4-Methyl-2-Pentanone	100	93.9	ug/L	94			67	130	
	Toluene	20	20.6	ug/L	103			80	121	
	t-1,3-Dichloropropene	20	20.9	ug/L	104			73	127	
	cis-1,3-Dichloropropene	20	21.3	ug/L	106			75	124	
	1,1,2-Trichloroethane	20	20.6	ug/L	103			80	119	
	2-Hexanone	100	98.2	ug/L	98			57	139	
	Dibromochloromethane	20	19.9	ug/L	100			74	126	
	1,2-Dibromoethane	20	20.8	ug/L	104			77	121	
	Tetrachloroethene	20	21.1	ug/L	106			74	129	
	Chlorobenzene	20	20.4	ug/L	102			82	118	
	Ethyl Benzene	20	20.7	ug/L	104			79	121	
	m/p-Xylenes	40	41.6	ug/L	104			80	121	
	o-Xylene	20	20.9	ug/L	104			78	122	
	Styrene	20	20.7	ug/L	104			78	123	
	Bromoform	20	19.1	ug/L	96			66	130	
	Isopropylbenzene	20	21.0	ug/L	105			72	131	
	1,1,2,2-Tetrachloroethane	20	19.5	ug/L	98			71	121	
	1,3-Dichlorobenzene	20	21.1	ug/L	106			80	119	
	1,4-Dichlorobenzene	20	20.3	ug/L	102			79	118	
	1,2-Dichlorobenzene	20	20.2	ug/L	101			80	119	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5280

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX044311.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VX1216WBS01	1,2-Dibromo-3-Chloropropane	20	16.8	ug/L	84			62	128	
	1,2,4-Trichlorobenzene	20	20.3	ug/L	102			69	130	
	1,2,3-Trichlorobenzene	20	19.9	ug/L	100			69	129	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5280

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX044378.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1218WBS01	Dichlorodifluoromethane	20	19.9	ug/L	100			32	152	
	Chloromethane	20	19.9	ug/L	100			50	139	
	Vinyl chloride	20	19.5	ug/L	98			58	137	
	Bromomethane	20	17.2	ug/L	86			53	141	
	Chloroethane	20	21.1	ug/L	106			60	138	
	Trichlorofluoromethane	20	15.8	ug/L	79			65	141	
	1,1,2-Trichlorotrifluoroethane	20	18.8	ug/L	94			70	136	
	1,1-Dichloroethene	20	19.7	ug/L	99			71	131	
	Acetone	100	80.4	ug/L	80			39	160	
	Carbon disulfide	20	19.1	ug/L	96			64	133	
	Methyl tert-butyl Ether	20	18.4	ug/L	92			71	124	
	Methyl Acetate	20	18.5	ug/L	93			56	136	
	Methylene Chloride	20	19.7	ug/L	99			74	124	
	trans-1,2-Dichloroethene	20	19.2	ug/L	96			75	124	
	1,1-Dichloroethane	20	19.6	ug/L	98			77	125	
	Cyclohexane	20	19.2	ug/L	96			71	130	
	2-Butanone	100	89.3	ug/L	89			56	143	
	Carbon Tetrachloride	20	18.4	ug/L	92			72	136	
	cis-1,2-Dichloroethene	20	18.9	ug/L	95			78	123	
	Bromochloromethane	20	17.5	ug/L	88			78	123	
	Chloroform	20	18.7	ug/L	94			79	124	
	1,1,1-Trichloroethane	20	18.7	ug/L	94			74	131	
	Methylcyclohexane	20	19.4	ug/L	97			72	132	
	Benzene	20	20.1	ug/L	101			79	120	
	1,2-Dichloroethane	20	18.9	ug/L	95			73	128	
	Trichloroethene	20	19.1	ug/L	96			79	123	
	1,2-Dichloropropane	20	19.3	ug/L	97			78	122	
	Bromodichloromethane	20	18.8	ug/L	94			79	125	
	4-Methyl-2-Pentanone	100	94.8	ug/L	95			67	130	
	Toluene	20	19.4	ug/L	97			80	121	
	t-1,3-Dichloropropene	20	18.8	ug/L	94			73	127	
	cis-1,3-Dichloropropene	20	19.5	ug/L	98			75	124	
	1,1,2-Trichloroethane	20	19.3	ug/L	97			80	119	
	2-Hexanone	100	93.3	ug/L	93			57	139	
	Dibromochloromethane	20	18.3	ug/L	92			74	126	
	1,2-Dibromoethane	20	19.9	ug/L	100			77	121	
	Tetrachloroethene	20	19.6	ug/L	98			74	129	
	Chlorobenzene	20	19.5	ug/L	98			82	118	
	Ethyl Benzene	20	19.3	ug/L	97			79	121	
	m/p-Xylenes	40	39.3	ug/L	98			80	121	
	o-Xylene	20	19.2	ug/L	96			78	122	
	Styrene	20	19.4	ug/L	97			78	123	
	Bromoform	20	18.7	ug/L	94			66	130	
	Isopropylbenzene	20	20.4	ug/L	102			72	131	
	1,1,2,2-Tetrachloroethane	20	19.9	ug/L	100			71	121	
	1,3-Dichlorobenzene	20	19.8	ug/L	99			80	119	
	1,4-Dichlorobenzene	20	19.5	ug/L	98			79	118	
	1,2-Dichlorobenzene	20	19.0	ug/L	95			80	119	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5280

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX044378.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VX1218WBS01	1,2-Dibromo-3-Chloropropane	20	17.5	ug/L	88			62	128	
	1,2,4-Trichlorobenzene	20	19.2	ug/L	96			69	130	
	1,2,3-Trichlorobenzene	20	19.7	ug/L	99			69	129	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5280

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX044441.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1220WBS01	Dichlorodifluoromethane	20	18.5	ug/L	93			32	152	
	Chloromethane	20	18.7	ug/L	94			50	139	
	Vinyl chloride	20	17.9	ug/L	90			58	137	
	Bromomethane	20	18.8	ug/L	94			53	141	
	Chloroethane	20	19.8	ug/L	99			60	138	
	Trichlorofluoromethane	20	15.3	ug/L	77			65	141	
	1,1,2-Trichlorotrifluoroethane	20	18.5	ug/L	93			70	136	
	1,1-Dichloroethene	20	18.9	ug/L	95			71	131	
	Acetone	100	75.7	ug/L	76			39	160	
	Carbon disulfide	20	16.4	ug/L	82			64	133	
	Methyl tert-butyl Ether	20	17.6	ug/L	88			71	124	
	Methyl Acetate	20	17.7	ug/L	89			56	136	
	Methylene Chloride	20	19.1	ug/L	96			74	124	
	trans-1,2-Dichloroethene	20	18.7	ug/L	94			75	124	
	1,1-Dichloroethane	20	18.9	ug/L	95			77	125	
	Cyclohexane	20	17.6	ug/L	88			71	130	
	2-Butanone	100	83.3	ug/L	83			56	143	
	Carbon Tetrachloride	20	18.9	ug/L	95			72	136	
	cis-1,2-Dichloroethene	20	18.0	ug/L	90			78	123	
	Bromochloromethane	20	15.8	ug/L	79			78	123	
	Chloroform	20	17.7	ug/L	89			79	124	
	1,1,1-Trichloroethane	20	17.5	ug/L	88			74	131	
	Methylcyclohexane	20	19.3	ug/L	97			72	132	
	Benzene	20	20.4	ug/L	102			79	120	
	1,2-Dichloroethane	20	19.2	ug/L	96			73	128	
	Trichloroethene	20	19.9	ug/L	100			79	123	
	1,2-Dichloropropane	20	19.1	ug/L	96			78	122	
	Bromodichloromethane	20	19.1	ug/L	96			79	125	
	4-Methyl-2-Pentanone	100	92.4	ug/L	92			67	130	
	Toluene	20	19.5	ug/L	98			80	121	
	t-1,3-Dichloropropene	20	17.9	ug/L	90			73	127	
	cis-1,3-Dichloropropene	20	19.2	ug/L	96			75	124	
	1,1,2-Trichloroethane	20	20.8	ug/L	104			80	119	
	2-Hexanone	100	90.4	ug/L	90			57	139	
	Dibromochloromethane	20	18.8	ug/L	94			74	126	
	1,2-Dibromoethane	20	20.2	ug/L	101			77	121	
	Tetrachloroethene	20	20.8	ug/L	104			74	129	
	Chlorobenzene	20	19.9	ug/L	100			82	118	
	Ethyl Benzene	20	19.2	ug/L	96			79	121	
	m/p-Xylenes	40	39.8	ug/L	100			80	121	
	o-Xylene	20	19.8	ug/L	99			78	122	
	Styrene	20	19.8	ug/L	99			78	123	
	Bromoform	20	19.3	ug/L	97			66	130	
	Isopropylbenzene	20	19.9	ug/L	100			72	131	
	1,1,2,2-Tetrachloroethane	20	19.6	ug/L	98			71	121	
	1,3-Dichlorobenzene	20	20.1	ug/L	101			80	119	
	1,4-Dichlorobenzene	20	20.0	ug/L	100			79	118	
	1,2-Dichlorobenzene	20	19.8	ug/L	99			80	119	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5280

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX044441.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VX1220WBS01	1,2-Dibromo-3-Chloropropane	20	16.5	ug/L	83			62	128	
	1,2,4-Trichlorobenzene	20	19.5	ug/L	98			69	130	
	1,2,3-Trichlorobenzene	20	19.4	ug/L	97			69	129	

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VX1216WBL01

Lab Name: CHEMTECH

Contract: AECO15

Lab Code: CHEM Case No.: P5280

SAS No.: P5280 SDG NO.: P5280

Lab File ID: VX044310.D

Lab Sample ID: VX1216WBL01

Date Analyzed: 12/16/2024

Time Analyzed: 09:50

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX1216WBS01	VX1216WBS01	VX044311.D	12/16/2024
BPOW6-7-20241212	P5280-01	VX044332.D	12/16/2024

COMMENTS: _____

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VX1218WBL01

Lab Name: CHEMTECH

Contract: AECO15

Lab Code: CHEM Case No.: P5280

SAS No.: P5280 SDG NO.: P5280

Lab File ID: VX044377.D

Lab Sample ID: VX1218WBL01

Date Analyzed: 12/18/2024

Time Analyzed: 12:48

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX1218WBS01	VX1218WBS01	VX044378.D	12/18/2024
BPOW6-8-20241212	P5280-02	VX044384.D	12/18/2024
BPOW6-9-20241212	P5280-06	VX044385.D	12/18/2024
BPOW6-10-20241212	P5280-07	VX044386.D	12/18/2024
DUP-03-20241212	P5280-08	VX044387.D	12/18/2024

COMMENTS: _____

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VX1220WBL01

Lab Name: CHEMTECH

Contract: AECO15

Lab Code: CHEM Case No.: P5280

SAS No.: P5280 SDG NO.: P5280

Lab File ID: VX044440.D

Lab Sample ID: VX1220WBL01

Date Analyzed: 12/20/2024

Time Analyzed: 10:23

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX1220WBS01	VX1220WBS01	VX044441.D	12/20/2024
BPOW6-11-20241212	P5280-03	VX044458.D	12/20/2024
BPOW6-11-20241212MS	P5280-04MS	VX044459.D	12/20/2024
BPOW6-11-20241212MSD	P5280-05MSD	VX044460.D	12/20/2024

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: P5280 SAS No.: P5280 SDG NO.: P5280
Lab File ID: VX044218.D BFB Injection Date: 12/11/2024
Instrument ID: MSVOA_X BFB Injection Time: 10:13
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	21.7
75	30.0 - 60.0% of mass 95	56.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.7
173	Less than 2.0% of mass 174	0.9 (1.2) 1
174	50.0 - 100.0% of mass 95	71.1
175	5.0 - 9.0% of mass 174	5.2 (7.3) 1
176	95.0 - 101.0% of mass 174	68.7 (96.6) 1
177	5.0 - 9.0% of mass 176	4.6 (6.8) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC001	VSTDIC001	VX044219.D	12/11/2024	10:41
VSTDIC005	VSTDIC005	VX044220.D	12/11/2024	11:27
VSTDIC020	VSTDIC020	VX044221.D	12/11/2024	11:50
VSTDIC050	VSTDIC050	VX044222.D	12/11/2024	12:14
VSTDIC100	VSTDIC100	VX044223.D	12/11/2024	12:37
VSTDIC150	VSTDIC150	VX044224.D	12/11/2024	13:00

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: P5280 SAS No.: P5280 SDG NO.: P5280
Lab File ID: VX044307.D BFB Injection Date: 12/16/2024
Instrument ID: MSVOA_X BFB Injection Time: 08:09
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.5
75	30.0 - 60.0% of mass 95	54.9
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.8
173	Less than 2.0% of mass 174	0.6 (0.8) 1
174	50.0 - 100.0% of mass 95	69.5
175	5.0 - 9.0% of mass 174	5.4 (7.8) 1
176	95.0 - 101.0% of mass 174	66.1 (95.2) 1
177	5.0 - 9.0% of mass 176	4.4 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX044308.D	12/16/2024	08:58
VX1216WBL01	VX1216WBL01	VX044310.D	12/16/2024	09:50
VX1216WBS01	VX1216WBS01	VX044311.D	12/16/2024	10:13
BPOW6-7-20241212	P5280-01	VX044332.D	12/16/2024	18:26
VSTDCCC050EC	VSTDCCC050	VX044333.D	12/16/2024	18:49

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: P5280 SAS No.: P5280 SDG NO.: P5280
Lab File ID: VX044374.D BFB Injection Date: 12/18/2024
Instrument ID: MSVOA_X BFB Injection Time: 08:15
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	20.4
75	30.0 - 60.0% of mass 95	55.1
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.5
173	Less than 2.0% of mass 174	0.6 (0.7) 1
174	50.0 - 100.0% of mass 95	79.8
175	5.0 - 9.0% of mass 174	5.7 (7.1) 1
176	95.0 - 101.0% of mass 174	76.6 (96) 1
177	5.0 - 9.0% of mass 176	4.6 (6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX044375.D	12/18/2024	11:57
VX1218WBL01	VX1218WBL01	VX044377.D	12/18/2024	12:48
VX1218WBS01	VX1218WBS01	VX044378.D	12/18/2024	13:11
BPOW6-8-20241212	P5280-02	VX044384.D	12/18/2024	15:34
BPOW6-9-20241212	P5280-06	VX044385.D	12/18/2024	15:58
BPOW6-10-20241212	P5280-07	VX044386.D	12/18/2024	16:21
DUP-03-20241212	P5280-08	VX044387.D	12/18/2024	16:45
VSTDCCC050EC	VSTDCCC050	VX044395.D	12/18/2024	19:52

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: P5280 SAS No.: P5280 SDG NO.: P5280
Lab File ID: VX044437.D BFB Injection Date: 12/20/2024
Instrument ID: MSVOA_X BFB Injection Time: 08:49
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	20.4
75	30.0 - 60.0% of mass 95	52.1
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7.8
173	Less than 2.0% of mass 174	0.9 (1.2) 1
174	50.0 - 100.0% of mass 95	77.7
175	5.0 - 9.0% of mass 174	5.6 (7.2) 1
176	95.0 - 101.0% of mass 174	76.6 (98.5) 1
177	5.0 - 9.0% of mass 176	4.7 (6.1) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX044438.D	12/20/2024	09:16
VX1220WBL01	VX1220WBL01	VX044440.D	12/20/2024	10:23
VX1220WBS01	VX1220WBS01	VX044441.D	12/20/2024	10:46
BPOW6-11-20241212	P5280-03	VX044458.D	12/20/2024	17:30
BPOW6-11-20241212MS	P5280-04MS	VX044459.D	12/20/2024	17:53
BPOW6-11-20241212MSD	P5280-05MSD	VX044460.D	12/20/2024	18:16
VSTDCCC050EC	VSTDCCC050	VX044464.D	12/20/2024	19:49

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: P5280 SAS No.: P5280 SDG NO.: P5280
Lab File ID: VX044308.D Date Analyzed: 12/16/2024
Instrument ID: MSVOA_X Time Analyzed: 08:58
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	139378	5.54	244474	6.75	205865	10.05
UPPER LIMIT	278756	6.038	488948	7.251	411730	10.549
LOWER LIMIT	69689	5.038	122237	6.251	102933	9.549
EPA SAMPLE NO.						
BPOW6-7-20241212	113355	5.55	210532	6.76	182215	10.06
VX1216WBL01	134837	5.54	253656	6.75	222159	10.05
VX1216WBS01	134529	5.54	242066	6.76	212439	10.06

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: P5280 SAS No.: P5280 SDG NO.: P5280
Lab File ID: VX044308.D Date Analyzed: 12/16/2024
Instrument ID: MSVOA_X Time Analyzed: 08:58
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	95969	12.018				
UPPER LIMIT	191938	12.518				
LOWER LIMIT	47984.5	11.518				
EPA SAMPLE NO.						
BPOW6-7-20241212	71993	12.02				
VX1216WBL01	94759	12.02				
VX1216WBS01	96021	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: P5280 SAS No.: P5280 SDG NO.: P5280
Lab File ID: VX044375.D Date Analyzed: 12/18/2024
Instrument ID: MSVOA_X Time Analyzed: 11:57
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	145222	5.54	244425	6.75	212426	10.05
UPPER LIMIT	290444	6.044	488850	7.251	424852	10.549
LOWER LIMIT	72611	5.044	122213	6.251	106213	9.549
EPA SAMPLE NO.						
BPOW6-8-20241212	119034	5.55	214052	6.76	187198	10.06
BPOW6-9-20241212	124049	5.55	223373	6.76	194667	10.06
BPOW6-10-20241212	119959	5.55	215469	6.76	187592	10.06
DUP-03-20241212	111360	5.54	204112	6.76	177491	10.06
VX1218WBL01	123193	5.55	230004	6.76	201476	10.06
VX1218WBS01	152020	5.54	272118	6.76	233700	10.06

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: P5280 SAS No.: P5280 SDG NO.: P5280
 Lab File ID: VX044375.D Date Analyzed: 12/18/2024
 Instrument ID: MSVOA_X Time Analyzed: 11:57
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	100489	12.018				
UPPER LIMIT	200978	12.518				
LOWER LIMIT	50244.5	11.518				
EPA SAMPLE NO.						
BPOW6-8-20241212	77415	12.02				
BPOW6-9-20241212	82050	12.02				
BPOW6-10-20241212	79273	12.02				
DUP-03-20241212	69992	12.02				
VX1218WBL01	82810	12.02				
VX1218WBS01	101950	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: P5280 SAS No.: P5280 SDG NO.: P5280
Lab File ID: VX044438.D Date Analyzed: 12/20/2024
Instrument ID: MSVOA_X Time Analyzed: 09:16
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	136445	5.54	226018	6.76	197543	10.06
UPPER LIMIT	272890	6.044	452036	7.257	395086	10.555
LOWER LIMIT	68222.5	5.044	113009	6.257	98771.5	9.555
EPA SAMPLE NO.						
BPOW6-11-20241212	114054	5.54	195214	6.76	167344	10.06
BPOW6-11-20241212MS	140154	5.54	239394	6.76	209816	10.06
BPOW6-11-20241212MSD	133359	5.55	232484	6.76	203785	10.05
VX1220WBL01	120864	5.55	218417	6.76	193836	10.05
VX1220WBS01	146515	5.55	249173	6.76	212384	10.05

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: P5280 SAS No.: P5280 SDG NO.: P5280
Lab File ID: VX044438.D Date Analyzed: 12/20/2024
Instrument ID: MSVOA_X Time Analyzed: 09:16
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	95339	12.018				
UPPER LIMIT	190678	12.518				
LOWER LIMIT	47669.5	11.518				
EPA SAMPLE NO.						
BPOW6-11-20241212	65426	12.02				
BPOW6-11-20241212MS	99964	12.02				
BPOW6-11-20241212MSD	96300	12.02				
VX1220WBL01	81005	12.02				
VX1220WBS01	96885	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX1216WBL01		SDG No.:	P5280
Lab Sample ID:	VX1216WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044310.D	1		12/16/24 09:50	VX121624

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX1216WBL01		SDG No.:	P5280
Lab Sample ID:	VX1216WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044310.D	1		12/16/24 09:50	VX121624

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	47.5		81 - 118		95%	SPK: 50
1868-53-7	Dibromofluoromethane	46.8		80 - 119		94%	SPK: 50
2037-26-5	Toluene-d8	51.6		89 - 112		103%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.4		85 - 114		99%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	135000	5.544				
540-36-3	1,4-Difluorobenzene	254000	6.751				
3114-55-4	Chlorobenzene-d5	222000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	94800	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX1216WBL01		SDG No.:	P5280
Lab Sample ID:	VX1216WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044310.D	1		12/16/24 09:50	VX121624

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX1218WBL01		SDG No.:	P5280
Lab Sample ID:	VX1218WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044377.D	1		12/18/24 12:48	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX1218WBL01		SDG No.:	P5280
Lab Sample ID:	VX1218WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044377.D	1		12/18/24 12:48	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	45.6		81 - 118		91%	SPK: 50
1868-53-7	Dibromofluoromethane	47.1		80 - 119		94%	SPK: 50
2037-26-5	Toluene-d8	51.2		89 - 112		102%	SPK: 50
460-00-4	4-Bromofluorobenzene	47.2		85 - 114		94%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	123000	5.55				
540-36-3	1,4-Difluorobenzene	230000	6.757				
3114-55-4	Chlorobenzene-d5	201000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	82800	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX1218WBL01		SDG No.:	P5280
Lab Sample ID:	VX1218WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044377.D	1		12/18/24 12:48	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX1220WBL01		SDG No.:	P5280
Lab Sample ID:	VX1220WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044440.D	1		12/20/24 10:23	VX122024

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX1220WBL01		SDG No.:	P5280
Lab Sample ID:	VX1220WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044440.D	1		12/20/24 10:23	VX122024

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	44.1		81 - 118		88%	SPK: 50
1868-53-7	Dibromofluoromethane	47.5		80 - 119		95%	SPK: 50
2037-26-5	Toluene-d8	52.1		89 - 112		104%	SPK: 50
460-00-4	4-Bromofluorobenzene	48.4		85 - 114		97%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	121000	5.55				
540-36-3	1,4-Difluorobenzene	218000	6.757				
3114-55-4	Chlorobenzene-d5	194000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	81000	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX1220WBL01		SDG No.:	P5280
Lab Sample ID:	VX1220WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044440.D	1		12/20/24 10:23	VX122024

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX1216WBS01		SDG No.:	P5280
Lab Sample ID:	VX1216WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044311.D	1		12/16/24 10:13	VX121624

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	19.9		0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	20.8		0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	19.7		0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	19.0		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	20.3		0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	16.0		0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	20.1		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	20.3		0.26	0.75	1.00	ug/L
67-64-1	Acetone	110		1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	19.3		0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	19.6		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	18.9		0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	20.6		0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	20.3		0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	20.4		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	19.4		1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	98.7		1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	20.4		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	19.3		0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	16.6		0.18	0.50	1.00	ug/L
67-66-3	Chloroform	19.6		0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	19.8		0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	20.6		0.19	0.50	1.00	ug/L
71-43-2	Benzene	20.5		0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	20.0		0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	20.4		0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	20.8		0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	20.8		0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	93.9		0.75	2.50	5.00	ug/L
108-88-3	Toluene	20.6		0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX1216WBS01		SDG No.:	P5280
Lab Sample ID:	VX1216WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044311.D	1		12/16/24 10:13	VX121624

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	20.9		0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	21.3		0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	20.6		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	98.2		1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	19.9		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	20.8		0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	21.1		0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	20.4		0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	20.7		0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	41.6		0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	20.9		0.14	0.50	1.00	ug/L
100-42-5	Styrene	20.7		0.16	0.50	1.00	ug/L
75-25-2	Bromoform	19.1		0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	21.0		0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	19.5		0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	21.1		0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	20.3		0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	20.2		0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	16.8		0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	20.3		0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	19.9		0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	47.2		81 - 118		94%	SPK: 50
1868-53-7	Dibromofluoromethane	48.5		80 - 119		97%	SPK: 50
2037-26-5	Toluene-d8	50.9		89 - 112		102%	SPK: 50
460-00-4	4-Bromofluorobenzene	48.6		85 - 114		97%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	135000	5.543				
540-36-3	1,4-Difluorobenzene	242000	6.757				
3114-55-4	Chlorobenzene-d5	212000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	96000	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX1216WBS01		SDG No.:	P5280
Lab Sample ID:	VX1216WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044311.D	1		12/16/24 10:13	VX121624

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX1218WBS01		SDG No.:	P5280
Lab Sample ID:	VX1218WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044378.D	1		12/18/24 13:11	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	19.9		0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	19.9		0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	19.5		0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	17.2		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	21.1		0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	15.8		0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	18.8		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	19.7		0.26	0.75	1.00	ug/L
67-64-1	Acetone	80.4		1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	19.1		0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	18.4		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	18.5		0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	19.7		0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	19.2		0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	19.6		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	19.2		1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	89.3		1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	18.4		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	18.9		0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	17.5		0.18	0.50	1.00	ug/L
67-66-3	Chloroform	18.7		0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	18.7		0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	19.4		0.19	0.50	1.00	ug/L
71-43-2	Benzene	20.1		0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	18.9		0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	19.1		0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	19.3		0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	18.8		0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	94.8		0.75	2.50	5.00	ug/L
108-88-3	Toluene	19.4		0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX1218WBS01		SDG No.:	P5280
Lab Sample ID:	VX1218WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044378.D	1		12/18/24 13:11	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	18.8		0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	19.5		0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	19.3		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	93.3		1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	18.3		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	19.9		0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	19.6		0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	19.5		0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	19.3		0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	39.3		0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	19.2		0.14	0.50	1.00	ug/L
100-42-5	Styrene	19.4		0.16	0.50	1.00	ug/L
75-25-2	Bromoform	18.7		0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	20.4		0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	19.9		0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	19.8		0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	19.5		0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	19.0		0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	17.5		0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	19.2		0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	19.7		0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	47.2		81 - 118		94%	SPK: 50
1868-53-7	Dibromofluoromethane	50.5		80 - 119		101%	SPK: 50
2037-26-5	Toluene-d8	50.8		89 - 112		102%	SPK: 50
460-00-4	4-Bromofluorobenzene	48.2		85 - 114		96%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	152000	5.544				
540-36-3	1,4-Difluorobenzene	272000	6.757				
3114-55-4	Chlorobenzene-d5	234000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	102000	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX1218WBS01		SDG No.:	P5280
Lab Sample ID:	VX1218WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044378.D	1		12/18/24 13:11	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX1220WBS01		SDG No.:	P5280
Lab Sample ID:	VX1220WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044441.D	1		12/20/24 10:46	VX122024

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	18.5		0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	18.7		0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	17.9		0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	18.8		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	19.8		0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	15.3		0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	18.5		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	18.9		0.26	0.75	1.00	ug/L
67-64-1	Acetone	75.7		1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	16.4		0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	17.6		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	17.7		0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	19.1		0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	18.7		0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	18.9		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	17.6		1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	83.3		1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	18.9		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	18.0		0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	15.8		0.18	0.50	1.00	ug/L
67-66-3	Chloroform	17.7		0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	17.5		0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	19.3		0.19	0.50	1.00	ug/L
71-43-2	Benzene	20.4		0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	19.2		0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	19.9		0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	19.1		0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	19.1		0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	92.4		0.75	2.50	5.00	ug/L
108-88-3	Toluene	19.5		0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX1220WBS01		SDG No.:	P5280
Lab Sample ID:	VX1220WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044441.D	1		12/20/24 10:46	VX122024

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	17.9		0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	19.2		0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	20.8		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	90.4		1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	18.8		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	20.2		0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	20.8		0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	19.9		0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	19.2		0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	39.8		0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	19.8		0.14	0.50	1.00	ug/L
100-42-5	Styrene	19.8		0.16	0.50	1.00	ug/L
75-25-2	Bromoform	19.3		0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	19.9		0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	19.6		0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	20.1		0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	20.0		0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	19.8		0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	16.5		0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	19.5		0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	19.4		0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	42.7		81 - 118		85%	SPK: 50
1868-53-7	Dibromofluoromethane	48.5		80 - 119		97%	SPK: 50
2037-26-5	Toluene-d8	48.7		89 - 112		97%	SPK: 50
460-00-4	4-Bromofluorobenzene	46.4		85 - 114		93%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	147000	5.55				
540-36-3	1,4-Difluorobenzene	249000	6.757				
3114-55-4	Chlorobenzene-d5	212000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	96900	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX1220WBS01		SDG No.:	P5280
Lab Sample ID:	VX1220WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044441.D	1		12/20/24 10:46	VX122024

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/12/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	BPOW6-11-20241212MS		SDG No.:	P5280	
Lab Sample ID:	P5280-04MS		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044459.D	1		12/20/24 17:53	VX122024

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	47.5		0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	49.4		0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	48.5		0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	44.4		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	49.5		0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	43.4		0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	47.2		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	50.7		0.26	0.75	1.00	ug/L
67-64-1	Acetone	210		1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	47.0		0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	45.4		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	44.2		0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	48.0		0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	48.3		0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	47.8		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	46.0		1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	230		1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	50.1		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	48.4		0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	41.8		0.18	0.50	1.00	ug/L
67-66-3	Chloroform	45.9		0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	47.3		0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	49.7		0.19	0.50	1.00	ug/L
71-43-2	Benzene	52.4		0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	48.0		0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	52.2		0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	49.5		0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	52.7		0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	250		0.75	2.50	5.00	ug/L
108-88-3	Toluene	50.2		0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/12/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	BPOW6-11-20241212MS		SDG No.:	P5280	
Lab Sample ID:	P5280-04MS		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044459.D	1		12/20/24 17:53	VX122024

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	48.0		0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	51.3		0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	52.8		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	250		1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	49.9		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	53.0		0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	51.9		0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	51.4		0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	50.5		0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	100		0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	51.3		0.14	0.50	1.00	ug/L
100-42-5	Styrene	54.0		0.16	0.50	1.00	ug/L
75-25-2	Bromoform	51.8		0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	49.7		0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	48.9		0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	52.3		0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	49.4		0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	50.4		0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	47.1		0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	52.0		0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	52.6		0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	44.6		81 - 118		89%	SPK: 50
1868-53-7	Dibromofluoromethane	51.9		80 - 119		104%	SPK: 50
2037-26-5	Toluene-d8	52.9		89 - 112		106%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.8		85 - 114		106%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	140000	5.544				
540-36-3	1,4-Difluorobenzene	239000	6.757				
3114-55-4	Chlorobenzene-d5	210000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	100000	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/12/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	BPOW6-11-20241212MS		SDG No.:	P5280	
Lab Sample ID:	P5280-04MS		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044459.D	1		12/20/24 17:53	VX122024

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/12/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	BPOW6-11-20241212MSD		SDG No.:	P5280	
Lab Sample ID:	P5280-05MSD		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044460.D	1		12/20/24 18:16	VX122024

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	48.3		0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	48.6		0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	48.6		0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	47.0		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	50.3		0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	44.6		0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	49.6		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	52.4		0.26	0.75	1.00	ug/L
67-64-1	Acetone	210		1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	50.3		0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	46.2		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	44.3		0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	49.9		0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	50.9		0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	49.6		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	46.4		1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	230		1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	51.2		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	48.5		0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	44.8		0.18	0.50	1.00	ug/L
67-66-3	Chloroform	47.0		0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	49.1		0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	51.4		0.19	0.50	1.00	ug/L
71-43-2	Benzene	51.6		0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	47.9		0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	53.5		0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	51.2		0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	53.2		0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	240		0.75	2.50	5.00	ug/L
108-88-3	Toluene	51.3		0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/12/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	BPOW6-11-20241212MSD		SDG No.:	P5280	
Lab Sample ID:	P5280-05MSD		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044460.D	1		12/20/24 18:16	VX122024

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	48.2		0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	51.7		0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	52.7		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	250		1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	50.5		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	52.3		0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	52.5		0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	51.9		0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	50.7		0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	110		0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	52.0		0.14	0.50	1.00	ug/L
100-42-5	Styrene	53.9		0.16	0.50	1.00	ug/L
75-25-2	Bromoform	52.9		0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	50.9		0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	49.5		0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	52.9		0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	50.8		0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	50.9		0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	47.9		0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	54.3		0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	53.6		0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	43.9		81 - 118		88%	SPK: 50
1868-53-7	Dibromofluoromethane	50.3		80 - 119		101%	SPK: 50
2037-26-5	Toluene-d8	50.8		89 - 112		102%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.1		85 - 114		100%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	133000	5.55				
540-36-3	1,4-Difluorobenzene	232000	6.757				
3114-55-4	Chlorobenzene-d5	204000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	96300	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/12/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	BPOW6-11-20241212MSD		SDG No.:	P5280	
Lab Sample ID:	P5280-05MSD		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044460.D	1		12/20/24 18:16	VX122024

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name:	<u>CHEMTECH</u>	Contract:	<u>AECO15</u>
Lab Code:	<u>CHEM</u>	Case No.:	<u>P5280</u>
Instrument ID:	<u>MSVOA_X</u>	SAS No.:	<u>P5280</u>
Heated Purge: (Y/N)	<u>N</u>	SDG No.:	<u>P5280</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18 (mm)</u>
		Calibration Date(s):	<u>12/11/2024</u>
		Calibration Time(s):	<u>10:41</u>

LAB FILE ID:	RRF001 = VX044219.D	RRF005 = VX044220.D	RRF020 = VX044221.D	RRF050 = VX044222.D	RRF100 = VX044223.D	RRF150 = VX044224.D		
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
Dichlorodifluoromethane	0.877	0.749	0.721	0.788	0.745	0.759	0.773	7.2
Chloromethane	0.743	0.747	0.735	0.756	0.715	0.740	0.739	1.8
Vinyl Chloride	0.805	0.790	0.782	0.786	0.722	0.737	0.770	4.3
Bromomethane		0.590	0.513	0.551	0.529	0.548	0.546	5.3
Chloroethane	0.464	0.520	0.515	0.532	0.422	0.426	0.480	10.2
Trichlorofluoromethane	1.582	1.519	1.514	1.541	1.478	1.414	1.508	3.8
1,1,2-Trichlorotrifluoroethane	0.633	0.628	0.592	0.615	0.573	0.620	0.610	3.8
1,1-Dichloroethene	0.512	0.560	0.544	0.596	0.554	0.596	0.560	5.7
Acetone	0.359	0.353	0.331	0.381	0.348	0.372	0.357	5
Carbon Disulfide	0.841	0.872	0.937	1.150	1.230	1.362	1.065	20
Methyl tert-butyl Ether	2.054	2.188	2.127	2.305	2.170	2.279	2.187	4.3
Methyl Acetate	0.965	0.889	0.883	0.953	0.888	0.951	0.922	4.2
Methylene Chloride	0.679	0.678	0.666	0.689	0.648	0.668	0.671	2.1
trans-1,2-Dichloroethene	0.600	0.556	0.577	0.619	0.595	0.626	0.596	4.4
1,1-Dichloroethane	1.071	1.173	1.154	1.244	1.171	1.218	1.172	5.1
Cyclohexane		0.931	0.963	0.977	0.922	0.938	0.946	2.4
2-Butanone	0.449	0.491	0.510	0.548	0.509	0.541	0.508	7.1
Carbon Tetrachloride	0.484	0.443	0.432	0.501	0.502	0.529	0.482	7.8
cis-1,2-Dichloroethene	0.809	0.737	0.728	0.772	0.731	0.762	0.756	4.1
Bromochloromethane	0.578	0.577	0.576	0.592	0.556	0.568	0.575	2.1
Chloroform	1.389	1.312	1.272	1.343	1.272	1.310	1.316	3.4
1,1,1-Trichloroethane	0.993	1.019	1.077	1.176	1.119	1.173	1.093	7.1
Methylcyclohexane	0.514	0.503	0.527	0.555	0.550	0.562	0.535	4.5
Benzene	1.353	1.353	1.332	1.410	1.353	1.353	1.359	1.9
1,2-Dichloroethane	0.554	0.557	0.545	0.584	0.556	0.564	0.560	2.4
Trichloroethene	0.356	0.325	0.325	0.349	0.337	0.342	0.339	3.8
1,2-Dichloropropane	0.368	0.345	0.330	0.357	0.341	0.342	0.347	3.8
Bromodichloromethane	0.380	0.411	0.436	0.508	0.518	0.538	0.465	13.9
4-Methyl-2-Pentanone	0.506	0.551	0.547	0.590	0.561	0.582	0.556	5.4
Toluene	0.908	0.830	0.816	0.877	0.841	0.849	0.853	4

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: P5280 SAS No.: P5280 SDG No.: P5280

Instrument ID: MSVOA_X Calibration Date(s): 12/11/2024 12/11/2024

Heated Purge: (Y/N) N Calibration Time(s): 10:41 13:00

GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:		RRF001 = VX044219.D		RRF005 = VX044220.D		RRF020 = VX044221.D		
		RRF050 = VX044222.D		RRF100 = VX044223.D		RRF150 = VX044224.D		
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
t-1,3-Dichloropropene	0.338	0.357	0.439	0.514	0.526	0.558	0.455	20.3
cis-1,3-Dichloropropene	0.404	0.441	0.488	0.567	0.558	0.583	0.507	14.6
1,1,2-Trichloroethane	0.337	0.332	0.335	0.358	0.334	0.340	0.339	2.8
2-Hexanone	0.323	0.386	0.408	0.439	0.420	0.438	0.403	10.8
Dibromochloromethane	0.254	0.248	0.306	0.370	0.378	0.398	0.326	20.2
1,2-Dibromoethane	0.303	0.322	0.344	0.369	0.364	0.370	0.345	8.1
Tetrachloroethene	0.359	0.354	0.317	0.321	0.312	0.327	0.332	6
Chlorobenzene	1.130	1.068	1.039	1.088	1.036	1.067	1.071	3.3
Ethyl Benzene	1.853	1.806	1.816	1.927	1.831	1.875	1.851	2.4
m/p-Xylenes	0.653	0.660	0.665	0.707	0.683	0.707	0.679	3.5
o-Xylene	0.686	0.683	0.685	0.703	0.671	0.697	0.687	1.6
Styrene	1.007	1.019	1.099	1.162	1.148	1.177	1.102	6.7
Bromoform	0.137	0.173	0.194	0.247	0.269	0.297	0.219	28
Isopropylbenzene	3.665	3.976	3.920	3.912	3.747	3.747	3.828	3.2
1,1,2,2-Tetrachloroethane	1.410	1.350	1.332	1.336	1.256	1.293	1.329	3.9
1,3-Dichlorobenzene	1.519	1.634	1.640	1.683	1.625	1.628	1.621	3.3
1,4-Dichlorobenzene	1.811	1.640	1.632	1.672	1.621	1.623	1.667	4.4
1,2-Dichlorobenzene	1.745	1.666	1.668	1.726	1.654	1.643	1.684	2.5
1,2-Dibromo-3-Chloropropane	0.222	0.226	0.241	0.273	0.279	0.306	0.258	13
1,2,4-Trichlorobenzene	0.770	0.827	0.891	0.978	1.013	1.058	0.923	12.2
1,2,3-Trichlorobenzene	0.854	0.885	0.984	1.019	1.055	1.077	0.979	9.3
1,2-Dichloroethane-d4		0.934	0.778	0.921	0.862	0.888	0.877	7.1
Dibromofluoromethane		0.353	0.298	0.367	0.360	0.354	0.346	8
Toluene-d8		1.148	1.010	1.251	1.229	1.203	1.168	8.3
4-Bromofluorobenzene		0.390	0.351	0.442	0.432	0.426	0.408	9.2

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: P5280 SAS No.: P5280 SDG No.: P5280

Instrument ID: MSVOA_X Calibration Date/Time: 12/16/2024 08:58

Lab File ID: VX044308.D Init. Calib. Date(s): 12/11/2024 12/11/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:41 13:00

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.773	0.794		2.72	20
Chloromethane	0.739	0.720	0.1	-2.57	20
Vinyl Chloride	0.770	0.735		-4.55	20
Bromomethane	0.546	0.499		-8.61	20
Chloroethane	0.480	0.450		-6.25	20
Trichlorofluoromethane	1.508	1.490		-1.19	20
1,1,2-Trichlorotrifluoroethane	0.610	0.650		6.56	20
1,1-Dichloroethene	0.560	0.580		3.57	20
Acetone	0.357	0.399		11.77	20
Carbon Disulfide	1.065	1.223		14.84	20
Methyl tert-butyl Ether	2.187	2.109		-3.57	20
Methyl Acetate	0.922	0.830		-9.98	20
Methylene Chloride	0.671	0.649		-3.28	20
trans-1,2-Dichloroethene	0.596	0.599		0.5	20
1,1-Dichloroethane	1.172	1.189	0.1	1.45	20
Cyclohexane	0.946	0.966		2.11	20
2-Butanone	0.508	0.480		-5.51	20
Carbon Tetrachloride	0.482	0.526		9.13	20
cis-1,2-Dichloroethene	0.756	0.720		-4.76	20
Bromochloromethane	0.575	0.551		-4.17	20
Chloroform	1.316	1.264		-3.95	20
1,1,1-Trichloroethane	1.093	1.118		2.29	20
Methylcyclohexane	0.535	0.604		12.9	20
Benzene	1.359	1.384		1.84	20
1,2-Dichloroethane	0.560	0.564		0.71	20
Trichloroethene	0.339	0.347		2.36	20
1,2-Dichloropropane	0.347	0.340		-2.02	20
Bromodichloromethane	0.465	0.508		9.25	20
4-Methyl-2-Pentanone	0.556	0.506		-8.99	20
Toluene	0.853	0.830		-2.7	20
t-1,3-Dichloropropene	0.455	0.509		11.87	20
cis-1,3-Dichloropropene	0.507	0.550		8.48	20
1,1,2-Trichloroethane	0.339	0.332		-2.07	20
2-Hexanone	0.403	0.393		-2.48	20
Dibromochloromethane	0.326	0.357		9.51	20
1,2-Dibromoethane	0.345	0.350		1.45	20
Tetrachloroethene	0.332	0.342		3.01	20
Chlorobenzene	1.071	1.071	0.3	0	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: P5280 SAS No.: P5280 SDG No.: P5280

Instrument ID: MSVOA_X Calibration Date/Time: 12/16/2024 08:58

Lab File ID: VX044308.D Init. Calib. Date(s): 12/11/2024 12/11/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:41 13:00

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.851	1.926		4	20
m/p-Xylenes	0.679	0.710		4.57	20
o-Xylene	0.687	0.689		0.29	20
Styrene	1.102	1.176		6.72	20
Bromoform	0.219	0.257	0.1	17.35	20
Isopropylbenzene	3.828	3.954		3.29	20
1,1,2,2-Tetrachloroethane	1.329	1.243	0.3	-6.47	20
1,3-Dichlorobenzene	1.621	1.669		2.96	20
1,4-Dichlorobenzene	1.667	1.689		1.32	20
1,2-Dichlorobenzene	1.684	1.724		2.38	20
1,2-Dibromo-3-Chloropropane	0.258	0.261		1.16	20
1,2,4-Trichlorobenzene	0.923	1.012		9.64	20
1,2,3-Trichlorobenzene	0.979	1.000		2.14	20
1,2-Dichloroethane-d4	0.877	0.799		-8.89	20
Dibromofluoromethane	0.346	0.341		-1.45	20
Toluene-d8	1.168	1.142		-2.23	20
4-Bromofluorobenzene	0.408	0.395		-3.19	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: P5280 SAS No.: P5280 SDG No.: P5280

Instrument ID: MSVOA_X Calibration Date/Time: 12/16/2024 18:49

Lab File ID: VX044333.D Init. Calib. Date(s): 12/11/2024 12/11/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:41 13:00

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.773	0.718		-7.11	50
Chloromethane	0.739	0.742	0.1	0.41	50
Vinyl Chloride	0.770	0.756		-1.82	50
Bromomethane	0.546	0.496		-9.16	50
Chloroethane	0.480	0.450		-6.25	50
Trichlorofluoromethane	1.508	1.420		-5.84	50
1,1,2-Trichlorotrifluoroethane	0.610	0.588		-3.61	50
1,1-Dichloroethene	0.560	0.555		-0.89	50
Acetone	0.357	0.322		-9.8	50
Carbon Disulfide	1.065	1.006		-5.54	50
Methyl tert-butyl Ether	2.187	2.102		-3.89	50
Methyl Acetate	0.922	0.915		-0.76	50
Methylene Chloride	0.671	0.670		-0.15	50
trans-1,2-Dichloroethene	0.596	0.599		0.5	50
1,1-Dichloroethane	1.172	1.204	0.1	2.73	50
Cyclohexane	0.946	0.934		-1.27	50
2-Butanone	0.508	0.492		-3.15	50
Carbon Tetrachloride	0.482	0.488		1.25	50
cis-1,2-Dichloroethene	0.756	0.748		-1.06	50
Bromochloromethane	0.575	0.542		-5.74	50
Chloroform	1.316	1.279		-2.81	50
1,1,1-Trichloroethane	1.093	1.065		-2.56	50
Methylcyclohexane	0.535	0.553		3.36	50
Benzene	1.359	1.435		5.59	50
1,2-Dichloroethane	0.560	0.571		1.96	50
Trichloroethene	0.339	0.357		5.31	50
1,2-Dichloropropane	0.347	0.361		4.03	50
Bromodichloromethane	0.465	0.496		6.67	50
4-Methyl-2-Pentanone	0.556	0.564		1.44	50
Toluene	0.853	0.864		1.29	50
t-1,3-Dichloropropene	0.455	0.476		4.61	50
cis-1,3-Dichloropropene	0.507	0.542		6.9	50
1,1,2-Trichloroethane	0.339	0.354		4.43	50
2-Hexanone	0.403	0.415		2.98	50
Dibromochloromethane	0.326	0.345		5.83	50
1,2-Dibromoethane	0.345	0.368		6.67	50
Tetrachloroethene	0.332	0.346		4.22	50
Chlorobenzene	1.071	1.084	0.3	1.21	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: P5280 SAS No.: P5280 SDG No.: P5280

Instrument ID: MSVOA_X Calibration Date/Time: 12/16/2024 18:49

Lab File ID: VX044333.D Init. Calib. Date(s): 12/11/2024 12/11/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:41 13:00

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.851	1.921		3.78	50
m/p-Xylenes	0.679	0.723		6.48	50
o-Xylene	0.687	0.718		4.51	50
Styrene	1.102	1.187		7.71	50
Bromoform	0.219	0.231	0.1	5.48	50
Isopropylbenzene	3.828	3.922		2.46	50
1,1,2,2-Tetrachloroethane	1.329	1.334	0.3	0.38	50
1,3-Dichlorobenzene	1.621	1.674		3.27	50
1,4-Dichlorobenzene	1.667	1.704		2.22	50
1,2-Dichlorobenzene	1.684	1.701		1.01	50
1,2-Dibromo-3-Chloropropane	0.258	0.253		-1.94	50
1,2,4-Trichlorobenzene	0.923	0.964		4.44	50
1,2,3-Trichlorobenzene	0.979	1.009		3.06	50
1,2-Dichloroethane-d4	0.877	0.838		-4.45	50
Dibromofluoromethane	0.346	0.353		2.02	50
Toluene-d8	1.168	1.212		3.77	50
4-Bromofluorobenzene	0.408	0.411		0.74	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: P5280 SAS No.: P5280 SDG No.: P5280

Instrument ID: MSVOA_X Calibration Date/Time: 12/18/2024 11:57

Lab File ID: VX044375.D Init. Calib. Date(s): 12/11/2024 12/11/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:41 13:00

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.773	0.800		3.49	20
Chloromethane	0.739	0.720	0.1	-2.57	20
Vinyl Chloride	0.770	0.743		-3.51	20
Bromomethane	0.546	0.475		-13	20
Chloroethane	0.480	0.494		2.92	20
Trichlorofluoromethane	1.508	1.275		-15.45	20
1,1,2-Trichlorotrifluoroethane	0.610	0.613		0.49	20
1,1-Dichloroethene	0.560	0.566		1.07	20
Acetone	0.357	0.298		-16.53	20
Carbon Disulfide	1.065	1.209		13.52	20
Methyl tert-butyl Ether	2.187	2.034		-7	20
Methyl Acetate	0.922	0.830		-9.98	20
Methylene Chloride	0.671	0.643		-4.17	20
trans-1,2-Dichloroethene	0.596	0.594		-0.34	20
1,1-Dichloroethane	1.172	1.124	0.1	-4.1	20
Cyclohexane	0.946	0.956		1.06	20
2-Butanone	0.508	0.450		-11.42	20
Carbon Tetrachloride	0.482	0.517		7.26	20
cis-1,2-Dichloroethene	0.756	0.723		-4.36	20
Bromochloromethane	0.575	0.531		-7.65	20
Chloroform	1.316	1.196		-9.12	20
1,1,1-Trichloroethane	1.093	1.072		-1.92	20
Methylcyclohexane	0.535	0.624		16.64	20
Benzene	1.359	1.428		5.08	20
1,2-Dichloroethane	0.560	0.555		-0.89	20
Trichloroethene	0.339	0.364		7.38	20
1,2-Dichloropropane	0.347	0.356		2.59	20
Bromodichloromethane	0.465	0.508		9.25	20
4-Methyl-2-Pentanone	0.556	0.546		-1.8	20
Toluene	0.853	0.900		5.51	20
t-1,3-Dichloropropene	0.455	0.521		14.51	20
cis-1,3-Dichloropropene	0.507	0.571		12.62	20
1,1,2-Trichloroethane	0.339	0.351		3.54	20
2-Hexanone	0.403	0.402		-0.25	20
Dibromochloromethane	0.326	0.365		11.96	20
1,2-Dibromoethane	0.345	0.365		5.8	20
Tetrachloroethene	0.332	0.358		7.83	20
Chlorobenzene	1.071	1.111	0.3	3.73	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: P5280 SAS No.: P5280 SDG No.: P5280

Instrument ID: MSVOA_X Calibration Date/Time: 12/18/2024 11:57

Lab File ID: VX044375.D Init. Calib. Date(s): 12/11/2024 12/11/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:41 13:00

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.851	1.955		5.62	20
m/p-Xylenes	0.679	0.737		8.54	20
o-Xylene	0.687	0.731		6.41	20
Styrene	1.102	1.210		9.8	20
Bromoform	0.219	0.261	0.1	19.18	20
Isopropylbenzene	3.828	4.011		4.78	20
1,1,2,2-Tetrachloroethane	1.329	1.294	0.3	-2.63	20
1,3-Dichlorobenzene	1.621	1.757		8.39	20
1,4-Dichlorobenzene	1.667	1.734		4.02	20
1,2-Dichlorobenzene	1.684	1.744		3.56	20
1,2-Dibromo-3-Chloropropane	0.258	0.244		-5.43	20
1,2,4-Trichlorobenzene	0.923	1.075		16.47	20
1,2,3-Trichlorobenzene	0.979	1.107		13.07	20
1,2-Dichloroethane-d4	0.877	0.742		-15.39	20
Dibromofluoromethane	0.346	0.340		-1.73	20
Toluene-d8	1.168	1.164		-0.34	20
4-Bromofluorobenzene	0.408	0.413		1.23	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: P5280 SAS No.: P5280 SDG No.: P5280

Instrument ID: MSVOA_X Calibration Date/Time: 12/18/2024 19:52

Lab File ID: VX044395.D Init. Calib. Date(s): 12/11/2024 12/11/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:41 13:00

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.773	0.785		1.55	50
Chloromethane	0.739	0.805	0.1	8.93	50
Vinyl Chloride	0.770	0.803		4.29	50
Bromomethane	0.546	0.526		-3.66	50
Chloroethane	0.480	0.557		16.04	50
Trichlorofluoromethane	1.508	1.236		-18.04	50
1,1,2-Trichlorotrifluoroethane	0.610	0.577		-5.41	50
1,1-Dichloroethene	0.560	0.579		3.39	50
Acetone	0.357	0.306		-14.29	50
Carbon Disulfide	1.065	1.056		-0.85	50
Methyl tert-butyl Ether	2.187	2.164		-1.05	50
Methyl Acetate	0.922	0.912		-1.09	50
Methylene Chloride	0.671	0.695		3.58	50
trans-1,2-Dichloroethene	0.596	0.603		1.17	50
1,1-Dichloroethane	1.172	1.201	0.1	2.47	50
Cyclohexane	0.946	0.875		-7.51	50
2-Butanone	0.508	0.497		-2.16	50
Carbon Tetrachloride	0.482	0.466		-3.32	50
cis-1,2-Dichloroethene	0.756	0.763		0.93	50
Bromochloromethane	0.575	0.561		-2.43	50
Chloroform	1.316	1.284		-2.43	50
1,1,1-Trichloroethane	1.093	1.060		-3.02	50
Methylcyclohexane	0.535	0.528		-1.31	50
Benzene	1.359	1.397		2.8	50
1,2-Dichloroethane	0.560	0.554		-1.07	50
Trichloroethene	0.339	0.347		2.36	50
1,2-Dichloropropane	0.347	0.349		0.58	50
Bromodichloromethane	0.465	0.496		6.67	50
4-Methyl-2-Pentanone	0.556	0.558		0.36	50
Toluene	0.853	0.833		-2.35	50
t-1,3-Dichloropropene	0.455	0.470		3.3	50
cis-1,3-Dichloropropene	0.507	0.526		3.75	50
1,1,2-Trichloroethane	0.339	0.360		6.2	50
2-Hexanone	0.403	0.410		1.74	50
Dibromochloromethane	0.326	0.346		6.14	50
1,2-Dibromoethane	0.345	0.361		4.64	50
Tetrachloroethene	0.332	0.328		-1.21	50
Chlorobenzene	1.071	1.072	0.3	0.09	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: P5280 SAS No.: P5280 SDG No.: P5280

Instrument ID: MSVOA_X Calibration Date/Time: 12/18/2024 19:52

Lab File ID: VX044395.D Init. Calib. Date(s): 12/11/2024 12/11/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:41 13:00

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.851	1.810		-2.21	50
m/p-Xylenes	0.679	0.687		1.18	50
o-Xylene	0.687	0.694		1.02	50
Styrene	1.102	1.146		3.99	50
Bromoform	0.219	0.241	0.1	10.05	50
Isopropylbenzene	3.828	3.704		-3.24	50
1,1,2,2-Tetrachloroethane	1.329	1.359	0.3	2.26	50
1,3-Dichlorobenzene	1.621	1.607		-0.86	50
1,4-Dichlorobenzene	1.667	1.616		-3.06	50
1,2-Dichlorobenzene	1.684	1.678		-0.36	50
1,2-Dibromo-3-Chloropropane	0.258	0.250		-3.1	50
1,2,4-Trichlorobenzene	0.923	0.932		0.98	50
1,2,3-Trichlorobenzene	0.979	0.985		0.61	50
1,2-Dichloroethane-d4	0.877	0.792		-9.69	50
Dibromofluoromethane	0.346	0.340		-1.73	50
Toluene-d8	1.168	1.114		-4.62	50
4-Bromofluorobenzene	0.408	0.378		-7.35	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: P5280 SAS No.: P5280 SDG No.: P5280

Instrument ID: MSVOA_X Calibration Date/Time: 12/20/2024 09:16

Lab File ID: VX044438.D Init. Calib. Date(s): 12/11/2024 12/11/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:41 13:00

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.773	0.854		10.48	20
Chloromethane	0.739	0.782	0.1	5.82	20
Vinyl Chloride	0.770	0.798		3.64	20
Bromomethane	0.546	0.532		-2.56	20
Chloroethane	0.480	0.473		-1.46	20
Trichlorofluoromethane	1.508	1.482		-1.72	20
1,1,2-Trichlorotrifluoroethane	0.610	0.676		10.82	20
1,1-Dichloroethene	0.560	0.624		11.43	20
Acetone	0.357	0.291		-18.49	20
Carbon Disulfide	1.065	1.106		3.85	20
Methyl tert-butyl Ether	2.187	2.093		-4.3	20
Methyl Acetate	0.922	0.896		-2.82	20
Methylene Chloride	0.671	0.689		2.68	20
trans-1,2-Dichloroethene	0.596	0.625		4.87	20
1,1-Dichloroethane	1.172	1.203	0.1	2.64	20
Cyclohexane	0.946	1.000		5.71	20
2-Butanone	0.508	0.457		-10.04	20
Carbon Tetrachloride	0.482	0.533		10.58	20
cis-1,2-Dichloroethene	0.756	0.755		-0.13	20
Bromochloromethane	0.575	0.553		-3.83	20
Chloroform	1.316	1.265		-3.88	20
1,1,1-Trichloroethane	1.093	1.099		0.55	20
Methylcyclohexane	0.535	0.648		21.12	20
Benzene	1.359	1.529		12.51	20
1,2-Dichloroethane	0.560	0.586		4.64	20
Trichloroethene	0.339	0.388		14.45	20
1,2-Dichloropropane	0.347	0.369		6.34	20
Bromodichloromethane	0.465	0.525		12.9	20
4-Methyl-2-Pentanone	0.556	0.573		3.06	20
Toluene	0.853	0.924		8.32	20
t-1,3-Dichloropropene	0.455	0.517		13.63	20
cis-1,3-Dichloropropene	0.507	0.577		13.81	20
1,1,2-Trichloroethane	0.339	0.374		10.32	20
2-Hexanone	0.403	0.414		2.73	20
Dibromochloromethane	0.326	0.381		16.87	20
1,2-Dibromoethane	0.345	0.385		11.59	20
Tetrachloroethene	0.332	0.378		13.85	20
Chlorobenzene	1.071	1.187	0.3	10.83	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: P5280 SAS No.: P5280 SDG No.: P5280

Instrument ID: MSVOA_X Calibration Date/Time: 12/20/2024 09:16

Lab File ID: VX044438.D Init. Calib. Date(s): 12/11/2024 12/11/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:41 13:00

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.851	2.034		9.89	20
m/p-Xylenes	0.679	0.780		14.88	20
o-Xylene	0.687	0.748		8.88	20
Styrene	1.102	1.275		15.7	20
Bromoform	0.219	0.266	0.1	21.46	20
Isopropylbenzene	3.828	4.095		6.97	20
1,1,2,2-Tetrachloroethane	1.329	1.387	0.3	4.36	20
1,3-Dichlorobenzene	1.621	1.811		11.72	20
1,4-Dichlorobenzene	1.667	1.781		6.84	20
1,2-Dichlorobenzene	1.684	1.820		8.08	20
1,2-Dibromo-3-Chloropropane	0.258	0.253		-1.94	20
1,2,4-Trichlorobenzene	0.923	1.098		18.96	20
1,2,3-Trichlorobenzene	0.979	1.140		16.44	20
1,2-Dichloroethane-d4	0.877	0.836		-4.68	20
Dibromofluoromethane	0.346	0.392		13.3	20
Toluene-d8	1.168	1.318		12.84	20
4-Bromofluorobenzene	0.408	0.454		11.27	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: P5280 SAS No.: P5280 SDG No.: P5280

Instrument ID: MSVOA_X Calibration Date/Time: 12/20/2024 19:49

Lab File ID: VX044464.D Init. Calib. Date(s): 12/11/2024 12/11/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:41 13:00

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.773	0.744		-3.75	50
Chloromethane	0.739	0.696	0.1	-5.82	50
Vinyl Chloride	0.770	0.730		-5.2	50
Bromomethane	0.546	0.498		-8.79	50
Chloroethane	0.480	0.507		5.63	50
Trichlorofluoromethane	1.508	1.202		-20.29	50
1,1,2-Trichlorotrifluoroethane	0.610	0.590		-3.28	50
1,1-Dichloroethene	0.560	0.560		0	50
Acetone	0.357	0.270		-24.37	50
Carbon Disulfide	1.065	1.089		2.25	50
Methyl tert-butyl Ether	2.187	1.972		-9.83	50
Methyl Acetate	0.922	0.860		-6.72	50
Methylene Chloride	0.671	0.636		-5.22	50
trans-1,2-Dichloroethene	0.596	0.578		-3.02	50
1,1-Dichloroethane	1.172	1.125	0.1	-4.01	50
Cyclohexane	0.946	0.871		-7.93	50
2-Butanone	0.508	0.433		-14.76	50
Carbon Tetrachloride	0.482	0.495		2.7	50
cis-1,2-Dichloroethene	0.756	0.719		-4.89	50
Bromochloromethane	0.575	0.510		-11.3	50
Chloroform	1.316	1.190		-9.57	50
1,1,1-Trichloroethane	1.093	1.029		-5.86	50
Methylcyclohexane	0.535	0.550		2.8	50
Benzene	1.359	1.393		2.5	50
1,2-Dichloroethane	0.560	0.534		-4.64	50
Trichloroethene	0.339	0.360		6.2	50
1,2-Dichloropropane	0.347	0.345		-0.58	50
Bromodichloromethane	0.465	0.491		5.59	50
4-Methyl-2-Pentanone	0.556	0.524		-5.76	50
Toluene	0.853	0.855		0.23	50
t-1,3-Dichloropropene	0.455	0.482		5.93	50
cis-1,3-Dichloropropene	0.507	0.531		4.73	50
1,1,2-Trichloroethane	0.339	0.345		1.77	50
2-Hexanone	0.403	0.381		-5.46	50
Dibromochloromethane	0.326	0.361		10.74	50
1,2-Dibromoethane	0.345	0.366		6.09	50
Tetrachloroethene	0.332	0.343		3.31	50
Chlorobenzene	1.071	1.094	0.3	2.15	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: P5280 SAS No.: P5280 SDG No.: P5280

Instrument ID: MSVOA_X Calibration Date/Time: 12/20/2024 19:49

Lab File ID: VX044464.D Init. Calib. Date(s): 12/11/2024 12/11/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:41 13:00

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.851	1.868		0.92	50
m/p-Xylenes	0.679	0.713		5.01	50
o-Xylene	0.687	0.708		3.06	50
Styrene	1.102	1.163		5.53	50
Bromoform	0.219	0.251	0.1	14.61	50
Isopropylbenzene	3.828	3.829		0.03	50
1,1,2,2-Tetrachloroethane	1.329	1.278	0.3	-3.84	50
1,3-Dichlorobenzene	1.621	1.669		2.96	50
1,4-Dichlorobenzene	1.667	1.653		-0.84	50
1,2-Dichlorobenzene	1.684	1.709		1.49	50
1,2-Dibromo-3-Chloropropane	0.258	0.238		-7.75	50
1,2,4-Trichlorobenzene	0.923	1.000		8.34	50
1,2,3-Trichlorobenzene	0.979	1.036		5.82	50
1,2-Dichloroethane-d4	0.877	0.778		-11.29	50
Dibromofluoromethane	0.346	0.363		4.91	50
Toluene-d8	1.168	1.213		3.85	50
4-Bromofluorobenzene	0.408	0.418		2.45	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.