

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: P5281 SAS No.: P5281 SDG No.: P5281

Instrument ID: BNA_N Calibration Date/Time: 12/20/2024 18:18

Lab File ID: BN035747.D Init. Calib. Date(s): 11/27/2024 11/27/2024

EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 15:34 19:09

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.626	0.618		-1.3	50.0
Fluoranthene-d10	1.134	1.034		-8.8	50.0
2-Fluorophenol	1.001	0.002		-99.8	50.0
Phenol-d6	1.204	0.025		-97.9	50.0
Nitrobenzene-d5	0.244	0.005		-98.0	50.0
2-Fluorobiphenyl	1.512	0.002		-99.9	50.0
2,4,6-Tribromophenol	0.284	0.001		-99.6	50.0
Terphenyl-d14	0.789	0.003		-99.6	50.0
1,4-Dioxane	0.382	0.448		17.3	50.0

All other compounds must meet a minimum RRF of 0.010.