

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P5281 SAS No.: P5281 SDG NO.: P5281

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 12/20/2024

Lab File ID: BN035731.D Time Analyzed: 08:04

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	3239	7.271	8431	10.02	5341	13.92
UPPER LIMIT	6478	7.771	16862	10.52	10682	14.424
LOWER LIMIT	1619.5	6.771	4215.5	9.52	2670.5	13.424
EPA SAMPLE NO.						
01 BPOW6-11-20241212MS	3051	7.27	8358	10.01	5197	13.92
02 BPOW6-11-20241212MSD	3357	7.27	9243	10.01	5773	13.92
03 RE114D1-20241212	3072	7.27	8100	10.01	5112	13.93

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.