

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P5281 SAS No.: P5281 SDG NO.: P5281

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 12/24/2024

Lab File ID: BN035811.D Time Analyzed: 08:59

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	3972	7.264	9483	10.00	5387	13.92
UPPER LIMIT	7944	7.764	18966	10.498	10774	14.424
LOWER LIMIT	1986	6.764	4741.5	9.498	2693.5	13.424
EPA SAMPLE NO.						
01 TT101D2-20241212DL	3005	7.26	6792	10.01	4271	13.92

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.