

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P5281 SAS No.: P5281 SDG NO.: P5281

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 12/26/2024

Lab File ID: BN035845.D Time Analyzed: 09:58

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	3755	7.257	9322	10.00	5387	13.91
UPPER LIMIT	7510	7.757	18644	10.499	10774	14.414
LOWER LIMIT	1877.5	6.757	4661	9.499	2693.5	13.414
EPA SAMPLE NO.						
01 RE115D1-20241213DL	2930	7.26	6901	10.00	4611	13.91
02 TT189D2-20241213DL	3270	7.26	7531	10.00	4933	13.91

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.