

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P5281 SAS No.: P5281 SDG NO.: P5281

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 12/26/2024

Lab File ID: BN035845.D Time Analyzed: 09:58

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	11262	16.686	9932	20.938	10817	23.021
UPPER LIMIT	22524	17.186	19864	21.438	21634	23.521
LOWER LIMIT	5631	16.186	4966	20.438	5408.5	22.521
EPA SAMPLE NO.						
01 RE115D1-20241213DL	10065	16.70	7379	20.95	6927	23.03
02 TT189D2-20241213DL	10717	16.69	7941	20.95	7194	23.03

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.