

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: AECO15

Lab Code: CHEM

SAS No.: P5281 SDG NO.: P5281

Lab File ID: BN035793.D

DFTPP Injection Date: 12/23/2024

Instrument ID: BNA_N

DFTPP Injection Time: 20:26

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	23.2
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	32.6
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	39.9
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	26.6
365	Greater than 1% of mass 198	3.7
441	Present, but less than mass 443	11.1
442	Greater than 50% of mass 198	100.0
443	15.0 - 24.0% of mass 442	13.8 (22.6) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN035794.D	12/23/2024	22:16
PB165667BL	PB165667BL	BN035795.D	12/23/2024	22:52
BPOW1-8-20241212	P5281-03	BN035796.D	12/23/2024	23:27
RE117D1-20241213	P5281-05	BN035797.D	12/24/2024	00:03
RE115D1-20241213	P5281-06	BN035798.D	12/24/2024	00:38
TT189D2-20241213	P5281-07	BN035799.D	12/24/2024	01:14
FB02-20241213	P5281-08	BN035800.D	12/24/2024	01:49
PB165667BS	PB165667BS	BN035808.D	12/24/2024	06:34
SSTDCCC0.4EC	SSTDCCC0.4	BN035809.D	12/24/2024	07:09