

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: AECO15

Lab Code: CHEM

SAS No.: P5281 SDG NO.: P5281

Lab File ID: BN035810.D

DFTPP Injection Date: 12/24/2024

Instrument ID: BNA_N

DFTPP Injection Time: 07:45

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	22.6
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	30.9
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	39.2
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	29.6
365	Greater than 1% of mass 198	4.5
441	Present, but less than mass 443	13.2
442	Greater than 50% of mass 198	100.0
443	15.0 - 24.0% of mass 442	15.2 (18.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN035811.D	12/24/2024	08:59
TT101D2-20241212DL	P5281-04DL	BN035813.D	12/24/2024	10:10
SSTDCCC0.4EC	SSTDCCC0.4	BN035828.D	12/24/2024	19:05