

LAB CHRONICLE

OrderID:	P5281	OrderDate:	12/13/2024 1:05:00 PM
Client:	AECOM Technical Services, Inc.	Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258
Contact:	Eleanor Vivadou	Location:	L61,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
P5281-01	RE114D1-20241212	Water	VOCMS Group1	8260-Low	12/12/24		12/18/24	12/13/24
P5281-01DL	RE114D1-20241212D L	Water	VOCMS Group1	8260-Low	12/12/24		12/20/24	12/13/24
P5281-02	BPOW1-7-20241212	Water	VOCMS Group1	8260-Low	12/12/24		12/20/24	12/13/24
P5281-03	BPOW1-8-20241212	Water	VOCMS Group1	8260-Low	12/12/24		12/18/24	12/13/24
P5281-04	TT101D2-20241212	Water	VOCMS Group1	8260-Low	12/12/24		12/18/24	12/13/24
P5281-05	RE117D1-20241213	Water	VOCMS Group1	8260-Low	12/13/24		12/18/24	12/13/24
P5281-06	RE115D1-20241213	Water	VOCMS Group1	8260-Low	12/13/24		12/18/24	12/13/24
P5281-06DL	RE115D1-20241213D L	Water	VOCMS Group1	8260-Low	12/13/24		12/20/24	12/13/24
P5281-07	TT189D2-20241213	Water	VOCMS Group1	8260-Low	12/13/24		12/18/24	12/13/24
P5281-08	FB02-20241213	Water	VOCMS Group1	8260-Low	12/13/24		12/18/24	12/13/24
P5281-09	TB	Water	VOCMS Group1	8260-Low	12/13/24		12/18/24	12/13/24

Hit Summary Sheet SW-846

SDG No.: P5281

Client: AECOM Technical Services, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID:	RE114D1-20241212								
P5281-01	RE114D1-20241212	Water	Dichlorodifluoromethane	1.00		0.21	0.50	1.00	ug/L
P5281-01	RE114D1-20241212	Water	1,1,2-Trichlorotrifluoroethane	21.7		0.25	0.50	1.00	ug/L
P5281-01	RE114D1-20241212	Water	1,1-Dichloroethene	3.50		0.26	0.75	1.00	ug/L
P5281-01	RE114D1-20241212	Water	1,1-Dichloroethane	0.97	J	0.23	0.50	1.00	ug/L
P5281-01	RE114D1-20241212	Water	Carbon Tetrachloride	1.80		0.25	0.50	1.00	ug/L
P5281-01	RE114D1-20241212	Water	cis-1,2-Dichloroethene	3.60		0.25	0.75	1.00	ug/L
P5281-01	RE114D1-20241212	Water	Chloroform	1.90		0.26	0.50	1.00	ug/L
P5281-01	RE114D1-20241212	Water	Trichloroethene	390	E	0.32	0.75	1.00	ug/L
P5281-01	RE114D1-20241212	Water	1,1,2-Trichloroethane	1.40		0.21	0.50	1.00	ug/L
P5281-01	RE114D1-20241212	Water	m/p-Xylenes	0.34	J	0.31	1.00	2.00	ug/L
			Total Voc :	426					
			Total Concentration:	426					
Client ID:	RE114D1-20241212DL								
P5281-01DL	RE114D1-20241212	Water	1,1,2-Trichlorotrifluoroethane	23.0	D	2.50	5.00	10.0	ug/L
P5281-01DL	RE114D1-20241212	Water	1,1-Dichloroethene	3.80	JD	2.60	7.50	10.0	ug/L
P5281-01DL	RE114D1-20241212	Water	cis-1,2-Dichloroethene	3.40	JD	2.50	7.50	10.0	ug/L
P5281-01DL	RE114D1-20241212	Water	Trichloroethene	410	D	3.20	7.50	10.0	ug/L
			Total Voc :	440					
			Total Concentration:	440					
Client ID:	BPOW1-7-20241212								
P5281-02	BPOW1-7-20241212	Water	Trichloroethene	0.61	J	0.32	0.75	1.00	ug/L
			Total Voc :	0.61					
			Total Concentration:	0.61					
Client ID:	BPOW1-8-20241212								
P5281-03	BPOW1-8-20241212	Water	Trichloroethene	0.59	J	0.32	0.75	1.00	ug/L
			Total Voc :	0.59					
			Total Concentration:	0.59					
Client ID:	TT101D2-20241212								
P5281-04	TT101D2-20241212	Water	1,1,2-Trichlorotrifluoroethane	32.4	J	10.0	20.0	40.0	ug/L
P5281-04	TT101D2-20241212	Water	1,1-Dichloroethene	13.2	J	10.4	30.0	40.0	ug/L
P5281-04	TT101D2-20241212	Water	Trichloroethene	1600		12.8	30.0	40.0	ug/L
			Total Voc :	1650					
			Total Concentration:	1650					
Client ID:	RE117D1-20241213								
P5281-05	RE117D1-20241213	Water	1,1,2-Trichlorotrifluoroethane	0.89	J	0.25	0.50	1.00	ug/L
P5281-05	RE117D1-20241213	Water	1,1-Dichloroethene	1.30		0.26	0.75	1.00	ug/L
P5281-05	RE117D1-20241213	Water	Carbon Tetrachloride	1.60		0.25	0.50	1.00	ug/L
P5281-05	RE117D1-20241213	Water	cis-1,2-Dichloroethene	1.20		0.25	0.75	1.00	ug/L

Hit Summary Sheet SW-846

SDG No.: P5281

Client: AECOM Technical Services, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
P5281-05	RE117D1-20241213	Water	Chloroform	0.44	J	0.26	0.50	1.00	ug/L
P5281-05	RE117D1-20241213	Water	Trichloroethene	130		0.32	0.75	1.00	ug/L
P5281-05	RE117D1-20241213	Water	1,1,2-Trichloroethane	0.61	J	0.21	0.50	1.00	ug/L
			Total Voc :	136					
			Total Concentration:	136					
Client ID:	RE115D1-20241213								
P5281-06	RE115D1-20241213	Water	Dichlorodifluoromethane	1.30		0.21	0.50	1.00	ug/L
P5281-06	RE115D1-20241213	Water	1,1,2-Trichlorotrifluoroethane	23.4		0.25	0.50	1.00	ug/L
P5281-06	RE115D1-20241213	Water	1,1-Dichloroethene	10.1		0.26	0.75	1.00	ug/L
P5281-06	RE115D1-20241213	Water	Methylene Chloride	0.41	J	0.32	0.50	1.00	ug/L
P5281-06	RE115D1-20241213	Water	1,1-Dichloroethane	1.30		0.23	0.50	1.00	ug/L
P5281-06	RE115D1-20241213	Water	Carbon Tetrachloride	2.90		0.25	0.50	1.00	ug/L
P5281-06	RE115D1-20241213	Water	cis-1,2-Dichloroethene	2.70		0.25	0.75	1.00	ug/L
P5281-06	RE115D1-20241213	Water	Chloroform	1.70		0.26	0.50	1.00	ug/L
P5281-06	RE115D1-20241213	Water	Trichloroethene	540	E	0.32	0.75	1.00	ug/L
P5281-06	RE115D1-20241213	Water	1,1,2-Trichloroethane	1.10		0.21	0.50	1.00	ug/L
P5281-06	RE115D1-20241213	Water	Tetrachloroethene	0.45	J	0.25	0.50	1.00	ug/L
			Total Voc :	585					
			Total Concentration:	585					
Client ID:	RE115D1-20241213DL								
P5281-06DL	RE115D1-20241213	Water	1,1,2-Trichlorotrifluoroethane	32.6	D	5.00	10.0	20.0	ug/L
P5281-06DL	RE115D1-20241213	Water	1,1-Dichloroethene	11.9	JD	5.20	15.0	20.0	ug/L
P5281-06DL	RE115D1-20241213	Water	Trichloroethene	670	D	6.40	15.0	20.0	ug/L
			Total Voc :	715					
			Total Concentration:	715					
Client ID:	TT189D2-20241213								
P5281-07	TT189D2-20241213	Water	Dichlorodifluoromethane	0.54	J	0.21	0.50	1.00	ug/L
P5281-07	TT189D2-20241213	Water	Chloromethane	0.54	J	0.35	0.50	1.00	ug/L
P5281-07	TT189D2-20241213	Water	1,1,2-Trichlorotrifluoroethane	21.4		0.25	0.50	1.00	ug/L
P5281-07	TT189D2-20241213	Water	1,1-Dichloroethene	1.20		0.26	0.75	1.00	ug/L
P5281-07	TT189D2-20241213	Water	Acetone	1.90	J	1.40	3.80	5.00	ug/L
P5281-07	TT189D2-20241213	Water	Carbon Tetrachloride	0.50	J	0.25	0.50	1.00	ug/L
P5281-07	TT189D2-20241213	Water	cis-1,2-Dichloroethene	2.80		0.25	0.75	1.00	ug/L
P5281-07	TT189D2-20241213	Water	Chloroform	0.42	J	0.26	0.50	1.00	ug/L
P5281-07	TT189D2-20241213	Water	Trichloroethene	94.8		0.32	0.75	1.00	ug/L
P5281-07	TT189D2-20241213	Water	Tetrachloroethene	0.39	J	0.25	0.50	1.00	ug/L
			Total Voc :	124					
			Total Concentration:	124					

Hit Summary Sheet
SW-846

SDG No.: P5281

Client: AECOM Technical Services, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID:	FB02-20241213								
P5281-08	FB02-20241213	Water	Acetone	2.60	J	1.40	3.80	5.00	ug/L
P5281-08	FB02-20241213	Water	Methylene Chloride	0.33	J	0.32	0.50	1.00	ug/L
P5281-08	FB02-20241213	Water	2-Butanone	2.20	J	1.30	2.50	5.00	ug/L
P5281-08	FB02-20241213	Water	Trichloroethene	0.40	J	0.32	0.75	1.00	ug/L
			Total Voc :	5.53					
			Total Concentration:	5.53					
Client ID:	TB								
P5281-09	TB	Water	Chloromethane	0.58	J	0.35	0.50	1.00	ug/L
P5281-09	TB	Water	Acetone	1.50	J	1.40	3.80	5.00	ug/L
P5281-09	TB	Water	Methylene Chloride	0.45	J	0.32	0.50	1.00	ug/L
			Total Voc :	2.53					
			Total Concentration:	2.53					



SAMPLE DATA

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/12/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	RE114D1-20241212		SDG No.:	P5281	
Lab Sample ID:	P5281-01		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044388.D	1		12/18/24 17:08	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	1.00		0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	21.7		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	3.50		0.26	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.97	J	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	1.80		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	3.60		0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
67-66-3	Chloroform	1.90		0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	390	E	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.			Date Collected:	12/12/24		
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258			Date Received:	12/13/24		
Client Sample ID:	RE114D1-20241212			SDG No.:	P5281		
Lab Sample ID:	P5281-01			Matrix:	Water		
Analytical Method:	SW8260			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	DB-624UI	ID :	0.18	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044388.D	1		12/18/24 17:08	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	1.40		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	0.34	J	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	45.2		81 - 118		90%	SPK: 50
1868-53-7	Dibromofluoromethane	47.8		80 - 119		96%	SPK: 50
2037-26-5	Toluene-d8	51.2		89 - 112		102%	SPK: 50
460-00-4	4-Bromofluorobenzene	48.8		85 - 114		98%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	116000	5.544				
540-36-3	1,4-Difluorobenzene	209000	6.757				
3114-55-4	Chlorobenzene-d5	183000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	75800	12.024				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/12/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	RE114D1-20241212		SDG No.:	P5281	
Lab Sample ID:	P5281-01		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044388.D	1		12/18/24 17:08	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.	Date Collected:	12/12/24
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258	Date Received:	12/13/24
Client Sample ID:	RE114D1-20241212DL	SDG No.:	P5281
Lab Sample ID:	P5281-01DL	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044444.D	10		12/20/24 12:03	VX122024

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	5.00	UD	2.10	5.00	10.0	ug/L
74-87-3	Chloromethane	5.00	UD	3.50	5.00	10.0	ug/L
75-01-4	Vinyl Chloride	7.50	UD	3.40	7.50	10.0	ug/L
74-83-9	Bromomethane	37.5	UD	13.6	37.5	50.0	ug/L
75-00-3	Chloroethane	7.50	UD	5.60	7.50	10.0	ug/L
75-69-4	Trichlorofluoromethane	5.00	UD	3.40	5.00	10.0	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	23.0	D	2.50	5.00	10.0	ug/L
75-35-4	1,1-Dichloroethene	3.80	JD	2.60	7.50	10.0	ug/L
67-64-1	Acetone	37.5	UD	13.9	37.5	50.0	ug/L
75-15-0	Carbon Disulfide	7.50	UD	3.20	7.50	10.0	ug/L
1634-04-4	Methyl tert-butyl Ether	5.00	UD	1.60	5.00	10.0	ug/L
79-20-9	Methyl Acetate	7.50	UD	6.00	7.50	10.0	ug/L
75-09-2	Methylene Chloride	5.00	UD	3.20	5.00	10.0	ug/L
156-60-5	trans-1,2-Dichloroethene	5.00	UD	2.50	5.00	10.0	ug/L
75-34-3	1,1-Dichloroethane	5.00	UD	2.30	5.00	10.0	ug/L
110-82-7	Cyclohexane	25.0	UD	16.2	25.0	50.0	ug/L
78-93-3	2-Butanone	25.0	UD	13.0	25.0	50.0	ug/L
56-23-5	Carbon Tetrachloride	5.00	UD	2.50	5.00	10.0	ug/L
156-59-2	cis-1,2-Dichloroethene	3.40	JD	2.50	7.50	10.0	ug/L
74-97-5	Bromochloromethane	5.00	UD	1.80	5.00	10.0	ug/L
67-66-3	Chloroform	5.00	UD	2.60	5.00	10.0	ug/L
71-55-6	1,1,1-Trichloroethane	5.00	UD	1.90	5.00	10.0	ug/L
108-87-2	Methylcyclohexane	5.00	UD	1.90	5.00	10.0	ug/L
71-43-2	Benzene	5.00	UD	1.60	5.00	10.0	ug/L
107-06-2	1,2-Dichloroethane	5.00	UD	2.40	5.00	10.0	ug/L
79-01-6	Trichloroethene	410	D	3.20	7.50	10.0	ug/L
78-87-5	1,2-Dichloropropane	5.00	UD	1.90	5.00	10.0	ug/L
75-27-4	Bromodichloromethane	5.00	UD	2.40	5.00	10.0	ug/L
108-10-1	4-Methyl-2-Pentanone	25.0	UD	7.50	25.0	50.0	ug/L
108-88-3	Toluene	5.00	UD	1.80	5.00	10.0	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.	Date Collected:	12/12/24
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258	Date Received:	12/13/24
Client Sample ID:	RE114D1-20241212DL	SDG No.:	P5281
Lab Sample ID:	P5281-01DL	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044444.D	10		12/20/24 12:03	VX122024

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	5.00	UD	2.10	5.00	10.0	ug/L
10061-01-5	cis-1,3-Dichloropropene	5.00	UD	1.80	5.00	10.0	ug/L
79-00-5	1,1,2-Trichloroethane	5.00	UD	2.10	5.00	10.0	ug/L
591-78-6	2-Hexanone	25.0	UD	11.3	25.0	50.0	ug/L
124-48-1	Dibromochloromethane	5.00	UD	1.80	5.00	10.0	ug/L
106-93-4	1,2-Dibromoethane	5.00	UD	1.60	5.00	10.0	ug/L
127-18-4	Tetrachloroethene	5.00	UD	2.50	5.00	10.0	ug/L
108-90-7	Chlorobenzene	5.00	UD	1.30	5.00	10.0	ug/L
100-41-4	Ethyl Benzene	5.00	UD	1.60	5.00	10.0	ug/L
179601-23-1	m/p-Xylenes	10.0	UD	3.10	10.0	20.0	ug/L
95-47-6	o-Xylene	5.00	UD	1.40	5.00	10.0	ug/L
100-42-5	Styrene	5.00	UD	1.60	5.00	10.0	ug/L
75-25-2	Bromoform	5.00	UD	2.10	5.00	10.0	ug/L
98-82-8	Isopropylbenzene	5.00	UD	1.30	5.00	10.0	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	5.00	UD	2.70	5.00	10.0	ug/L
541-73-1	1,3-Dichlorobenzene	5.00	UD	2.40	5.00	10.0	ug/L
106-46-7	1,4-Dichlorobenzene	5.00	UD	2.70	5.00	10.0	ug/L
95-50-1	1,2-Dichlorobenzene	5.00	UD	1.90	5.00	10.0	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	7.50	UD	4.60	7.50	10.0	ug/L
120-82-1	1,2,4-Trichlorobenzene	5.00	UD	4.20	5.00	10.0	ug/L
87-61-6	1,2,3-Trichlorobenzene	7.50	UD	5.10	7.50	10.0	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	42.9		81 - 118		86%	SPK: 50
1868-53-7	Dibromofluoromethane	48.3		80 - 119		97%	SPK: 50
2037-26-5	Toluene-d8	50.7		89 - 112		101%	SPK: 50
460-00-4	4-Bromofluorobenzene	46.3		85 - 114		93%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	119000	5.55				
540-36-3	1,4-Difluorobenzene	212000	6.757				
3114-55-4	Chlorobenzene-d5	185000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	75000	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/12/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	RE114D1-20241212DL		SDG No.:	P5281	
Lab Sample ID:	P5281-01DL		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044444.D	10		12/20/24 12:03	VX122024

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/12/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	BPOW1-7-20241212		SDG No.:	P5281	
Lab Sample ID:	P5281-02		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044451.D	1		12/20/24 14:46	VX122024

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.61	J	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.			Date Collected:	12/12/24		
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258			Date Received:	12/13/24		
Client Sample ID:	BPOW1-7-20241212			SDG No.:	P5281		
Lab Sample ID:	P5281-02			Matrix:	Water		
Analytical Method:	SW8260			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	DB-624UI	ID :	0.18	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044451.D	1		12/20/24 14:46	VX122024

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	42.4		81 - 118		85%	SPK: 50
1868-53-7	Dibromofluoromethane	47.4		80 - 119		95%	SPK: 50
2037-26-5	Toluene-d8	50.5		89 - 112		101%	SPK: 50
460-00-4	4-Bromofluorobenzene	43.8		85 - 114		88%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	117000	5.55				
540-36-3	1,4-Difluorobenzene	206000	6.757				
3114-55-4	Chlorobenzene-d5	173000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	66500	12.024				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/12/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	BPOW1-7-20241212		SDG No.:	P5281	
Lab Sample ID:	P5281-02		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044451.D	1		12/20/24 14:46	VX122024

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/12/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	BPOW1-8-20241212		SDG No.:	P5281	
Lab Sample ID:	P5281-03		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044390.D	1		12/18/24 17:55	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.59	J	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/12/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	BPOW1-8-20241212		SDG No.:	P5281	
Lab Sample ID:	P5281-03		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044390.D	1		12/18/24 17:55	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	44.5		81 - 118		89%	SPK: 50
1868-53-7	Dibromofluoromethane	47.0		80 - 119		94%	SPK: 50
2037-26-5	Toluene-d8	51.6		89 - 112		103%	SPK: 50
460-00-4	4-Bromofluorobenzene	47.8		85 - 114		96%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	114000	5.543				
540-36-3	1,4-Difluorobenzene	208000	6.757				
3114-55-4	Chlorobenzene-d5	184000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	77800	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/12/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	BPOW1-8-20241212		SDG No.:	P5281	
Lab Sample ID:	P5281-03		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044390.D	1		12/18/24 17:55	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/12/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	TT101D2-20241212		SDG No.:	P5281	
Lab Sample ID:	P5281-04		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044381.D	40		12/18/24 14:24	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	20.0	U	8.40	20.0	40.0	ug/L
74-87-3	Chloromethane	20.0	U	14.0	20.0	40.0	ug/L
75-01-4	Vinyl Chloride	30.0	U	13.6	30.0	40.0	ug/L
74-83-9	Bromomethane	150	U	54.4	150	200	ug/L
75-00-3	Chloroethane	30.0	U	22.4	30.0	40.0	ug/L
75-69-4	Trichlorofluoromethane	20.0	U	13.6	20.0	40.0	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	32.4	J	10.0	20.0	40.0	ug/L
75-35-4	1,1-Dichloroethene	13.2	J	10.4	30.0	40.0	ug/L
67-64-1	Acetone	150	U	55.6	150	200	ug/L
75-15-0	Carbon Disulfide	30.0	U	12.8	30.0	40.0	ug/L
1634-04-4	Methyl tert-butyl Ether	20.0	U	6.40	20.0	40.0	ug/L
79-20-9	Methyl Acetate	30.0	U	24.0	30.0	40.0	ug/L
75-09-2	Methylene Chloride	20.0	U	12.8	20.0	40.0	ug/L
156-60-5	trans-1,2-Dichloroethene	20.0	U	10.0	20.0	40.0	ug/L
75-34-3	1,1-Dichloroethane	20.0	U	9.20	20.0	40.0	ug/L
110-82-7	Cyclohexane	100	U	64.8	100	200	ug/L
78-93-3	2-Butanone	100	U	52.0	100	200	ug/L
56-23-5	Carbon Tetrachloride	20.0	U	10.0	20.0	40.0	ug/L
156-59-2	cis-1,2-Dichloroethene	30.0	U	10.0	30.0	40.0	ug/L
74-97-5	Bromochloromethane	20.0	U	7.20	20.0	40.0	ug/L
67-66-3	Chloroform	20.0	U	10.4	20.0	40.0	ug/L
71-55-6	1,1,1-Trichloroethane	20.0	U	7.60	20.0	40.0	ug/L
108-87-2	Methylcyclohexane	20.0	U	7.60	20.0	40.0	ug/L
71-43-2	Benzene	20.0	U	6.40	20.0	40.0	ug/L
107-06-2	1,2-Dichloroethane	20.0	U	9.60	20.0	40.0	ug/L
79-01-6	Trichloroethene	1600		12.8	30.0	40.0	ug/L
78-87-5	1,2-Dichloropropane	20.0	U	7.60	20.0	40.0	ug/L
75-27-4	Bromodichloromethane	20.0	U	9.60	20.0	40.0	ug/L
108-10-1	4-Methyl-2-Pentanone	100	U	30.0	100	200	ug/L
108-88-3	Toluene	20.0	U	7.20	20.0	40.0	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.			Date Collected:	12/12/24		
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258			Date Received:	12/13/24		
Client Sample ID:	TT101D2-20241212			SDG No.:	P5281		
Lab Sample ID:	P5281-04			Matrix:	Water		
Analytical Method:	SW8260			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	DB-624UI	ID :	0.18	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044381.D	40		12/18/24 14:24	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	20.0	U	8.40	20.0	40.0	ug/L
10061-01-5	cis-1,3-Dichloropropene	20.0	U	7.20	20.0	40.0	ug/L
79-00-5	1,1,2-Trichloroethane	20.0	U	8.40	20.0	40.0	ug/L
591-78-6	2-Hexanone	100	U	45.2	100	200	ug/L
124-48-1	Dibromochloromethane	20.0	U	7.20	20.0	40.0	ug/L
106-93-4	1,2-Dibromoethane	20.0	U	6.40	20.0	40.0	ug/L
127-18-4	Tetrachloroethene	20.0	U	10.0	20.0	40.0	ug/L
108-90-7	Chlorobenzene	20.0	U	5.20	20.0	40.0	ug/L
100-41-4	Ethyl Benzene	20.0	U	6.40	20.0	40.0	ug/L
179601-23-1	m/p-Xylenes	40.0	U	12.4	40.0	80.0	ug/L
95-47-6	o-Xylene	20.0	U	5.60	20.0	40.0	ug/L
100-42-5	Styrene	20.0	U	6.40	20.0	40.0	ug/L
75-25-2	Bromoform	20.0	U	8.40	20.0	40.0	ug/L
98-82-8	Isopropylbenzene	20.0	U	5.20	20.0	40.0	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	20.0	U	10.8	20.0	40.0	ug/L
541-73-1	1,3-Dichlorobenzene	20.0	U	9.60	20.0	40.0	ug/L
106-46-7	1,4-Dichlorobenzene	20.0	U	10.8	20.0	40.0	ug/L
95-50-1	1,2-Dichlorobenzene	20.0	U	7.60	20.0	40.0	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	30.0	U	18.4	30.0	40.0	ug/L
120-82-1	1,2,4-Trichlorobenzene	20.0	U	16.8	20.0	40.0	ug/L
87-61-6	1,2,3-Trichlorobenzene	30.0	U	20.4	30.0	40.0	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	43.6		81 - 118		87%	SPK: 50
1868-53-7	Dibromofluoromethane	47.1		80 - 119		94%	SPK: 50
2037-26-5	Toluene-d8	52.2		89 - 112		104%	SPK: 50
460-00-4	4-Bromofluorobenzene	47.4		85 - 114		95%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	121000	5.55				
540-36-3	1,4-Difluorobenzene	214000	6.757				
3114-55-4	Chlorobenzene-d5	187000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	76900	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/12/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	TT101D2-20241212		SDG No.:	P5281	
Lab Sample ID:	P5281-04		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044381.D	40		12/18/24 14:24	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/13/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	RE117D1-20241213		SDG No.:	P5281	
Lab Sample ID:	P5281-05		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044391.D	1		12/18/24 18:18	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.89	J	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	1.30		0.26	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	1.60		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	1.20		0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
67-66-3	Chloroform	0.44	J	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	130		0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/13/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	RE117D1-20241213		SDG No.:	P5281	
Lab Sample ID:	P5281-05		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044391.D	1		12/18/24 18:18	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.61	J	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	44.9		81 - 118		90%	SPK: 50
1868-53-7	Dibromofluoromethane	47.0		80 - 119		94%	SPK: 50
2037-26-5	Toluene-d8	51.1		89 - 112		102%	SPK: 50
460-00-4	4-Bromofluorobenzene	45.8		85 - 114		92%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	112000	5.55				
540-36-3	1,4-Difluorobenzene	202000	6.757				
3114-55-4	Chlorobenzene-d5	173000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	67900	12.024				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/13/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	RE117D1-20241213		SDG No.:	P5281	
Lab Sample ID:	P5281-05		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044391.D	1		12/18/24 18:18	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/13/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	RE115D1-20241213		SDG No.:	P5281	
Lab Sample ID:	P5281-06		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044392.D	1		12/18/24 18:42	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	1.30		0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	23.4		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	10.1		0.26	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.41	J	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	1.30		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	2.90		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	2.70		0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
67-66-3	Chloroform	1.70		0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	540	E	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/13/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	RE115D1-20241213		SDG No.:	P5281	
Lab Sample ID:	P5281-06		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044392.D	1		12/18/24 18:42	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	1.10		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.45	J	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	43.0		81 - 118		86%	SPK: 50
1868-53-7	Dibromofluoromethane	46.9		80 - 119		94%	SPK: 50
2037-26-5	Toluene-d8	50.6		89 - 112		101%	SPK: 50
460-00-4	4-Bromofluorobenzene	46.0		85 - 114		92%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	116000	5.55				
540-36-3	1,4-Difluorobenzene	207000	6.757				
3114-55-4	Chlorobenzene-d5	178000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	70800	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/13/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	RE115D1-20241213		SDG No.:	P5281	
Lab Sample ID:	P5281-06		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044392.D	1		12/18/24 18:42	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/13/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	RE115D1-20241213DL		SDG No.:	P5281	
Lab Sample ID:	P5281-06DL		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044455.D	20		12/20/24 16:19	VX122024

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	10.0	UD	4.20	10.0	20.0	ug/L
74-87-3	Chloromethane	10.0	UD	7.00	10.0	20.0	ug/L
75-01-4	Vinyl Chloride	15.0	UD	6.80	15.0	20.0	ug/L
74-83-9	Bromomethane	75.0	UD	27.2	75.0	100	ug/L
75-00-3	Chloroethane	15.0	UD	11.2	15.0	20.0	ug/L
75-69-4	Trichlorofluoromethane	10.0	UD	6.80	10.0	20.0	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	32.6	D	5.00	10.0	20.0	ug/L
75-35-4	1,1-Dichloroethene	11.9	JD	5.20	15.0	20.0	ug/L
67-64-1	Acetone	75.0	UD	27.8	75.0	100	ug/L
75-15-0	Carbon Disulfide	15.0	UD	6.40	15.0	20.0	ug/L
1634-04-4	Methyl tert-butyl Ether	10.0	UD	3.20	10.0	20.0	ug/L
79-20-9	Methyl Acetate	15.0	UD	12.0	15.0	20.0	ug/L
75-09-2	Methylene Chloride	10.0	UD	6.40	10.0	20.0	ug/L
156-60-5	trans-1,2-Dichloroethene	10.0	UD	5.00	10.0	20.0	ug/L
75-34-3	1,1-Dichloroethane	10.0	UD	4.60	10.0	20.0	ug/L
110-82-7	Cyclohexane	50.0	UD	32.4	50.0	100	ug/L
78-93-3	2-Butanone	50.0	UD	26.0	50.0	100	ug/L
56-23-5	Carbon Tetrachloride	10.0	UD	5.00	10.0	20.0	ug/L
156-59-2	cis-1,2-Dichloroethene	15.0	UD	5.00	15.0	20.0	ug/L
74-97-5	Bromochloromethane	10.0	UD	3.60	10.0	20.0	ug/L
67-66-3	Chloroform	10.0	UD	5.20	10.0	20.0	ug/L
71-55-6	1,1,1-Trichloroethane	10.0	UD	3.80	10.0	20.0	ug/L
108-87-2	Methylcyclohexane	10.0	UD	3.80	10.0	20.0	ug/L
71-43-2	Benzene	10.0	UD	3.20	10.0	20.0	ug/L
107-06-2	1,2-Dichloroethane	10.0	UD	4.80	10.0	20.0	ug/L
79-01-6	Trichloroethene	670	D	6.40	15.0	20.0	ug/L
78-87-5	1,2-Dichloropropane	10.0	UD	3.80	10.0	20.0	ug/L
75-27-4	Bromodichloromethane	10.0	UD	4.80	10.0	20.0	ug/L
108-10-1	4-Methyl-2-Pentanone	50.0	UD	15.0	50.0	100	ug/L
108-88-3	Toluene	10.0	UD	3.60	10.0	20.0	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.			Date Collected:	12/13/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258			Date Received:	12/13/24	
Client Sample ID:	RE115D1-20241213DL			SDG No.:	P5281	
Lab Sample ID:	P5281-06DL			Matrix:	Water	
Analytical Method:	SW8260			% Solid:	0	
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID :	0.18	Level :	LOW	
Prep Method :						

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044455.D	20		12/20/24 16:19	VX122024

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	10.0	UD	4.20	10.0	20.0	ug/L
10061-01-5	cis-1,3-Dichloropropene	10.0	UD	3.60	10.0	20.0	ug/L
79-00-5	1,1,2-Trichloroethane	10.0	UD	4.20	10.0	20.0	ug/L
591-78-6	2-Hexanone	50.0	UD	22.6	50.0	100	ug/L
124-48-1	Dibromochloromethane	10.0	UD	3.60	10.0	20.0	ug/L
106-93-4	1,2-Dibromoethane	10.0	UD	3.20	10.0	20.0	ug/L
127-18-4	Tetrachloroethene	10.0	UD	5.00	10.0	20.0	ug/L
108-90-7	Chlorobenzene	10.0	UD	2.60	10.0	20.0	ug/L
100-41-4	Ethyl Benzene	10.0	UD	3.20	10.0	20.0	ug/L
179601-23-1	m/p-Xylenes	20.0	UD	6.20	20.0	40.0	ug/L
95-47-6	o-Xylene	10.0	UD	2.80	10.0	20.0	ug/L
100-42-5	Styrene	10.0	UD	3.20	10.0	20.0	ug/L
75-25-2	Bromoform	10.0	UD	4.20	10.0	20.0	ug/L
98-82-8	Isopropylbenzene	10.0	UD	2.60	10.0	20.0	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	10.0	UD	5.40	10.0	20.0	ug/L
541-73-1	1,3-Dichlorobenzene	10.0	UD	4.80	10.0	20.0	ug/L
106-46-7	1,4-Dichlorobenzene	10.0	UD	5.40	10.0	20.0	ug/L
95-50-1	1,2-Dichlorobenzene	10.0	UD	3.80	10.0	20.0	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	15.0	UD	9.20	15.0	20.0	ug/L
120-82-1	1,2,4-Trichlorobenzene	10.0	UD	8.40	10.0	20.0	ug/L
87-61-6	1,2,3-Trichlorobenzene	15.0	UD	10.2	15.0	20.0	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	42.6		81 - 118		85%	SPK: 50
1868-53-7	Dibromofluoromethane	46.1		80 - 119		92%	SPK: 50
2037-26-5	Toluene-d8	50.3		89 - 112		101%	SPK: 50
460-00-4	4-Bromofluorobenzene	45.6		85 - 114		91%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	121000	5.55				
540-36-3	1,4-Difluorobenzene	216000	6.757				
3114-55-4	Chlorobenzene-d5	192000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	80400	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/13/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	RE115D1-20241213DL		SDG No.:	P5281	
Lab Sample ID:	P5281-06DL		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044455.D	20		12/20/24 16:19	VX122024

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/13/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	TT189D2-20241213		SDG No.:	P5281	
Lab Sample ID:	P5281-07		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044393.D	1		12/18/24 19:05	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.54	J	0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	0.54	J	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	21.4		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	1.20		0.26	0.75	1.00	ug/L
67-64-1	Acetone	1.90	J	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	J	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	2.80		0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
67-66-3	Chloroform	0.42	J	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	94.8		0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/13/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	TT189D2-20241213		SDG No.:	P5281	
Lab Sample ID:	P5281-07		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044393.D	1		12/18/24 19:05	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.39	J	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	44.0		81 - 118		88%	SPK: 50
1868-53-7	Dibromofluoromethane	46.0		80 - 119		92%	SPK: 50
2037-26-5	Toluene-d8	50.7		89 - 112		101%	SPK: 50
460-00-4	4-Bromofluorobenzene	46.8		85 - 114		94%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	121000	5.55				
540-36-3	1,4-Difluorobenzene	218000	6.757				
3114-55-4	Chlorobenzene-d5	190000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	76700	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/13/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	TT189D2-20241213		SDG No.:	P5281	
Lab Sample ID:	P5281-07		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044393.D	1		12/18/24 19:05	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/13/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	FB02-20241213		SDG No.:	P5281	
Lab Sample ID:	P5281-08		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044382.D	1		12/18/24 14:48	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	2.60	J	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.33	J	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	2.20	J	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.40	J	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.			Date Collected:	12/13/24		
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258			Date Received:	12/13/24		
Client Sample ID:	FB02-20241213			SDG No.:	P5281		
Lab Sample ID:	P5281-08			Matrix:	Water		
Analytical Method:	SW8260			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	DB-624UI	ID :	0.18	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044382.D	1		12/18/24 14:48	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	44.8		81 - 118		90%	SPK: 50
1868-53-7	Dibromofluoromethane	47.5		80 - 119		95%	SPK: 50
2037-26-5	Toluene-d8	51.1		89 - 112		102%	SPK: 50
460-00-4	4-Bromofluorobenzene	46.2		85 - 114		92%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	120000	5.55				
540-36-3	1,4-Difluorobenzene	218000	6.757				
3114-55-4	Chlorobenzene-d5	186000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	75100	12.024				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/13/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	FB02-20241213		SDG No.:	P5281	
Lab Sample ID:	P5281-08		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044382.D	1		12/18/24 14:48	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/13/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	TB		SDG No.:	P5281	
Lab Sample ID:	P5281-09		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044383.D	1		12/18/24 15:11	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	0.58	J	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	1.50	J	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.45	J	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/13/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	TB		SDG No.:	P5281	
Lab Sample ID:	P5281-09		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044383.D	1		12/18/24 15:11	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	43.7		81 - 118		87%	SPK: 50
1868-53-7	Dibromofluoromethane	46.4		80 - 119		93%	SPK: 50
2037-26-5	Toluene-d8	51.8		89 - 112		104%	SPK: 50
460-00-4	4-Bromofluorobenzene	45.3		85 - 114		91%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	117000	5.544				
540-36-3	1,4-Difluorobenzene	209000	6.757				
3114-55-4	Chlorobenzene-d5	182000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	71000	12.024				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/13/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/13/24	
Client Sample ID:	TB		SDG No.:	P5281	
Lab Sample ID:	P5281-09		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044383.D	1		12/18/24 15:11	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SDG No.: **P5281**

Client: **AECOM Technical Services, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P5281-01	RE114D1-20241212	1,2-Dichloroethane-d4	50	45.2	90	81	118
		Dibromofluoromethane	50	47.8	96	80	119
		Toluene-d8	50	51.2	102	89	112
		4-Bromofluorobenzene	50	48.8	98	85	114
P5281-01DL	RE114D1-20241212DL	1,2-Dichloroethane-d4	50	42.9	86	81	118
		Dibromofluoromethane	50	48.3	97	80	119
		Toluene-d8	50	50.7	101	89	112
		4-Bromofluorobenzene	50	46.3	93	85	114
P5281-02	BPOW1-7-20241212	1,2-Dichloroethane-d4	50	42.4	85	81	118
		Dibromofluoromethane	50	47.4	95	80	119
		Toluene-d8	50	50.5	101	89	112
		4-Bromofluorobenzene	50	43.8	88	85	114
P5281-03	BPOW1-8-20241212	1,2-Dichloroethane-d4	50	44.5	89	81	118
		Dibromofluoromethane	50	47.0	94	80	119
		Toluene-d8	50	51.6	103	89	112
		4-Bromofluorobenzene	50	47.8	96	85	114
P5281-04	TT101D2-20241212	1,2-Dichloroethane-d4	50	43.6	87	81	118
		Dibromofluoromethane	50	47.1	94	80	119
		Toluene-d8	50	52.2	104	89	112
		4-Bromofluorobenzene	50	47.4	95	85	114
P5281-05	RE117D1-20241213	1,2-Dichloroethane-d4	50	44.9	90	81	118
		Dibromofluoromethane	50	47.0	94	80	119
		Toluene-d8	50	51.1	102	89	112
		4-Bromofluorobenzene	50	45.8	92	85	114
P5281-06	RE115D1-20241213	1,2-Dichloroethane-d4	50	43.0	86	81	118
		Dibromofluoromethane	50	46.9	94	80	119
		Toluene-d8	50	50.6	101	89	112
		4-Bromofluorobenzene	50	46.0	92	85	114
P5281-06DL	RE115D1-20241213DL	1,2-Dichloroethane-d4	50	42.6	85	81	118
		Dibromofluoromethane	50	46.1	92	80	119
		Toluene-d8	50	50.3	101	89	112
		4-Bromofluorobenzene	50	45.6	91	85	114
P5281-07	TT189D2-20241213	1,2-Dichloroethane-d4	50	44.0	88	81	118
		Dibromofluoromethane	50	46.0	92	80	119
		Toluene-d8	50	50.7	101	89	112
		4-Bromofluorobenzene	50	46.8	94	85	114
P5281-08	FB02-20241213	1,2-Dichloroethane-d4	50	44.8	90	81	118
		Dibromofluoromethane	50	47.5	95	80	119
		Toluene-d8	50	51.1	102	89	112
		4-Bromofluorobenzene	50	46.2	92	85	114
P5281-09	TB	1,2-Dichloroethane-d4	50	43.7	87	81	118
		Dibromofluoromethane	50	46.4	93	80	119
		Toluene-d8	50	51.8	104	89	112
		4-Bromofluorobenzene	50	45.3	91	85	114
VX1218WBL01	VX1218WBL01	1,2-Dichloroethane-d4	50	45.6	91	81	118
		Dibromofluoromethane	50	47.1	94	80	119
		Toluene-d8	50	51.2	102	89	112
		4-Bromofluorobenzene	50	47.2	94	85	114
VX1218WBS01	VX1218WBS01	1,2-Dichloroethane-d4	50	47.2	94	81	118
		Dibromofluoromethane	50	50.5	101	80	119

Surrogate Summary

SDG No.: P5281

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
VX1218WBS01	VX1218WBS01	Toluene-d8	50	50.8	102	89	112
		4-Bromofluorobenzene	50	48.2	96	85	114
VX1218WBSD0	VX1218WBSD01	1,2-Dichloroethane-d4	50	46.8	94	81	118
		Dibromofluoromethane	50	49.7	99	80	119
		Toluene-d8	50	50.6	101	89	112
		4-Bromofluorobenzene	50	50.6	101	85	114
VX1220WBL01	VX1220WBL01	1,2-Dichloroethane-d4	50	44.1	88	81	118
		Dibromofluoromethane	50	47.5	95	80	119
		Toluene-d8	50	52.0	104	89	112
		4-Bromofluorobenzene	50	48.4	97	85	114
VX1220WBS01	VX1220WBS01	1,2-Dichloroethane-d4	50	42.7	85	81	118
		Dibromofluoromethane	50	48.5	97	80	119
		Toluene-d8	50	48.7	97	89	112
		4-Bromofluorobenzene	50	46.4	93	85	114

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5281

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX044378.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1218WBS01	Dichlorodifluoromethane	20	19.9	ug/L	100			32	152	
	Chloromethane	20	19.9	ug/L	100			50	139	
	Vinyl chloride	20	19.5	ug/L	98			58	137	
	Bromomethane	20	17.2	ug/L	86			53	141	
	Chloroethane	20	21.1	ug/L	106			60	138	
	Trichlorofluoromethane	20	15.8	ug/L	79			65	141	
	1,1,2-Trichlorotrifluoroethane	20	18.8	ug/L	94			70	136	
	1,1-Dichloroethene	20	19.7	ug/L	99			71	131	
	Acetone	100	80.4	ug/L	80			39	160	
	Carbon disulfide	20	19.1	ug/L	96			64	133	
	Methyl tert-butyl Ether	20	18.4	ug/L	92			71	124	
	Methyl Acetate	20	18.5	ug/L	93			56	136	
	Methylene Chloride	20	19.7	ug/L	99			74	124	
	trans-1,2-Dichloroethene	20	19.2	ug/L	96			75	124	
	1,1-Dichloroethane	20	19.6	ug/L	98			77	125	
	Cyclohexane	20	19.2	ug/L	96			71	130	
	2-Butanone	100	89.3	ug/L	89			56	143	
	Carbon Tetrachloride	20	18.4	ug/L	92			72	136	
	cis-1,2-Dichloroethene	20	18.9	ug/L	95			78	123	
	Bromochloromethane	20	17.5	ug/L	88			78	123	
	Chloroform	20	18.7	ug/L	94			79	124	
	1,1,1-Trichloroethane	20	18.7	ug/L	94			74	131	
	Methylcyclohexane	20	19.4	ug/L	97			72	132	
	Benzene	20	20.1	ug/L	101			79	120	
	1,2-Dichloroethane	20	18.9	ug/L	95			73	128	
	Trichloroethene	20	19.1	ug/L	96			79	123	
	1,2-Dichloropropane	20	19.3	ug/L	97			78	122	
	Bromodichloromethane	20	18.8	ug/L	94			79	125	
	4-Methyl-2-Pentanone	100	94.8	ug/L	95			67	130	
	Toluene	20	19.4	ug/L	97			80	121	
	t-1,3-Dichloropropene	20	18.8	ug/L	94			73	127	
	cis-1,3-Dichloropropene	20	19.5	ug/L	98			75	124	
	1,1,2-Trichloroethane	20	19.3	ug/L	97			80	119	
	2-Hexanone	100	93.3	ug/L	93			57	139	
	Dibromochloromethane	20	18.3	ug/L	92			74	126	
	1,2-Dibromoethane	20	19.9	ug/L	100			77	121	
	Tetrachloroethene	20	19.6	ug/L	98			74	129	
	Chlorobenzene	20	19.5	ug/L	98			82	118	
	Ethyl Benzene	20	19.3	ug/L	97			79	121	
	m/p-Xylenes	40	39.3	ug/L	98			80	121	
	o-Xylene	20	19.2	ug/L	96			78	122	
	Styrene	20	19.4	ug/L	97			78	123	
	Bromoform	20	18.7	ug/L	94			66	130	
	Isopropylbenzene	20	20.4	ug/L	102			72	131	
	1,1,2,2-Tetrachloroethane	20	19.9	ug/L	100			71	121	
	1,3-Dichlorobenzene	20	19.8	ug/L	99			80	119	
	1,4-Dichlorobenzene	20	19.5	ug/L	98			79	118	
	1,2-Dichlorobenzene	20	19.0	ug/L	95			80	119	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5281

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX044378.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VX1218WBS01	1,2-Dibromo-3-Chloropropane	20	17.5	ug/L	88			62	128	
	1,2,4-Trichlorobenzene	20	19.2	ug/L	96			69	130	
	1,2,3-Trichlorobenzene	20	19.7	ug/L	99			69	129	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5281

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX044379.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1218WBSD01	Dichlorodifluoromethane	20	19.8	ug/L	99	1		32	152	20
	Chloromethane	20	19.1	ug/L	96	4		50	139	20
	Vinyl chloride	20	18.8	ug/L	94	4		58	137	20
	Bromomethane	20	17.1	ug/L	86	0		53	141	20
	Chloroethane	20	21.6	ug/L	108	2		60	138	20
	Trichlorofluoromethane	20	15.8	ug/L	79	0		65	141	20
	1,1,2-Trichlorotrifluoroethane	20	19.2	ug/L	96	2		70	136	20
	1,1-Dichloroethene	20	19.6	ug/L	98	1		71	131	20
	Acetone	100	79.6	ug/L	80	0		39	160	20
	Carbon disulfide	20	18.8	ug/L	94	2		64	133	20
	Methyl tert-butyl Ether	20	18.9	ug/L	95	3		71	124	20
	Methyl Acetate	20	18.7	ug/L	94	1		56	136	20
	Methylene Chloride	20	18.8	ug/L	94	5		74	124	20
	trans-1,2-Dichloroethene	20	18.5	ug/L	93	3		75	124	20
	1,1-Dichloroethane	20	18.7	ug/L	94	4		77	125	20
	Cyclohexane	20	19.0	ug/L	95	1		71	130	20
	2-Butanone	100	90.9	ug/L	91	2		56	143	20
	Carbon Tetrachloride	20	18.7	ug/L	94	2		72	136	20
	cis-1,2-Dichloroethene	20	18.2	ug/L	91	4		78	123	20
	Bromochloromethane	20	17.4	ug/L	87	1		78	123	20
	Chloroform	20	18.2	ug/L	91	3		79	124	20
	1,1,1-Trichloroethane	20	18.0	ug/L	90	4		74	131	20
	Methylcyclohexane	20	21.5	ug/L	108	11		72	132	20
	Benzene	20	20.2	ug/L	101	0		79	120	20
	1,2-Dichloroethane	20	19.8	ug/L	99	4		73	128	20
	Trichloroethene	20	19.4	ug/L	97	1		79	123	20
	1,2-Dichloropropane	20	19.3	ug/L	97	0		78	122	20
	Bromodichloromethane	20	19.6	ug/L	98	4		79	125	20
	4-Methyl-2-Pentanone	100	97.7	ug/L	98	3		67	130	20
	Toluene	20	19.8	ug/L	99	2		80	121	20
	t-1,3-Dichloropropene	20	19.8	ug/L	99	5		73	127	20
	cis-1,3-Dichloropropene	20	20.4	ug/L	102	4		75	124	20
	1,1,2-Trichloroethane	20	20.2	ug/L	101	4		80	119	20
	2-Hexanone	100	98.6	ug/L	99	6		57	139	20
	Dibromochloromethane	20	18.9	ug/L	95	3		74	126	20
	1,2-Dibromoethane	20	20.5	ug/L	103	3		77	121	20
	Tetrachloroethene	20	19.2	ug/L	96	2		74	129	20
	Chlorobenzene	20	19.6	ug/L	98	0		82	118	20
	Ethyl Benzene	20	19.2	ug/L	96	1		79	121	20
	m/p-Xylenes	40	39.5	ug/L	99	1		80	121	20
	o-Xylene	20	19.3	ug/L	97	1		78	122	20
	Styrene	20	19.7	ug/L	99	2		78	123	20
	Bromoform	20	18.9	ug/L	95	1		66	130	20
	Isopropylbenzene	20	19.8	ug/L	99	3		72	131	20
	1,1,2,2-Tetrachloroethane	20	19.5	ug/L	98	2		71	121	20
	1,3-Dichlorobenzene	20	19.9	ug/L	100	1		80	119	20
	1,4-Dichlorobenzene	20	19.1	ug/L	96	2		79	118	20
	1,2-Dichlorobenzene	20	19.1	ug/L	96	1		80	119	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5281

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX044379.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1218WBSD01	1,2-Dibromo-3-Chloropropane	20	16.4	ug/L	82	7		62	128	20
	1,2,4-Trichlorobenzene	20	19.3	ug/L	97	1		69	130	20
	1,2,3-Trichlorobenzene	20	20.4	ug/L	102	3		69	129	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5281

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX044441.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1220WBS01	Dichlorodifluoromethane	20	18.5	ug/L	93			32	152	
	Chloromethane	20	18.7	ug/L	94			50	139	
	Vinyl chloride	20	17.9	ug/L	90			58	137	
	Bromomethane	20	18.8	ug/L	94			53	141	
	Chloroethane	20	19.8	ug/L	99			60	138	
	Trichlorofluoromethane	20	15.3	ug/L	77			65	141	
	1,1,2-Trichlorotrifluoroethane	20	18.5	ug/L	93			70	136	
	1,1-Dichloroethene	20	18.9	ug/L	95			71	131	
	Acetone	100	75.7	ug/L	76			39	160	
	Carbon disulfide	20	16.4	ug/L	82			64	133	
	Methyl tert-butyl Ether	20	17.6	ug/L	88			71	124	
	Methyl Acetate	20	17.7	ug/L	89			56	136	
	Methylene Chloride	20	19.1	ug/L	96			74	124	
	trans-1,2-Dichloroethene	20	18.7	ug/L	94			75	124	
	1,1-Dichloroethane	20	18.9	ug/L	95			77	125	
	Cyclohexane	20	17.6	ug/L	88			71	130	
	2-Butanone	100	83.3	ug/L	83			56	143	
	Carbon Tetrachloride	20	18.9	ug/L	95			72	136	
	cis-1,2-Dichloroethene	20	18.0	ug/L	90			78	123	
	Bromochloromethane	20	15.8	ug/L	79			78	123	
	Chloroform	20	17.7	ug/L	89			79	124	
	1,1,1-Trichloroethane	20	17.5	ug/L	88			74	131	
	Methylcyclohexane	20	19.3	ug/L	97			72	132	
	Benzene	20	20.4	ug/L	102			79	120	
	1,2-Dichloroethane	20	19.2	ug/L	96			73	128	
	Trichloroethene	20	19.9	ug/L	100			79	123	
	1,2-Dichloropropane	20	19.1	ug/L	96			78	122	
	Bromodichloromethane	20	19.1	ug/L	96			79	125	
	4-Methyl-2-Pentanone	100	92.4	ug/L	92			67	130	
	Toluene	20	19.5	ug/L	98			80	121	
	t-1,3-Dichloropropene	20	17.9	ug/L	90			73	127	
	cis-1,3-Dichloropropene	20	19.2	ug/L	96			75	124	
	1,1,2-Trichloroethane	20	20.8	ug/L	104			80	119	
	2-Hexanone	100	90.4	ug/L	90			57	139	
	Dibromochloromethane	20	18.8	ug/L	94			74	126	
	1,2-Dibromoethane	20	20.2	ug/L	101			77	121	
	Tetrachloroethene	20	20.8	ug/L	104			74	129	
	Chlorobenzene	20	19.9	ug/L	100			82	118	
	Ethyl Benzene	20	19.2	ug/L	96			79	121	
	m/p-Xylenes	40	39.8	ug/L	100			80	121	
	o-Xylene	20	19.8	ug/L	99			78	122	
	Styrene	20	19.8	ug/L	99			78	123	
	Bromoform	20	19.3	ug/L	97			66	130	
	Isopropylbenzene	20	19.9	ug/L	100			72	131	
	1,1,2,2-Tetrachloroethane	20	19.6	ug/L	98			71	121	
	1,3-Dichlorobenzene	20	20.1	ug/L	101			80	119	
	1,4-Dichlorobenzene	20	20.0	ug/L	100			79	118	
	1,2-Dichlorobenzene	20	19.8	ug/L	99			80	119	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5281

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX044441.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1220WBS01	1,2-Dibromo-3-Chloropropane	20	16.5	ug/L	83			62	128	
	1,2,4-Trichlorobenzene	20	19.5	ug/L	98			69	130	
	1,2,3-Trichlorobenzene	20	19.4	ug/L	97			69	129	

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VX1218WBL01

Lab Name: CHEMTECH

Contract: AECO15

Lab Code: CHEM Case No.: P5281

SAS No.: P5281 SDG NO.: P5281

Lab File ID: VX044377.D

Lab Sample ID: VX1218WBL01

Date Analyzed: 12/18/2024

Time Analyzed: 12:48

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX1218WBS01	VX1218WBS01	VX044378.D	12/18/2024
VX1218WBSD01	VX1218WBSD01	VX044379.D	12/18/2024
TT101D2-20241212	P5281-04	VX044381.D	12/18/2024
FB02-20241213	P5281-08	VX044382.D	12/18/2024
TB	P5281-09	VX044383.D	12/18/2024
RE114D1-20241212	P5281-01	VX044388.D	12/18/2024
BPOW1-8-20241212	P5281-03	VX044390.D	12/18/2024
RE117D1-20241213	P5281-05	VX044391.D	12/18/2024
RE115D1-20241213	P5281-06	VX044392.D	12/18/2024
TT189D2-20241213	P5281-07	VX044393.D	12/18/2024

COMMENTS: _____

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VX1220WBL01

Lab Name: CHEMTECH

Contract: AECO15

Lab Code: CHEM Case No.: P5281

SAS No.: P5281 SDG NO.: P5281

Lab File ID: VX044440.D

Lab Sample ID: VX1220WBL01

Date Analyzed: 12/20/2024

Time Analyzed: 10:23

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX1220WBS01	VX1220WBS01	VX044441.D	12/20/2024
RE114D1-20241212DL	P5281-01DL	VX044444.D	12/20/2024
BPOW1-7-20241212	P5281-02	VX044451.D	12/20/2024
RE115D1-20241213DL	P5281-06DL	VX044455.D	12/20/2024

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: P5281 SAS No.: P5281 SDG NO.: P5281
Lab File ID: VX044218.D BFB Injection Date: 12/11/2024
Instrument ID: MSVOA_X BFB Injection Time: 10:13
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	21.7
75	30.0 - 60.0% of mass 95	56.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.7
173	Less than 2.0% of mass 174	0.9 (1.2) 1
174	50.0 - 100.0% of mass 95	71.1
175	5.0 - 9.0% of mass 174	5.2 (7.3) 1
176	95.0 - 101.0% of mass 174	68.7 (96.6) 1
177	5.0 - 9.0% of mass 176	4.6 (6.8) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC001	VSTDIC001	VX044219.D	12/11/2024	10:41
VSTDIC005	VSTDIC005	VX044220.D	12/11/2024	11:27
VSTDIC020	VSTDIC020	VX044221.D	12/11/2024	11:50
VSTDIC050	VSTDIC050	VX044222.D	12/11/2024	12:14
VSTDIC100	VSTDIC100	VX044223.D	12/11/2024	12:37
VSTDIC150	VSTDIC150	VX044224.D	12/11/2024	13:00

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: P5281 SAS No.: P5281 SDG NO.: P5281
Lab File ID: VX044374.D BFB Injection Date: 12/18/2024
Instrument ID: MSVOA_X BFB Injection Time: 08:15
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	20.4
75	30.0 - 60.0% of mass 95	55.1
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.5
173	Less than 2.0% of mass 174	0.6 (0.7) 1
174	50.0 - 100.0% of mass 95	79.8
175	5.0 - 9.0% of mass 174	5.7 (7.1) 1
176	95.0 - 101.0% of mass 174	76.6 (96) 1
177	5.0 - 9.0% of mass 176	4.6 (6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX044375.D	12/18/2024	11:57
VX1218WBL01	VX1218WBL01	VX044377.D	12/18/2024	12:48
VX1218WBS01	VX1218WBS01	VX044378.D	12/18/2024	13:11
VX1218WBSD01	VX1218WBSD01	VX044379.D	12/18/2024	13:38
TT101D2-20241212	P5281-04	VX044381.D	12/18/2024	14:24
FB02-20241213	P5281-08	VX044382.D	12/18/2024	14:48
TB	P5281-09	VX044383.D	12/18/2024	15:11
RE114D1-20241212	P5281-01	VX044388.D	12/18/2024	17:08
BPOW1-8-20241212	P5281-03	VX044390.D	12/18/2024	17:55
RE117D1-20241213	P5281-05	VX044391.D	12/18/2024	18:18
RE115D1-20241213	P5281-06	VX044392.D	12/18/2024	18:42
TT189D2-20241213	P5281-07	VX044393.D	12/18/2024	19:05
VSTDCCC050EC	VSTDCCC050	VX044395.D	12/18/2024	19:52

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: P5281 SAS No.: P5281 SDG NO.: P5281
Lab File ID: VX044437.D BFB Injection Date: 12/20/2024
Instrument ID: MSVOA_X BFB Injection Time: 08:49
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	20.4
75	30.0 - 60.0% of mass 95	52.1
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7.8
173	Less than 2.0% of mass 174	0.9 (1.2) 1
174	50.0 - 100.0% of mass 95	77.7
175	5.0 - 9.0% of mass 174	5.6 (7.2) 1
176	95.0 - 101.0% of mass 174	76.6 (98.5) 1
177	5.0 - 9.0% of mass 176	4.7 (6.1) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX044438.D	12/20/2024	09:16
VX1220WBL01	VX1220WBL01	VX044440.D	12/20/2024	10:23
VX1220WBS01	VX1220WBS01	VX044441.D	12/20/2024	10:46
RE114D1-20241212DL	P5281-01DL	VX044444.D	12/20/2024	12:03
BPOW1-7-20241212	P5281-02	VX044451.D	12/20/2024	14:46
RE115D1-20241213DL	P5281-06DL	VX044455.D	12/20/2024	16:19
VSTDCCC050EC	VSTDCCC050	VX044464.D	12/20/2024	19:49

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: P5281 SAS No.: P5281 SDG NO.: P5281
Lab File ID: VX044375.D Date Analyzed: 12/18/2024
Instrument ID: MSVOA_X Time Analyzed: 11:57
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	145222	5.54	244425	6.75	212426	10.05
UPPER LIMIT	290444	6.044	488850	7.251	424852	10.549
LOWER LIMIT	72611	5.044	122213	6.251	106213	9.549
EPA SAMPLE NO.						
RE114D1-20241212	115604	5.54	209270	6.76	182535	10.06
BPOW1-8-20241212	114414	5.54	208386	6.76	184180	10.06
TT101D2-20241212	120810	5.55	214177	6.76	186747	10.06
RE117D1-20241213	111808	5.55	202274	6.76	172810	10.06
RE115D1-20241213	115938	5.55	206968	6.76	178386	10.06
TT189D2-20241213	120659	5.55	218055	6.76	189518	10.06
FB02-20241213	120337	5.55	217882	6.76	185612	10.06
TB	116899	5.54	209253	6.76	182054	10.06
VX1218WBL01	123193	5.55	230004	6.76	201476	10.06
VX1218WBS01	152020	5.54	272118	6.76	233700	10.06
VX1218WBSD01	147797	5.54	256001	6.76	224705	10.06

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: P5281 SAS No.: P5281 SDG NO.: P5281
Lab File ID: VX044375.D Date Analyzed: 12/18/2024
Instrument ID: MSVOA_X Time Analyzed: 11:57
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	100489	12.018				
UPPER LIMIT	200978	12.518				
LOWER LIMIT	50244.5	11.518				
EPA SAMPLE NO.						
RE114D1-20241212	75803	12.02				
BPOW1-8-20241212	77822	12.02				
TT101D2-20241212	76850	12.02				
RE117D1-20241213	67944	12.02				
RE115D1-20241213	70750	12.02				
TT189D2-20241213	76670	12.02				
FB02-20241213	75075	12.02				
TB	70980	12.02				
VX1218WBL01	82810	12.02				
VX1218WBS01	101950	12.02				
VX1218WBSD01	101474	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: P5281 SAS No.: P5281 SDG NO.: P5281
Lab File ID: VX044438.D Date Analyzed: 12/20/2024
Instrument ID: MSVOA_X Time Analyzed: 09:16
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	136445	5.54	226018	6.76	197543	10.06
UPPER LIMIT	272890	6.044	452036	7.257	395086	10.555
LOWER LIMIT	68222.5	5.044	113009	6.257	98771.5	9.555
EPA SAMPLE NO.						
RE114D1-20241212DL	118538	5.55	211612	6.76	184574	10.06
BPOW1-7-20241212	117082	5.55	205938	6.76	173219	10.06
RE115D1-20241213DL	121258	5.55	216230	6.76	191670	10.06
VX1220WBL01	120864	5.55	218417	6.76	193836	10.05
VX1220WBS01	146515	5.55	249173	6.76	212384	10.05

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: P5281 SAS No.: P5281 SDG NO.: P5281
 Lab File ID: VX044438.D Date Analyzed: 12/20/2024
 Instrument ID: MSVOA_X Time Analyzed: 09:16
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	95339	12.018				
UPPER LIMIT	190678	12.518				
LOWER LIMIT	47669.5	11.518				
EPA SAMPLE NO.						
RE114D1-20241212DL	74955	12.02				
BPOW1-7-20241212	66450	12.02				
RE115D1-20241213DL	80370	12.02				
VX1220WBL01	81005	12.02				
VX1220WBS01	96885	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX1218WBL01		SDG No.:	P5281
Lab Sample ID:	VX1218WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044377.D	1		12/18/24 12:48	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX1218WBL01		SDG No.:	P5281
Lab Sample ID:	VX1218WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044377.D	1		12/18/24 12:48	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	45.6		81 - 118		91%	SPK: 50
1868-53-7	Dibromofluoromethane	47.1		80 - 119		94%	SPK: 50
2037-26-5	Toluene-d8	51.2		89 - 112		102%	SPK: 50
460-00-4	4-Bromofluorobenzene	47.2		85 - 114		94%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	123000	5.55				
540-36-3	1,4-Difluorobenzene	230000	6.757				
3114-55-4	Chlorobenzene-d5	201000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	82800	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX1218WBL01		SDG No.:	P5281
Lab Sample ID:	VX1218WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044377.D	1		12/18/24 12:48	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX1220WBL01		SDG No.:	P5281
Lab Sample ID:	VX1220WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044440.D	1		12/20/24 10:23	VX122024

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX1220WBL01		SDG No.:	P5281
Lab Sample ID:	VX1220WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044440.D	1		12/20/24 10:23	VX122024

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	44.1		81 - 118		88%	SPK: 50
1868-53-7	Dibromofluoromethane	47.5		80 - 119		95%	SPK: 50
2037-26-5	Toluene-d8	52.1		89 - 112		104%	SPK: 50
460-00-4	4-Bromofluorobenzene	48.4		85 - 114		97%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	121000	5.55				
540-36-3	1,4-Difluorobenzene	218000	6.757				
3114-55-4	Chlorobenzene-d5	194000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	81000	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX1220WBL01		SDG No.:	P5281
Lab Sample ID:	VX1220WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044440.D	1		12/20/24 10:23	VX122024

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX1218WBS01		SDG No.:	P5281
Lab Sample ID:	VX1218WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044378.D	1		12/18/24 13:11	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	19.9		0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	19.9		0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	19.5		0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	17.2		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	21.1		0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	15.8		0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	18.8		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	19.7		0.26	0.75	1.00	ug/L
67-64-1	Acetone	80.4		1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	19.1		0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	18.4		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	18.5		0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	19.7		0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	19.2		0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	19.6		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	19.2		1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	89.3		1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	18.4		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	18.9		0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	17.5		0.18	0.50	1.00	ug/L
67-66-3	Chloroform	18.7		0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	18.7		0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	19.4		0.19	0.50	1.00	ug/L
71-43-2	Benzene	20.1		0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	18.9		0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	19.1		0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	19.3		0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	18.8		0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	94.8		0.75	2.50	5.00	ug/L
108-88-3	Toluene	19.4		0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX1218WBS01		SDG No.:	P5281
Lab Sample ID:	VX1218WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044378.D	1		12/18/24 13:11	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	18.8		0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	19.5		0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	19.3		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	93.3		1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	18.3		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	19.9		0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	19.6		0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	19.5		0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	19.3		0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	39.3		0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	19.2		0.14	0.50	1.00	ug/L
100-42-5	Styrene	19.4		0.16	0.50	1.00	ug/L
75-25-2	Bromoform	18.7		0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	20.4		0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	19.9		0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	19.8		0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	19.5		0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	19.0		0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	17.5		0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	19.2		0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	19.7		0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	47.2		81 - 118		94%	SPK: 50
1868-53-7	Dibromofluoromethane	50.5		80 - 119		101%	SPK: 50
2037-26-5	Toluene-d8	50.8		89 - 112		102%	SPK: 50
460-00-4	4-Bromofluorobenzene	48.2		85 - 114		96%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	152000	5.544				
540-36-3	1,4-Difluorobenzene	272000	6.757				
3114-55-4	Chlorobenzene-d5	234000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	102000	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX1218WBS01		SDG No.:	P5281
Lab Sample ID:	VX1218WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044378.D	1		12/18/24 13:11	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX1220WBS01		SDG No.:	P5281
Lab Sample ID:	VX1220WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044441.D	1		12/20/24 10:46	VX122024

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	18.5		0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	18.7		0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	17.9		0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	18.8		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	19.8		0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	15.3		0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	18.5		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	18.9		0.26	0.75	1.00	ug/L
67-64-1	Acetone	75.7		1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	16.4		0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	17.6		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	17.7		0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	19.1		0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	18.7		0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	18.9		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	17.6		1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	83.3		1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	18.9		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	18.0		0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	15.8		0.18	0.50	1.00	ug/L
67-66-3	Chloroform	17.7		0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	17.5		0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	19.3		0.19	0.50	1.00	ug/L
71-43-2	Benzene	20.4		0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	19.2		0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	19.9		0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	19.1		0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	19.1		0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	92.4		0.75	2.50	5.00	ug/L
108-88-3	Toluene	19.5		0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX1220WBS01		SDG No.:	P5281
Lab Sample ID:	VX1220WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044441.D	1		12/20/24 10:46	VX122024

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	17.9		0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	19.2		0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	20.8		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	90.4		1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	18.8		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	20.2		0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	20.8		0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	19.9		0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	19.2		0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	39.8		0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	19.8		0.14	0.50	1.00	ug/L
100-42-5	Styrene	19.8		0.16	0.50	1.00	ug/L
75-25-2	Bromoform	19.3		0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	19.9		0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	19.6		0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	20.1		0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	20.0		0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	19.8		0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	16.5		0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	19.5		0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	19.4		0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	42.7		81 - 118		85%	SPK: 50
1868-53-7	Dibromofluoromethane	48.5		80 - 119		97%	SPK: 50
2037-26-5	Toluene-d8	48.7		89 - 112		97%	SPK: 50
460-00-4	4-Bromofluorobenzene	46.4		85 - 114		93%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	147000	5.55				
540-36-3	1,4-Difluorobenzene	249000	6.757				
3114-55-4	Chlorobenzene-d5	212000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	96900	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX1220WBS01		SDG No.:	P5281
Lab Sample ID:	VX1220WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044441.D	1		12/20/24 10:46	VX122024

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX1218WBSD01		SDG No.:	P5281
Lab Sample ID:	VX1218WBSD01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044379.D	1		12/18/24 13:38	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	19.8		0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	19.1		0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	18.8		0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	17.1		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	21.6		0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	15.8		0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	19.2		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	19.6		0.26	0.75	1.00	ug/L
67-64-1	Acetone	79.6		1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	18.8		0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	18.9		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	18.7		0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	18.8		0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	18.5		0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	18.7		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	19.0		1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	90.9		1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	18.7		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	18.2		0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	17.4		0.18	0.50	1.00	ug/L
67-66-3	Chloroform	18.2		0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	18.0		0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	21.5		0.19	0.50	1.00	ug/L
71-43-2	Benzene	20.2		0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	19.8		0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	19.4		0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	19.3		0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	19.6		0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	97.7		0.75	2.50	5.00	ug/L
108-88-3	Toluene	19.8		0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX1218WBSD01		SDG No.:	P5281
Lab Sample ID:	VX1218WBSD01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044379.D	1		12/18/24 13:38	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	19.8		0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	20.4		0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	20.2		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	98.6		1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	18.9		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	20.5		0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	19.2		0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	19.6		0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	19.2		0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	39.5		0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	19.3		0.14	0.50	1.00	ug/L
100-42-5	Styrene	19.7		0.16	0.50	1.00	ug/L
75-25-2	Bromoform	18.9		0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	19.8		0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	19.5		0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	19.9		0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	19.1		0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	19.1		0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	16.4		0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	19.3		0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	20.4		0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	46.8		81 - 118		94%	SPK: 50
1868-53-7	Dibromofluoromethane	49.7		80 - 119		99%	SPK: 50
2037-26-5	Toluene-d8	50.6		89 - 112		101%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.6		85 - 114		101%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	148000	5.544				
540-36-3	1,4-Difluorobenzene	256000	6.757				
3114-55-4	Chlorobenzene-d5	225000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	101000	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX1218WBSD01		SDG No.:	P5281
Lab Sample ID:	VX1218WBSD01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044379.D	1		12/18/24 13:38	VX121824

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name:	<u>CHEMTECH</u>	Contract:	<u>AECO15</u>
Lab Code:	<u>CHEM</u>	Case No.:	<u>P5281</u>
Instrument ID:	<u>MSVOA_X</u>	SAS No.:	<u>P5281</u>
Heated Purge: (Y/N)	<u>N</u>	SDG No.:	<u>P5281</u>
GC Column:	<u>DB-624UI</u>	Calibration Date(s):	<u>12/11/2024</u>
ID:	<u>0.18 (mm)</u>	Calibration Time(s):	<u>10:41</u>
			<u>13:00</u>

LAB FILE ID:	RRF001 = VX044219.D	RRF005 = VX044220.D	RRF020 = VX044221.D	RRF050 = VX044222.D	RRF100 = VX044223.D	RRF150 = VX044224.D		
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
Dichlorodifluoromethane	0.877	0.749	0.721	0.788	0.745	0.759	0.773	7.2
Chloromethane	0.743	0.747	0.735	0.756	0.715	0.740	0.739	1.8
Vinyl Chloride	0.805	0.790	0.782	0.786	0.722	0.737	0.770	4.3
Bromomethane		0.590	0.513	0.551	0.529	0.548	0.546	5.3
Chloroethane	0.464	0.520	0.515	0.532	0.422	0.426	0.480	10.2
Trichlorofluoromethane	1.582	1.519	1.514	1.541	1.478	1.414	1.508	3.8
1,1,2-Trichlorotrifluoroethane	0.633	0.628	0.592	0.615	0.573	0.620	0.610	3.8
1,1-Dichloroethene	0.512	0.560	0.544	0.596	0.554	0.596	0.560	5.7
Acetone	0.359	0.353	0.331	0.381	0.348	0.372	0.357	5
Carbon Disulfide	0.841	0.872	0.937	1.150	1.230	1.362	1.065	20
Methyl tert-butyl Ether	2.054	2.188	2.127	2.305	2.170	2.279	2.187	4.3
Methyl Acetate	0.965	0.889	0.883	0.953	0.888	0.951	0.922	4.2
Methylene Chloride	0.679	0.678	0.666	0.689	0.648	0.668	0.671	2.1
trans-1,2-Dichloroethene	0.600	0.556	0.577	0.619	0.595	0.626	0.596	4.4
1,1-Dichloroethane	1.071	1.173	1.154	1.244	1.171	1.218	1.172	5.1
Cyclohexane		0.931	0.963	0.977	0.922	0.938	0.946	2.4
2-Butanone	0.449	0.491	0.510	0.548	0.509	0.541	0.508	7.1
Carbon Tetrachloride	0.484	0.443	0.432	0.501	0.502	0.529	0.482	7.8
cis-1,2-Dichloroethene	0.809	0.737	0.728	0.772	0.731	0.762	0.756	4.1
Bromochloromethane	0.578	0.577	0.576	0.592	0.556	0.568	0.575	2.1
Chloroform	1.389	1.312	1.272	1.343	1.272	1.310	1.316	3.4
1,1,1-Trichloroethane	0.993	1.019	1.077	1.176	1.119	1.173	1.093	7.1
Methylcyclohexane	0.514	0.503	0.527	0.555	0.550	0.562	0.535	4.5
Benzene	1.353	1.353	1.332	1.410	1.353	1.353	1.359	1.9
1,2-Dichloroethane	0.554	0.557	0.545	0.584	0.556	0.564	0.560	2.4
Trichloroethene	0.356	0.325	0.325	0.349	0.337	0.342	0.339	3.8
1,2-Dichloropropane	0.368	0.345	0.330	0.357	0.341	0.342	0.347	3.8
Bromodichloromethane	0.380	0.411	0.436	0.508	0.518	0.538	0.465	13.9
4-Methyl-2-Pentanone	0.506	0.551	0.547	0.590	0.561	0.582	0.556	5.4
Toluene	0.908	0.830	0.816	0.877	0.841	0.849	0.853	4

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: P5281 SAS No.: P5281 SDG No.: P5281

Instrument ID: MSVOA_X Calibration Date(s): 12/11/2024 12/11/2024

Heated Purge: (Y/N) N Calibration Time(s): 10:41 13:00

GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:								
RRF001 = VX044219.D			RRF005 = VX044220.D			RRF020 = VX044221.D		
RRF050 = VX044222.D			RRF100 = VX044223.D			RRF150 = VX044224.D		
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
t-1,3-Dichloropropene	0.338	0.357	0.439	0.514	0.526	0.558	0.455	20.3
cis-1,3-Dichloropropene	0.404	0.441	0.488	0.567	0.558	0.583	0.507	14.6
1,1,2-Trichloroethane	0.337	0.332	0.335	0.358	0.334	0.340	0.339	2.8
2-Hexanone	0.323	0.386	0.408	0.439	0.420	0.438	0.403	10.8
Dibromochloromethane	0.254	0.248	0.306	0.370	0.378	0.398	0.326	20.2
1,2-Dibromoethane	0.303	0.322	0.344	0.369	0.364	0.370	0.345	8.1
Tetrachloroethene	0.359	0.354	0.317	0.321	0.312	0.327	0.332	6
Chlorobenzene	1.130	1.068	1.039	1.088	1.036	1.067	1.071	3.3
Ethyl Benzene	1.853	1.806	1.816	1.927	1.831	1.875	1.851	2.4
m/p-Xylenes	0.653	0.660	0.665	0.707	0.683	0.707	0.679	3.5
o-Xylene	0.686	0.683	0.685	0.703	0.671	0.697	0.687	1.6
Styrene	1.007	1.019	1.099	1.162	1.148	1.177	1.102	6.7
Bromoform	0.137	0.173	0.194	0.247	0.269	0.297	0.219	28
Isopropylbenzene	3.665	3.976	3.920	3.912	3.747	3.747	3.828	3.2
1,1,2,2-Tetrachloroethane	1.410	1.350	1.332	1.336	1.256	1.293	1.329	3.9
1,3-Dichlorobenzene	1.519	1.634	1.640	1.683	1.625	1.628	1.621	3.3
1,4-Dichlorobenzene	1.811	1.640	1.632	1.672	1.621	1.623	1.667	4.4
1,2-Dichlorobenzene	1.745	1.666	1.668	1.726	1.654	1.643	1.684	2.5
1,2-Dibromo-3-Chloropropane	0.222	0.226	0.241	0.273	0.279	0.306	0.258	13
1,2,4-Trichlorobenzene	0.770	0.827	0.891	0.978	1.013	1.058	0.923	12.2
1,2,3-Trichlorobenzene	0.854	0.885	0.984	1.019	1.055	1.077	0.979	9.3
1,2-Dichloroethane-d4		0.934	0.778	0.921	0.862	0.888	0.877	7.1
Dibromofluoromethane		0.353	0.298	0.367	0.360	0.354	0.346	8
Toluene-d8		1.148	1.010	1.251	1.229	1.203	1.168	8.3
4-Bromofluorobenzene		0.390	0.351	0.442	0.432	0.426	0.408	9.2

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: P5281 SAS No.: P5281 SDG No.: P5281

Instrument ID: MSVOA_X Calibration Date/Time: 12/18/2024 11:57

Lab File ID: VX044375.D Init. Calib. Date(s): 12/11/2024 12/11/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:41 13:00

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.773	0.800		3.49	20
Chloromethane	0.739	0.720	0.1	-2.57	20
Vinyl Chloride	0.770	0.743		-3.51	20
Bromomethane	0.546	0.475		-13	20
Chloroethane	0.480	0.494		2.92	20
Trichlorofluoromethane	1.508	1.275		-15.45	20
1,1,2-Trichlorotrifluoroethane	0.610	0.613		0.49	20
1,1-Dichloroethene	0.560	0.566		1.07	20
Acetone	0.357	0.298		-16.53	20
Carbon Disulfide	1.065	1.209		13.52	20
Methyl tert-butyl Ether	2.187	2.034		-7	20
Methyl Acetate	0.922	0.830		-9.98	20
Methylene Chloride	0.671	0.643		-4.17	20
trans-1,2-Dichloroethene	0.596	0.594		-0.34	20
1,1-Dichloroethane	1.172	1.124	0.1	-4.1	20
Cyclohexane	0.946	0.956		1.06	20
2-Butanone	0.508	0.450		-11.42	20
Carbon Tetrachloride	0.482	0.517		7.26	20
cis-1,2-Dichloroethene	0.756	0.723		-4.36	20
Bromochloromethane	0.575	0.531		-7.65	20
Chloroform	1.316	1.196		-9.12	20
1,1,1-Trichloroethane	1.093	1.072		-1.92	20
Methylcyclohexane	0.535	0.624		16.64	20
Benzene	1.359	1.428		5.08	20
1,2-Dichloroethane	0.560	0.555		-0.89	20
Trichloroethene	0.339	0.364		7.38	20
1,2-Dichloropropane	0.347	0.356		2.59	20
Bromodichloromethane	0.465	0.508		9.25	20
4-Methyl-2-Pentanone	0.556	0.546		-1.8	20
Toluene	0.853	0.900		5.51	20
t-1,3-Dichloropropene	0.455	0.521		14.51	20
cis-1,3-Dichloropropene	0.507	0.571		12.62	20
1,1,2-Trichloroethane	0.339	0.351		3.54	20
2-Hexanone	0.403	0.402		-0.25	20
Dibromochloromethane	0.326	0.365		11.96	20
1,2-Dibromoethane	0.345	0.365		5.8	20
Tetrachloroethene	0.332	0.358		7.83	20
Chlorobenzene	1.071	1.111	0.3	3.73	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: P5281 SAS No.: P5281 SDG No.: P5281

Instrument ID: MSVOA_X Calibration Date/Time: 12/18/2024 11:57

Lab File ID: VX044375.D Init. Calib. Date(s): 12/11/2024 12/11/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:41 13:00

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.851	1.955		5.62	20
m/p-Xylenes	0.679	0.737		8.54	20
o-Xylene	0.687	0.731		6.41	20
Styrene	1.102	1.210		9.8	20
Bromoform	0.219	0.261	0.1	19.18	20
Isopropylbenzene	3.828	4.011		4.78	20
1,1,2,2-Tetrachloroethane	1.329	1.294	0.3	-2.63	20
1,3-Dichlorobenzene	1.621	1.757		8.39	20
1,4-Dichlorobenzene	1.667	1.734		4.02	20
1,2-Dichlorobenzene	1.684	1.744		3.56	20
1,2-Dibromo-3-Chloropropane	0.258	0.244		-5.43	20
1,2,4-Trichlorobenzene	0.923	1.075		16.47	20
1,2,3-Trichlorobenzene	0.979	1.107		13.07	20
1,2-Dichloroethane-d4	0.877	0.742		-15.39	20
Dibromofluoromethane	0.346	0.340		-1.73	20
Toluene-d8	1.168	1.164		-0.34	20
4-Bromofluorobenzene	0.408	0.413		1.23	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: P5281 SAS No.: P5281 SDG No.: P5281

Instrument ID: MSVOA_X Calibration Date/Time: 12/18/2024 19:52

Lab File ID: VX044395.D Init. Calib. Date(s): 12/11/2024 12/11/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:41 13:00

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.773	0.785		1.55	50
Chloromethane	0.739	0.805	0.1	8.93	50
Vinyl Chloride	0.770	0.803		4.29	50
Bromomethane	0.546	0.526		-3.66	50
Chloroethane	0.480	0.557		16.04	50
Trichlorofluoromethane	1.508	1.236		-18.04	50
1,1,2-Trichlorotrifluoroethane	0.610	0.577		-5.41	50
1,1-Dichloroethene	0.560	0.579		3.39	50
Acetone	0.357	0.306		-14.29	50
Carbon Disulfide	1.065	1.056		-0.85	50
Methyl tert-butyl Ether	2.187	2.164		-1.05	50
Methyl Acetate	0.922	0.912		-1.09	50
Methylene Chloride	0.671	0.695		3.58	50
trans-1,2-Dichloroethene	0.596	0.603		1.17	50
1,1-Dichloroethane	1.172	1.201	0.1	2.47	50
Cyclohexane	0.946	0.875		-7.51	50
2-Butanone	0.508	0.497		-2.16	50
Carbon Tetrachloride	0.482	0.466		-3.32	50
cis-1,2-Dichloroethene	0.756	0.763		0.93	50
Bromochloromethane	0.575	0.561		-2.43	50
Chloroform	1.316	1.284		-2.43	50
1,1,1-Trichloroethane	1.093	1.060		-3.02	50
Methylcyclohexane	0.535	0.528		-1.31	50
Benzene	1.359	1.397		2.8	50
1,2-Dichloroethane	0.560	0.554		-1.07	50
Trichloroethene	0.339	0.347		2.36	50
1,2-Dichloropropane	0.347	0.349		0.58	50
Bromodichloromethane	0.465	0.496		6.67	50
4-Methyl-2-Pentanone	0.556	0.558		0.36	50
Toluene	0.853	0.833		-2.35	50
t-1,3-Dichloropropene	0.455	0.470		3.3	50
cis-1,3-Dichloropropene	0.507	0.526		3.75	50
1,1,2-Trichloroethane	0.339	0.360		6.2	50
2-Hexanone	0.403	0.410		1.74	50
Dibromochloromethane	0.326	0.346		6.14	50
1,2-Dibromoethane	0.345	0.361		4.64	50
Tetrachloroethene	0.332	0.328		-1.21	50
Chlorobenzene	1.071	1.072	0.3	0.09	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: P5281 SAS No.: P5281 SDG No.: P5281

Instrument ID: MSVOA_X Calibration Date/Time: 12/18/2024 19:52

Lab File ID: VX044395.D Init. Calib. Date(s): 12/11/2024 12/11/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:41 13:00

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.851	1.810		-2.21	50
m/p-Xylenes	0.679	0.687		1.18	50
o-Xylene	0.687	0.694		1.02	50
Styrene	1.102	1.146		3.99	50
Bromoform	0.219	0.241	0.1	10.05	50
Isopropylbenzene	3.828	3.704		-3.24	50
1,1,2,2-Tetrachloroethane	1.329	1.359	0.3	2.26	50
1,3-Dichlorobenzene	1.621	1.607		-0.86	50
1,4-Dichlorobenzene	1.667	1.616		-3.06	50
1,2-Dichlorobenzene	1.684	1.678		-0.36	50
1,2-Dibromo-3-Chloropropane	0.258	0.250		-3.1	50
1,2,4-Trichlorobenzene	0.923	0.932		0.98	50
1,2,3-Trichlorobenzene	0.979	0.985		0.61	50
1,2-Dichloroethane-d4	0.877	0.792		-9.69	50
Dibromofluoromethane	0.346	0.340		-1.73	50
Toluene-d8	1.168	1.114		-4.62	50
4-Bromofluorobenzene	0.408	0.378		-7.35	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: P5281 SAS No.: P5281 SDG No.: P5281

Instrument ID: MSVOA_X Calibration Date/Time: 12/20/2024 09:16

Lab File ID: VX044438.D Init. Calib. Date(s): 12/11/2024 12/11/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:41 13:00

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.773	0.854		10.48	20
Chloromethane	0.739	0.782	0.1	5.82	20
Vinyl Chloride	0.770	0.798		3.64	20
Bromomethane	0.546	0.532		-2.56	20
Chloroethane	0.480	0.473		-1.46	20
Trichlorofluoromethane	1.508	1.482		-1.72	20
1,1,2-Trichlorotrifluoroethane	0.610	0.676		10.82	20
1,1-Dichloroethene	0.560	0.624		11.43	20
Acetone	0.357	0.291		-18.49	20
Carbon Disulfide	1.065	1.106		3.85	20
Methyl tert-butyl Ether	2.187	2.093		-4.3	20
Methyl Acetate	0.922	0.896		-2.82	20
Methylene Chloride	0.671	0.689		2.68	20
trans-1,2-Dichloroethene	0.596	0.625		4.87	20
1,1-Dichloroethane	1.172	1.203	0.1	2.64	20
Cyclohexane	0.946	1.000		5.71	20
2-Butanone	0.508	0.457		-10.04	20
Carbon Tetrachloride	0.482	0.533		10.58	20
cis-1,2-Dichloroethene	0.756	0.755		-0.13	20
Bromochloromethane	0.575	0.553		-3.83	20
Chloroform	1.316	1.265		-3.88	20
1,1,1-Trichloroethane	1.093	1.099		0.55	20
Methylcyclohexane	0.535	0.648		21.12	20
Benzene	1.359	1.529		12.51	20
1,2-Dichloroethane	0.560	0.586		4.64	20
Trichloroethene	0.339	0.388		14.45	20
1,2-Dichloropropane	0.347	0.369		6.34	20
Bromodichloromethane	0.465	0.525		12.9	20
4-Methyl-2-Pentanone	0.556	0.573		3.06	20
Toluene	0.853	0.924		8.32	20
t-1,3-Dichloropropene	0.455	0.517		13.63	20
cis-1,3-Dichloropropene	0.507	0.577		13.81	20
1,1,2-Trichloroethane	0.339	0.374		10.32	20
2-Hexanone	0.403	0.414		2.73	20
Dibromochloromethane	0.326	0.381		16.87	20
1,2-Dibromoethane	0.345	0.385		11.59	20
Tetrachloroethene	0.332	0.378		13.85	20
Chlorobenzene	1.071	1.187	0.3	10.83	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: P5281 SAS No.: P5281 SDG No.: P5281

Instrument ID: MSVOA_X Calibration Date/Time: 12/20/2024 09:16

Lab File ID: VX044438.D Init. Calib. Date(s): 12/11/2024 12/11/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:41 13:00

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.851	2.034		9.89	20
m/p-Xylenes	0.679	0.780		14.88	20
o-Xylene	0.687	0.748		8.88	20
Styrene	1.102	1.275		15.7	20
Bromoform	0.219	0.266	0.1	21.46	20
Isopropylbenzene	3.828	4.095		6.97	20
1,1,2,2-Tetrachloroethane	1.329	1.387	0.3	4.36	20
1,3-Dichlorobenzene	1.621	1.811		11.72	20
1,4-Dichlorobenzene	1.667	1.781		6.84	20
1,2-Dichlorobenzene	1.684	1.820		8.08	20
1,2-Dibromo-3-Chloropropane	0.258	0.253		-1.94	20
1,2,4-Trichlorobenzene	0.923	1.098		18.96	20
1,2,3-Trichlorobenzene	0.979	1.140		16.44	20
1,2-Dichloroethane-d4	0.877	0.836		-4.68	20
Dibromofluoromethane	0.346	0.392		13.3	20
Toluene-d8	1.168	1.318		12.84	20
4-Bromofluorobenzene	0.408	0.454		11.27	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: P5281 SAS No.: P5281 SDG No.: P5281

Instrument ID: MSVOA_X Calibration Date/Time: 12/20/2024 19:49

Lab File ID: VX044464.D Init. Calib. Date(s): 12/11/2024 12/11/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:41 13:00

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.773	0.744		-3.75	50
Chloromethane	0.739	0.696	0.1	-5.82	50
Vinyl Chloride	0.770	0.730		-5.2	50
Bromomethane	0.546	0.498		-8.79	50
Chloroethane	0.480	0.507		5.63	50
Trichlorofluoromethane	1.508	1.202		-20.29	50
1,1,2-Trichlorotrifluoroethane	0.610	0.590		-3.28	50
1,1-Dichloroethene	0.560	0.560		0	50
Acetone	0.357	0.270		-24.37	50
Carbon Disulfide	1.065	1.089		2.25	50
Methyl tert-butyl Ether	2.187	1.972		-9.83	50
Methyl Acetate	0.922	0.860		-6.72	50
Methylene Chloride	0.671	0.636		-5.22	50
trans-1,2-Dichloroethene	0.596	0.578		-3.02	50
1,1-Dichloroethane	1.172	1.125	0.1	-4.01	50
Cyclohexane	0.946	0.871		-7.93	50
2-Butanone	0.508	0.433		-14.76	50
Carbon Tetrachloride	0.482	0.495		2.7	50
cis-1,2-Dichloroethene	0.756	0.719		-4.89	50
Bromochloromethane	0.575	0.510		-11.3	50
Chloroform	1.316	1.190		-9.57	50
1,1,1-Trichloroethane	1.093	1.029		-5.86	50
Methylcyclohexane	0.535	0.550		2.8	50
Benzene	1.359	1.393		2.5	50
1,2-Dichloroethane	0.560	0.534		-4.64	50
Trichloroethene	0.339	0.360		6.2	50
1,2-Dichloropropane	0.347	0.345		-0.58	50
Bromodichloromethane	0.465	0.491		5.59	50
4-Methyl-2-Pentanone	0.556	0.524		-5.76	50
Toluene	0.853	0.855		0.23	50
t-1,3-Dichloropropene	0.455	0.482		5.93	50
cis-1,3-Dichloropropene	0.507	0.531		4.73	50
1,1,2-Trichloroethane	0.339	0.345		1.77	50
2-Hexanone	0.403	0.381		-5.46	50
Dibromochloromethane	0.326	0.361		10.74	50
1,2-Dibromoethane	0.345	0.366		6.09	50
Tetrachloroethene	0.332	0.343		3.31	50
Chlorobenzene	1.071	1.094	0.3	2.15	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: P5281 SAS No.: P5281 SDG No.: P5281

Instrument ID: MSVOA_X Calibration Date/Time: 12/20/2024 19:49

Lab File ID: VX044464.D Init. Calib. Date(s): 12/11/2024 12/11/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:41 13:00

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.851	1.868		0.92	50
m/p-Xylenes	0.679	0.713		5.01	50
o-Xylene	0.687	0.708		3.06	50
Styrene	1.102	1.163		5.53	50
Bromoform	0.219	0.251	0.1	14.61	50
Isopropylbenzene	3.828	3.829		0.03	50
1,1,2,2-Tetrachloroethane	1.329	1.278	0.3	-3.84	50
1,3-Dichlorobenzene	1.621	1.669		2.96	50
1,4-Dichlorobenzene	1.667	1.653		-0.84	50
1,2-Dichlorobenzene	1.684	1.709		1.49	50
1,2-Dibromo-3-Chloropropane	0.258	0.238		-7.75	50
1,2,4-Trichlorobenzene	0.923	1.000		8.34	50
1,2,3-Trichlorobenzene	0.979	1.036		5.82	50
1,2-Dichloroethane-d4	0.877	0.778		-11.29	50
Dibromofluoromethane	0.346	0.363		4.91	50
Toluene-d8	1.168	1.213		3.85	50
4-Bromofluorobenzene	0.408	0.418		2.45	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.