

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: P5281 SAS No.: P5281 SDG NO.: P5281
Lab File ID: VX044375.D Date Analyzed: 12/18/2024
Instrument ID: MSVOA_X Time Analyzed: 11:57
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	100489	12.018				
UPPER LIMIT	200978	12.518				
LOWER LIMIT	50244.5	11.518				
EPA SAMPLE NO.						
RE114D1-20241212	75803	12.02				
BPOW1-8-20241212	77822	12.02				
TT101D2-20241212	76850	12.02				
RE117D1-20241213	67944	12.02				
RE115D1-20241213	70750	12.02				
TT189D2-20241213	76670	12.02				
FB02-20241213	75075	12.02				
TB	70980	12.02				
VX1218WBL01	82810	12.02				
VX1218WBS01	101950	12.02				
VX1218WBSD01	101474	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.