

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: P5282 SAS No.: P5282 SDG No.: P5282

Instrument ID: BNA_N Calibration Date/Time: 12/20/2024 08:04

Lab File ID: BN035731.D Init. Calib. Date(s): 11/27/2024 11/27/2024

EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 15:34 19:09

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.626	0.562		-10.2	20.0
Fluoranthene-d10	1.134	0.985		-13.1	20.0
2-Fluorophenol	1.001	0.990		-1.1	20.0
Phenol-d6	1.204	1.140		-5.3	20.0
Nitrobenzene-d5	0.244	0.284		16.4	20.0
2-Fluorobiphenyl	1.512	1.376		-9.0	20.0
2,4,6-Tribromophenol	0.284	0.221		-22.2	20.0
Terphenyl-d14	0.789	0.839		6.3	20.0
1,4-Dioxane	0.382	0.405		6.0	20.0

All other compounds must meet a minimum RRF of 0.010.