



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Matrix Spike/Matrix Spike Duplicate Summary
SW-846

SDG No.: P5282

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8270-Modified

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	P5280-05MSD	Client Sample ID:	BPOW6-11-20241212MSD					DataFile:	BN035736.D		
1,4-Dioxane	0.4	0	0.55	ug/L	138	*	4		70	130	20