

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: P5282 SDG NO.: P5282

Lab File ID: BN035730.D

DFTPP Injection Date: 12/20/2024

Instrument ID: BNA_N

DFTPP Injection Time: 07:25

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	22.5
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	31.3
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	38.3
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	28.6
365	Greater than 1% of mass 198	4.1
441	Present, but less than mass 443	13.2
442	Greater than 50% of mass 198	78.5
443	15.0 - 24.0% of mass 442	15.5 (19.8) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN035731.D	12/20/2024	08:04
BPOW6-11-20241212MS	P5280-04MS	BN035735.D	12/20/2024	11:05
BPOW6-11-20241212MSD	P5280-05MSD	BN035736.D	12/20/2024	11:41
SSTDCCC0.4EC	SSTDCCC0.4	BN035747.D	12/20/2024	18:18