

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: P5282 SDG NO.: P5282

Lab File ID: BN035844.D

DFTPP Injection Date: 12/26/2024

Instrument ID: BNA_N

DFTPP Injection Time: 09:19

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	24
68	Less than 2.0% of mass 69	0.2 (0.6) 1
69	Mass 69 relative abundance	32.9
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	40
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.7
275	10.0 - 60.0% of mass 198	27.1
365	Greater than 1% of mass 198	3.9
441	Present, but less than mass 443	11.5
442	Greater than 50% of mass 198	68.6
443	15.0 - 24.0% of mass 442	14.1 (20.5) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN035845.D	12/26/2024	09:58
RW7-SP100-20241212DL	P5282-01DL	BN035857.D	12/26/2024	17:18
SSTDCCC0.4EC	SSTDCCC0.4	BN035861.D	12/26/2024	19:41