

Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB133961

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/16/2024	09:00
2	CAL2	1	Water	NA	NA	20.2	7.00	12/16/2024	09:01
3	CAL3	1	Water	NA	NA	20.3	10.02	12/16/2024	09:05
4	ICV	1	Water	NA	NA	20.3	7.01	12/16/2024	09:11
5	CCV1	1	Water	NA	NA	20.3	2.01	12/16/2024	09:15
6	P5283-02	1	Water	NA	NA	20.1	5.28	12/16/2024	09:25
7	P5286-01	1	Water	NA	NA	20.2	6.02	12/16/2024	09:35
8	P5286-02	1	Water	NA	NA	20.1	4.40	12/16/2024	09:44
9	P5286-03	1	Water	NA	NA	20.5	5.01	12/16/2024	09:47
10	P5286-04	1	Water	NA	NA	20.5	4.98	12/16/2024	09:55
11	P5286-05	1	Water	NA	NA	20.5	5.11	12/16/2024	10:00
12	P5286-06	1	Water	NA	NA	20.8	5.03	12/16/2024	10:05
13	P5287-05	1	Water	NA	NA	20.4	6.42	12/16/2024	10:10
14	P5291-13	1	Water	NA	NA	20.7	6.59	12/16/2024	10:15
15	P5291-13DUP	1	Water	NA	NA	20.8	6.60	12/16/2024	10:16
16	CCV2	1	Water	NA	NA	20.2	12.02	12/16/2024	10:20

WORKLIST(Hardcopy Internal Chain)

133961

WorkList Name : ph p5286 WorkList ID : 186348 Department : Wet-Chemistry Date : 12-16-2024 08:07:18

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5283-02	TT-RW10A-IDWGW-20241213	Water	pH	Cool 4 deg C	TETR06	L61	12/13/2024	9040C
P5286-01	120424-COMP	Water	pH	Cool 4 deg C	PSEG03	L41	12/13/2024	9040C
P5286-02	34850	Water	pH	Cool 4 deg C	PSEG03	L41	12/13/2024	9040C
P5286-03	34851	Water	pH	Cool 4 deg C	PSEG03	L41	12/13/2024	9040C
P5286-04	34852	Water	pH	Cool 4 deg C	PSEG03	L41	12/13/2024	9040C
P5286-05	34853	Water	pH	Cool 4 deg C	PSEG03	L41	12/13/2024	9040C
P5286-06	34854	Water	pH	Cool 4 deg C	PSEG03	L41	12/13/2024	9040C
P5287-05	220-2-TRANS-WATER	Water	pH	Cool 4 deg C	PSEG03	L61	12/13/2024	9040C
P5291-13	WC-20241213	Water	pH	Cool 4 deg C	PARS02	L51	12/13/2024	9040C

Date/Time 12-16-24 08:40
Raw Sample Received by: 78 (WC)
Raw Sample Relinquished by: e6sor

Date/Time 12-16-24 12:30
Raw Sample Received by: cf sor
Raw Sample Relinquished by: sf col