



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Tetra Tech NUS, Inc.

SDG No.: P5283

Project: CTO WE13

RunNo.: LB133961

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	7.01	7	100	90-110	12/16/2024
Sample ID:	CCV1						
pH		pH	2.01	2.00	101	90-110	12/16/2024
Sample ID:	CCV2						
pH		pH	12.02	12.00	100	90-110	12/16/2024