

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: Tetra Tech NUS, Inc. **SDG No.:** P5283
Contract: TETR06 **Lab Code:** CHEM **Case No.:** P5283 **SAS No.:** P5283
Matrix: Water **Level:** LOW **Client ID:** 220-2-TRANS-WATERA
Sample ID: P5287-05 **Spiked ID:** P5287-05A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Potassium	ug/L	86 - 114	8030		4080		5000	79		P