

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc. **SDG No.:** P5283
Contract: TETR06 **Lab Code:** CHEM **Case No.:** P5283 **SAS No.:** P5283

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165659BS							
Aluminum	ug/L	1000	963		96	86 - 115	P
Antimony	ug/L	400	411		103	88 - 113	P
Arsenic	ug/L	400	412		103	87 - 113	P
Barium	ug/L	100	92.8		93	88 - 113	P
Beryllium	ug/L	100	90.7		91	89 - 112	P
Cadmium	ug/L	100	94.9		95	88 - 113	P
Calcium	ug/L	500	494	J	99	87 - 113	P
Chromium	ug/L	200	201		100	90 - 113	P
Cobalt	ug/L	100	97.5		98	89 - 114	P
Copper	ug/L	150	160		107	86 - 114	P
Iron	ug/L	1500	1530		102	87 - 115	P
Lead	ug/L	500	482		96	86 - 113	P
Magnesium	ug/L	1000	905	J	90	85 - 113	P
Manganese	ug/L	100	94.6		95	90 - 114	P
Nickel	ug/L	250	245		98	88 - 113	P
Potassium	ug/L	5000	5060		101	86 - 114	P
Selenium	ug/L	1000	1020		102	83 - 114	P
Silver	ug/L	37.5	38.0		101	84 - 115	P
Sodium	ug/L	1500	1430		95	87 - 115	P
Thallium	ug/L	1000	992		99	85 - 114	P
Vanadium	ug/L	150	143		95	90 - 111	P
Zinc	ug/L	100	101		101	87 - 115	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: P5283

Contract: TETR06

Lab Code: CHEM

Case No.: P5283

SAS No.: P5283

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165693BS Mercury	ug/L	4.0	4.49		112	82 - 119	CV