

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: P5283 SAS No.: P5283 SDG NO.: P5283

Continuing Calib Date: 12/16/2024 Initial Calibration Date(s): 12/06/2024 12/06/2024

Continuing Calib Time: 11:05 Initial Calibration Time(s): 14:19 22:34

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.61	5.62	5.52	5.72	0.01
Aroclor-1254-2 (2)	5.76	5.77	5.67	5.87	0.01
Aroclor-1254-3 (3)	6.17	6.17	6.07	6.27	0.00
Aroclor-1254-4 (4)	6.40	6.40	6.30	6.50	0.00
Aroclor-1254-5 (5)	6.82	6.83	6.73	6.93	0.01
Tetrachloro-m-xylene	3.71	3.71	3.61	3.81	0.00
Decachlorobiphenyl	8.77	8.79	8.69	8.89	0.02

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: P5283 SAS No.: P5283 SDG NO.: P5283

Continuing Calib Date: 12/16/2024 Initial Calibration Date(s): 12/06/2024 12/06/2024

Continuing Calib Time: 11:05 Initial Calibration Time(s): 14:19 22:34

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.59	5.60	5.50	5.70	0.01
Aroclor-1254-2 (2)	5.74	5.75	5.65	5.85	0.01
Aroclor-1254-3 (3)	6.14	6.15	6.05	6.25	0.01
Aroclor-1254-4 (4)	6.37	6.38	6.28	6.48	0.01
Aroclor-1254-5 (5)	6.79	6.80	6.70	6.90	0.01
Tetrachloro-m-xylene	3.70	3.71	3.61	3.81	0.01
Decachlorobiphenyl	8.72	8.74	8.64	8.84	0.02