

Surrogate Summary

SDG No.: **P5283**

Client: **Tetra Tech NUS, Inc.**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PO108361.D	PIBLK-PO108361.D	Tetrachloro-m-xylene	1	20	23.0	115		60	140
		Decachlorobiphenyl	1	20	23.6	118		60	140
		Tetrachloro-m-xylene	2	20	21.6	108		60	140
		Decachlorobiphenyl	2	20	23.6	118		60	140
I.BLK-PO108559.D	PIBLK-PO108559.D	Tetrachloro-m-xylene	1	20	17.6	88		60	140
		Decachlorobiphenyl	1	20	17.3	86		60	140
		Tetrachloro-m-xylene	2	20	18.9	94		60	140
		Decachlorobiphenyl	2	20	21.0	105		60	140
PB165649BL	PB165649BL	Tetrachloro-m-xylene	1	20	17.8	89		35	137
		Decachlorobiphenyl	1	20	18.0	90		40	135
		Tetrachloro-m-xylene	2	20	18.8	94		35	137
		Decachlorobiphenyl	2	20	21.3	106		40	135
PB165649BS	PB165649BS	Tetrachloro-m-xylene	1	20	17.4	87		35	137
		Decachlorobiphenyl	1	20	18.6	93		40	135
		Tetrachloro-m-xylene	2	20	18.2	91		35	137
		Decachlorobiphenyl	2	20	22.2	111		40	135
PB165649BSD	PB165649BSD	Tetrachloro-m-xylene	1	20	17.4	87		35	137
		Decachlorobiphenyl	1	20	18.7	94		40	135
		Tetrachloro-m-xylene	2	20	18.3	92		35	137
		Decachlorobiphenyl	2	20	22.2	111		40	135
P5283-02	TT-RW10A-IDWGW-20241213	Tetrachloro-m-xylene	1	20	19.0	95		35	137
		Decachlorobiphenyl	1	20	10.7	53		40	135
		Tetrachloro-m-xylene	2	20	20.1	101		35	137
		Decachlorobiphenyl	2	20	12.7	64		40	135
I.BLK-PO108572.D	PIBLK-PO108572.D	Tetrachloro-m-xylene	1	20	18.9	94		60	140
		Decachlorobiphenyl	1	20	18.9	95		60	140
		Tetrachloro-m-xylene	2	20	19.9	99		60	140
		Decachlorobiphenyl	2	20	22.2	111		60	140