

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\
 Method File : 8270-SIM-BN112724.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Nov 27 23:03:24 2024
 Response Via : Initial Calibration

Calibration Files

0.1 =BN035350.D 0.2 =BN035351.D 0.4 =BN035352.D 0.8 =BN035353.D 1.6 =BN035354.D 3.2 =BN035355.D 5.0 =BN035356.D

	Compound	0.1	0.2	0.4	0.8	1.6	3.2	5.0	Avg	%RSD

1) I	1,4-Dichlorobenzen...	-----ISTD-----								
2)	1,4-Dioxane	0.406	0.417	0.376	0.380	0.392	0.357	0.348	0.382	6.52
3)	n-Nitrosodimet...	0.334	0.302	0.326	0.315	0.332	0.310	0.309	0.319	3.92
4) S	2-Fluorophenol	1.025	1.112	1.018	0.958	0.998	0.954	0.942	1.001	5.88
5) S	Phenol-d6	1.227	1.186	1.193	1.143	1.235	1.215	1.229	1.204	2.69
6)	bis(2-Chloroet...	1.035	1.021	0.992	0.993	1.051	0.997	0.991	1.012	2.39
7) I	Naphthalene-d8	-----ISTD-----								
8) S	Nitrobenzene-d5	0.227	0.232	0.235	0.248	0.257	0.251	0.261	0.244	5.31
9)	Naphthalene	1.062	1.029	1.047	1.032	1.096	1.049	1.070	1.055	2.22
10)	Hexachlorobuta...	0.245	0.242	0.247	0.241	0.255	0.236	0.238	0.243	2.60
11) SURR2	Methylnaphth...	0.591	0.603	0.619	0.615	0.659	0.639	0.656	0.626	4.16
12)	2-Methylnaphth...	0.724	0.716	0.740	0.747	0.795	0.771	0.795	0.755	4.25
13) I	Acenaphthene-d10	-----ISTD-----								
14) S	2,4,6-Tribromo...	0.273	0.258	0.257	0.268	0.293	0.311	0.328	0.284	9.67
15) S	2-Fluorobiphenyl	1.489	1.491	1.510	1.508	1.566	1.511	1.511	1.512	1.68
16)	Acenaphthylene	1.643	1.600	1.595	1.638	1.737	1.763	1.781	1.680	4.68
17)	Acenaphthene	1.121	1.084	1.086	1.108	1.145	1.122	1.140	1.115	2.17
18)	Fluorene	1.589	1.549	1.543	1.600	1.652	1.614	1.625	1.596	2.47
19) I	Phenanthrene-d10	-----ISTD-----								
20)	4,6-Dinitro-2-...	0.038	0.031	0.036	0.041	0.051			0.039	19.30
21)	4-Bromophenyl-...	0.226	0.218	0.226	0.233	0.249	0.242	0.244	0.234	4.85
22)	Hexachlorobenzene	0.265	0.266	0.273	0.276	0.288	0.278	0.277	0.275	2.82
23)	Atrazine	0.155	0.155	0.154	0.156	0.175	0.179	0.191	0.167	8.98
24)	Pentachlorophenol	0.140	0.090	0.095	0.103	0.121	0.136	0.150	0.120	19.86
25)	Phenanthrene	1.092	1.046	1.067	1.092	1.148	1.121	1.125	1.099	3.20
26)	Anthracene	0.964	0.923	0.940	0.973	1.050	1.042	1.064	0.994	5.76
27) SURR	Fluoranthene-d10	1.203	1.086	1.077	1.105	1.165	1.138	1.164	1.134	4.10
28)	Fluoranthene	1.538	1.396	1.416	1.456	1.539	1.497	1.526	1.481	3.99
29) I	Chrysene-d12	-----ISTD-----								
30)	Pyrene	1.583	1.445	1.475	1.443	1.519	1.440	1.431	1.477	3.79
31) S	Terphenyl-d14	0.832	0.777	0.791	0.771	0.812	0.772	0.769	0.789	3.08
32)	Benzo(a)anthra...	1.431	1.343	1.355	1.375	1.451	1.411	1.429	1.399	2.98
33)	Chrysene	1.463	1.452	1.441	1.415	1.487	1.422	1.420	1.443	1.84
34)	Bis(2-ethylhex...	0.710	0.558	0.516	0.505	0.520	0.516	0.544	0.553	12.96
35) I	Perylene-d12	-----ISTD-----								

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\

Method File : 8270-SIM-BN112724.M

36)	Indeno(1,2,3-c...	1.411	1.489	1.532	1.554	1.660	1.615	1.685	1.564	6.22
37)	Benzo(b)fluora...	1.305	1.348	1.313	1.378	1.827	1.463	1.608	1.463	13.12
38)	Benzo(k)fluora...	1.444	1.376	1.402	1.419	1.527	1.447	1.468	1.440	3.39
39) C	Benzo(a)pyrene	1.204	1.156	1.146	1.171	1.256	1.232	1.271	1.205	4.11
40)	Dibenzo(a,h)an...	1.104	1.187	1.194	1.226	1.315	1.280	1.332	1.234	6.55
41)	Benzo(g,h,i)pe...	1.188	1.238	1.248	1.269	1.360	1.330	1.394	1.289	5.71

(#) = Out of Range