

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: P5284 SAS No.: P5284 SDG No.: P5284

Instrument ID: BNA_N Calibration Date/Time: 12/26/2024 09:58

Lab File ID: BN035845.D Init. Calib. Date(s): 11/27/2024 11/27/2024

EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 15:34 19:09

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.626	0.545		-12.9	20.0
Fluoranthene-d10	1.134	0.991		-12.6	20.0
2-Fluorophenol	1.001	1.006		0.5	20.0
Phenol-d6	1.204	1.112		-7.6	20.0
Nitrobenzene-d5	0.244	0.298		22.1	20.0
2-Fluorobiphenyl	1.512	1.504		-0.5	20.0
2,4,6-Tribromophenol	0.284	0.248		-12.7	20.0
Terphenyl-d14	0.789	0.784		-0.6	20.0
1,4-Dioxane	0.382	0.405		6.0	20.0

All other compounds must meet a minimum RRF of 0.010.