

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112724\
 Data File : BN035352.D
 Acq On : 27 Nov 2024 16:46
 Operator : RC/JU
 Sample : SSTDICCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICCC0.4

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/29/2024
 Supervised By :mohammad ahmed 12/03/2024

Quant Time: Nov 27 22:52:52 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 27 22:48:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.308	152	2048	0.400	ng	0.00
7) Naphthalene-d8	10.052	136	5229	0.400	ng	0.00
13) Acenaphthene-d10	13.967	164	3799	0.400	ng	0.00
19) Phenanthrene-d10	16.736	188	9490	0.400	ng	0.00
29) Chrysene-d12	20.974	240	9527	0.400	ng	0.00
35) Perylene-d12	23.067	264	10842	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	4.968	112	2085	0.401	ng	0.00
5) Phenol-d6	6.513	99	2444	0.375	ng	0.00
8) Nitrobenzene-d5	8.440	82	1228m	0.270	ng	0.00
11) 2-Methylnaphthalene-d10	11.656	152	3237	0.347	ng	0.00
14) 2,4,6-Tribromophenol	15.475	330	975	0.356	ng	0.00
15) 2-Fluorobiphenyl	12.574	172	5735	0.372	ng	0.00
27) Fluoranthene-d10	18.785	212	10223	0.352	ng	0.00
31) Terphenyl-d14	19.412	244	7533	0.377	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.003	88	771	0.414	ng	100
3) n-Nitrosodimethylamine	3.292	42	668	0.385	ng	100
6) bis(2-Chloroethyl)ether	6.759	93	2032	0.415	ng	100
9) Naphthalene	10.105	128	5475	0.401	ng	100
10) Hexachlorobutadiene	10.404	225	1292	0.323	ng	# 100
12) 2-Methylnaphthalene	11.732	142	3867	0.384	ng	100
16) Acenaphthylene	13.679	152	6058	0.373	ng	100
17) Acenaphthene	14.031	154	4125	0.388	ng	100
18) Fluorene	15.026	166	5863	0.375	ng	100
20) 4,6-Dinitro-2-methylph...	15.133	198	337	0.171	ng	100
21) 4-Bromophenyl-phenylether	15.941	248	2146	0.355	ng	100
22) Hexachlorobenzene	16.040	284	2590	0.413	ng	100
23) Atrazine	16.227	200	1461	0.269	ng	100
24) Pentachlorophenol	16.400	266	906	0.309	ng	87
25) Phenanthrene	16.773	178	10127	0.405	ng	100
26) Anthracene	16.860	178	8921	0.389	ng	100
28) Fluoranthene	18.817	202	13438	0.391	ng	100
30) Pyrene	19.179	202	14053	0.443	ng	100
32) Benzo(a)anthracene	20.956	228	12908	0.389	ng	100
33) Chrysene	21.009	228	13729	0.418	ng	100
34) Bis(2-ethylhexyl)phtha...	20.929	149	4912	0.282	ng	100
36) Indeno(1,2,3-cd)pyrene	25.108	276	16615	0.384	ng	100
37) Benzo(b)fluoranthene	22.453	252	14235	0.390	ng	100
38) Benzo(k)fluoranthene	22.494	252	15199	0.416	ng	100
39) Benzo(a)pyrene	22.977	252	12425	0.387	ng	100
40) Dibenzo(a,h)anthracene	25.126	278	12948	0.378	ng	100
41) Benzo(g,h,i)perylene	25.722	276	13529	0.371	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112724\
Data File : BN035352.D
Acq On : 27 Nov 2024 16:46
Operator : RC/JU
Sample : SSTDICCC0.4
Misc :
ALS Vial : 4 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4

Quant Time: Nov 27 22:52:52 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 27 22:48:24 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/29/2024
Supervised By :mohammad ahmed 12/03/2024






#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.308 min Scan# 613
Delta R.T. -0.000 min
Lab File: BN035352.D
Acq: 27 Nov 2024 16:46

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4






Tgt Ion:152 Resp: 2048
Ion Ratio Lower Upper
152 100
150 155.0 124.0 186.0
115 62.0 49.6 74.4

Manual Integrations

APPROVED




Reviewed By :Yogesh Patel 11/29/2024


Supervised By :mohammad ahmed 12/03/2024



#2
1,4-Dioxane
Concen: 0.414 ng
RT: 3.003 min Scan# 17
Delta R.T. 0.000 min
Lab File: BN035352.D
Acq: 27 Nov 2024 16:46


Tgt Ion: 88 Resp: 771
Ion Ratio Lower Upper
88 100
43 21.5 17.2 25.8
58 55.6 44.5 66.7





#3
n-Nitrosodimethylamine
Concen: 0.385 ng
RT: 3.292 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN035352.D
Acq: 27 Nov 2024 16:46




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4




Tgt Ion: 42 Resp: 668
Ion Ratio Lower Upper
42 100
74 156.1 124.9 187.3
44 2.8 2.2 3.4

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 11/29/2024
Supervised By :mohammad ahmed 12/03/2024




#4
2-Fluorophenol
Concen: 0.401 ng
RT: 4.968 min Scan# 289
Delta R.T. 0.000 min
Lab File: BN035352.D
Acq: 27 Nov 2024 16:46




Tgt Ion:112 Resp: 2085
Ion Ratio Lower Upper
112 100
64 49.8 39.8 59.8
63 26.3 21.0 31.6





#5
Phenol-d6
Concen: 0.375 ng
RT: 6.513 min Scan# 503
Delta R.T. 0.000 min
Lab File: BN035352.D
Acq: 27 Nov 2024 16:46




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4




Tgt Ion	Ratio	Lower	Upper
99	100		
42	14.3	11.4	17.2
71	36.6	29.3	43.9

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 11/29/2024
Supervised By :mohammad ahmed 12/03/2024




#6
bis(2-Chloroethyl)ether
Concen: 0.415 ng
RT: 6.759 min Scan# 537
Delta R.T. 0.000 min
Lab File: BN035352.D
Acq: 27 Nov 2024 16:46



Tgt Ion	Ratio	Lower	Upper
93	100		
63	63.0	50.4	75.6
95	32.1	25.7	38.5





#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.052 min Scan# 915
Delta R.T. 0.000 min
Lab File: BN035352.D
Acq: 27 Nov 2024 16:46

Instrument :

BNA_N

ClientSampleId :





SSTDICCC0.4

Tgt Ion:136 Resp: 5229
Ion Ratio Lower Upper
136 100
137 12.7 10.2 15.2
54 7.6 6.1 9.1
68 8.0 6.4 9.6

Manual Integrations




APPROVED


Reviewed By :Yogesh Patel 11/29/2024
Supervised By :mohammad ahmed 12/03/2024



#8
Nitrobenzene-d5
Concen: 0.270 ng m
RT: 8.440 min Scan# 764
Delta R.T. 0.000 min
Lab File: BN035352.D
Acq: 27 Nov 2024 16:46

Tgt Ion: 82 Resp: 1228
Ion Ratio Lower Upper
82 100
128 41.7 33.4 50.0
54 45.9 36.7 55.1









#9
Naphthalene
Concen: 0.401 ng
RT: 10.105 min Scan# 911
Delta R.T. 0.000 min
Lab File: BN035352.D
Acq: 27 Nov 2024 16:46

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4

Tgt Ion:128 Resp: 547
Ion Ratio Lower Upper
128 100
129 12.2 9.8 14.6
127 14.3 11.4 17.2




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 11/29/2024
Supervised By :mohammad ahmed 12/03/2024



#10
Hexachlorobutadiene
Concen: 0.323 ng
RT: 10.404 min Scan# 948
Delta R.T. 0.000 min
Lab File: BN035352.D
Acq: 27 Nov 2024 16:46


Tgt Ion:225 Resp: 1292
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.1 51.3 76.9





#11
2-Methylnaphthalene-d10
Concen: 0.347 ng
RT: 11.656 min Scan# 1081
Delta R.T. 0.000 min
Lab File: BN035352.D
Acq: 27 Nov 2024 16:46




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4



Tgt Ion:152 Resp: 323
Ion Ratio Lower Upper
152 100
151 20.8 16.6 25.0




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 11/29/2024
Supervised By :mohammad ahmed 12/03/2024



#12
2-Methylnaphthalene
Concen: 0.384 ng
RT: 11.732 min Scan# 1096
Delta R.T. 0.000 min
Lab File: BN035352.D
Acq: 27 Nov 2024 16:46


Tgt Ion:142 Resp: 3867
Ion Ratio Lower Upper
142 100
141 90.3 72.2 108.4
115 39.2 31.4 47.0





#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 13.967 min Scan# 1437
Delta R.T. 0.000 min
Lab File: BN035352.D
Acq: 27 Nov 2024 16:46




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4




Tgt Ion:164 Resp: 3799
Ion Ratio Lower Upper
164 100
162 102.7 82.2 123.2
160 50.1 40.1 60.1

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 11/29/2024
Supervised By :mohammad ahmed 12/03/2024




#14
2,4,6-Tribromophenol
Concen: 0.356 ng
RT: 15.475 min Scan# 1578
Delta R.T. 0.000 min
Lab File: BN035352.D
Acq: 27 Nov 2024 16:46




Tgt Ion:330 Resp: 975
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 33.3 26.6 40.0





#15
2-Fluorobiphenyl
Concen: 0.372 ng
RT: 12.574 min Scan# 1262
Delta R.T. 0.000 min
Lab File: BN035352.D
Acq: 27 Nov 2024 16:46




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4




Tgt Ion:172 Resp: 573
Ion Ratio Lower Upper
172 100
171 36.2 29.0 43.4
170 24.8 19.8 29.8

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 11/29/2024
Supervised By :mohammad ahmed 12/03/2024




#16
Acenaphthylene
Concen: 0.373 ng
RT: 13.679 min Scan# 1410
Delta R.T. 0.000 min
Lab File: BN035352.D
Acq: 27 Nov 2024 16:46




Tgt Ion:152 Resp: 6058
Ion Ratio Lower Upper
152 100
151 20.2 16.2 24.2
153 13.0 10.4 15.6





#17
Acenaphthene
Concen: 0.388 ng
RT: 14.031 min Scan# 1443
Delta R.T. 0.000 min
Lab File: BN035352.D
Acq: 27 Nov 2024 16:46




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4



Tgt Ion:154 Resp: 4125
Ion Ratio Lower Upper
154 100
153 115.8 92.6 139.0
152 61.3 49.0 73.6




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 11/29/2024
Supervised By :mohammad ahmed 12/03/2024



#18
Fluorene
Concen: 0.375 ng
RT: 15.026 min Scan# 1536
Delta R.T. 0.000 min
Lab File: BN035352.D
Acq: 27 Nov 2024 16:46

Tgt Ion:166 Resp: 5863
Ion Ratio Lower Upper
166 100
165 99.6 79.7 119.5
167 13.5 10.8 16.2









#19
 Phenanthrene-d10
 Concen: 0.400 ng
 RT: 16.736 min Scan# 1680
 Delta R.T. 0.000 min
 Lab File: BN035352.D
 Acq: 27 Nov 2024 16:46

Instrument : BNA_N
 ClientSampleId : SSTDICC0.4

Tgt Ion:188	Resp:	9490
Ion Ratio	Lower	Upper
188	100	
94	0.0	0.0
80	5.7	6.8




Manual Integrations
 APPROVED


Reviewed By :Yogesh Patel 11/29/2024
 Supervised By :mohammad ahmed 12/03/2024



#20
 4,6-Dinitro-2-methylphenol
 Concen: 0.171 ng
 RT: 15.133 min Scan# 1546
 Delta R.T. 0.000 min
 Lab File: BN035352.D
 Acq: 27 Nov 2024 16:46

Tgt Ion:198	Resp:	337
Ion Ratio	Lower	Upper
198	100	
51	58.1	46.5
105	56.6	45.3





#21
4-Bromophenyl-phenylether
Concen: 0.355 ng
RT: 15.941 min Scan# 1616
Delta R.T. 0.000 min
Lab File: BN035352.D
Acq: 27 Nov 2024 16:46

Instrument :

BNA_N

ClientSampleId :





SSTDICCC0.4

Tgt Ion:248 Resp: 2140
Ion Ratio Lower Upper
248 100
250 100.7 80.6 120.8
141 39.4 31.5 47.3

Manual Integrations




APPROVED


Reviewed By :Yogesh Patel 11/29/2024
Supervised By :mohammad ahmed 12/03/2024



#22
Hexachlorobenzene
Concen: 0.413 ng
RT: 16.040 min Scan# 1624
Delta R.T. 0.000 min
Lab File: BN035352.D
Acq: 27 Nov 2024 16:46


Tgt Ion:284 Resp: 2590
Ion Ratio Lower Upper
284 100
142 33.4 26.7 40.1
249 30.7 24.6 36.8





#23
Atrazine
Concen: 0.269 ng
RT: 16.227 min Scan# 1639
Delta R.T. 0.000 min
Lab File: BN035352.D
Acq: 27 Nov 2024 16:46




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4



Tgt Ion:200 Resp: 146
Ion Ratio Lower Upper
200 100
173 30.1 24.1 36.1
215 46.1 36.9 55.3




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 11/29/2024
Supervised By :mohammad ahmed 12/03/2024



#24
Pentachlorophenol
Concen: 0.309 ng
RT: 16.400 min Scan# 1653
Delta R.T. 0.000 min
Lab File: BN035352.D
Acq: 27 Nov 2024 16:46

Tgt Ion:266 Resp: 906
Ion Ratio Lower Upper
266 100
264 62.4 42.3 63.5
268 63.8 43.3 64.9









#25
Phenanthrene
Concen: 0.405 ng
RT: 16.773 min Scan# 1683
Delta R.T. 0.000 min
Lab File: BN035352.D
Acq: 27 Nov 2024 16:46

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4

Tgt Ion:178 Resp: 1012
Ion Ratio Lower Upper
178 100
176 19.3 15.4 23.2
179 15.4 12.3 18.5




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 11/29/2024
Supervised By :mohammad ahmed 12/03/2024



#26
Anthracene
Concen: 0.389 ng
RT: 16.860 min Scan# 1690
Delta R.T. 0.000 min
Lab File: BN035352.D
Acq: 27 Nov 2024 16:46

Tgt Ion:178 Resp: 8921
Ion Ratio Lower Upper
178 100
176 18.8 15.0 22.6
179 15.7 12.6 18.8





#27
Fluoranthene-d10
Concen: 0.352 ng
RT: 18.785 min Scan# 1882
Delta R.T. 0.000 min
Lab File: BN035352.D
Acq: 27 Nov 2024 16:46

Instrument :

BNA_N

ClientSampleId :





SSTDICCC0.4

Tgt Ion:212 Resp: 1022
Ion Ratio Lower Upper
212 100
106 11.5 9.2 13.8
104 6.6 5.3 7.9

Manual Integrations




APPROVED


Reviewed By :Yogesh Patel 11/29/2024
Supervised By :mohammad ahmed 12/03/2024



#28
Fluoranthene
Concen: 0.391 ng
RT: 18.817 min Scan# 1889
Delta R.T. 0.000 min
Lab File: BN035352.D
Acq: 27 Nov 2024 16:46


Tgt Ion:202 Resp: 13438
Ion Ratio Lower Upper
202 100
101 9.2 7.4 11.0
203 17.1 13.7 20.5





#29
Chrysene-d12
Concen: 0.400 ng
RT: 20.974 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN035352.D
Acq: 27 Nov 2024 16:46




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4



Tgt Ion:240 Resp: 952
Ion Ratio Lower Upper
240 100
120 9.9 7.9 11.9
236 28.6 22.9 34.3




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 11/29/2024
Supervised By :mohammad ahmed 12/03/2024



#30
Pyrene
Concen: 0.443 ng
RT: 19.179 min Scan# 1967
Delta R.T. 0.000 min
Lab File: BN035352.D
Acq: 27 Nov 2024 16:46

Tgt Ion:202 Resp: 14053
Ion Ratio Lower Upper
202 100
200 21.2 17.0 25.4
203 17.9 14.3 21.5





#31
 Terphenyl-d14
 Concen: 0.377 ng
 RT: 19.412 min Scan# 2017
 Delta R.T. 0.000 min
 Lab File: BN035352.D
 Acq: 27 Nov 2024 16:46

Instrument :

BNA_N

ClientSampleId :





SSTDICCC0.4

Tgt Ion:244 Resp: 753
 Ion Ratio Lower Upper
 244 100
 212 10.1 8.1 12.1
 122 12.9 10.3 15.5

Manual Integrations




APPROVED


Reviewed By :Yogesh Patel 11/29/2024
 Supervised By :mohammad ahmed 12/03/2024



#32
 Benzo(a)anthracene
 Concen: 0.389 ng
 RT: 20.956 min Scan# 2251
 Delta R.T. 0.000 min
 Lab File: BN035352.D
 Acq: 27 Nov 2024 16:46

Tgt Ion:228 Resp: 12908
 Ion Ratio Lower Upper
 228 100
 226 28.1 22.5 33.7
 229 19.8 15.8 23.8









#33
 Chrysene
 Concen: 0.418 ng
 RT: 21.009 min Scan# 2100
 Delta R.T. 0.000 min
 Lab File: BN035352.D
 Acq: 27 Nov 2024 16:46

Instrument : BNA_N
 ClientSampleId : SSTDICC0.4

Tgt Ion:	228	Resp:	13729
Ion Ratio	Lower	Upper	
228	100		
226	30.8	24.6	37.0
229	19.9	15.9	23.9




Manual Integrations
 APPROVED


Reviewed By :Yogesh Patel 11/29/2024
 Supervised By :mohammad ahmed 12/03/2024



#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.282 ng
 RT: 20.929 min Scan# 2248
 Delta R.T. 0.000 min
 Lab File: BN035352.D
 Acq: 27 Nov 2024 16:46


Tgt Ion:	149	Resp:	4912
Ion Ratio	Lower	Upper	
149	100		
167	27.8	22.2	33.4
279	3.4	2.7	4.1





#35
Perylene-d12
Concen: 0.400 ng
RT: 23.067 min Scan# 2639
Delta R.T. 0.000 min
Lab File: BN035352.D
Acq: 27 Nov 2024 16:46



Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4



Tgt Ion:264 Resp: 1084
Ion Ratio Lower Upper
264 100
260 26.8 21.4 32.2
265 50.3 40.2 60.4




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 11/29/2024
Supervised By :mohammad ahmed 12/03/2024



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.384 ng
RT: 25.108 min Scan# 3337
Delta R.T. 0.000 min
Lab File: BN035352.D
Acq: 27 Nov 2024 16:46


Tgt Ion:276 Resp: 16615
Ion Ratio Lower Upper
276 100
138 24.2 19.4 29.0
277 24.7 19.8 29.6





#37
Benzo(b)fluoranthene
Concen: 0.390 ng
RT: 22.453 min Scan# 2429
Delta R.T. 0.000 min
Lab File: BN035352.D
Acq: 27 Nov 2024 16:46




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4




Tgt Ion:252 Resp: 14235
Ion Ratio Lower Upper
252 100
253 24.5 19.6 29.4
125 12.0 9.6 14.4

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 11/29/2024
Supervised By :mohammad ahmed 12/03/2024




#38
Benzo(k)fluoranthene
Concen: 0.416 ng
RT: 22.494 min Scan# 2443
Delta R.T. 0.000 min
Lab File: BN035352.D
Acq: 27 Nov 2024 16:46



Tgt Ion:252 Resp: 15199
Ion Ratio Lower Upper
252 100
253 24.4 19.5 29.3
125 12.8 10.2 15.4






#39
Benzo(a)pyrene
Concen: 0.387 ng
RT: 22.977 min Scan# 2608
Delta R.T. 0.000 min
Lab File: BN035352.D
Acq: 27 Nov 2024 16:46

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4






Tgt Ion:252 Resp: 12425
Ion Ratio Lower Upper
252 100
253 25.3 20.2 30.4
125 13.6 10.9 16.3

Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 11/29/2024



Supervised By :mohammad ahmed 12/03/2024




#40
Dibenzo(a,h)anthracene
Concen: 0.378 ng
RT: 25.126 min Scan# 3343
Delta R.T. 0.000 min
Lab File: BN035352.D
Acq: 27 Nov 2024 16:46




Tgt Ion:278 Resp: 12948
Ion Ratio Lower Upper
278 100
139 17.8 14.2 21.4
279 25.6 20.5 30.7





#41
Benzo(g,h,i)perylene
Concen: 0.371 ng
RT: 25.722 min Scan# 31
Delta R.T. 0.000 min
Lab File: BN035352.D
Acq: 27 Nov 2024 16:46

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4



Tgt Ion:276 Resp: 13529
Ion Ratio Lower Upper
276 100
277 24.9 19.9 29.9
138 22.3 17.8 26.8

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/29/2024
Supervised By :mohammad ahmed 12/03/2024

