

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:		
Project:	CTO WE13		Date Received:		
Client Sample ID:	PB165667BS		SDG No.:	P5284	
Lab Sample ID:	PB165667BS		Matrix:	Water	
Analytical Method:	SW8270ESIM		% Solid:	0	
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN035808.D	1	12/16/24 11:30	12/24/24 06:34	PB165667

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.37		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.46		30 - 150		115%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.37		30 - 150		92%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.47	*	55 - 111		116%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.40		53 - 106		100%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.43		58 - 132		108%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	4070	7.264				
1146-65-2	Naphthalene-d8	9750	10.009				
15067-26-2	Acenaphthene-d10	5530	13.914				
1517-22-2	Phenanthrene-d10	10900	16.698				
1719-03-5	Chrysene-d12	9470	20.947				
1520-96-3	Perylene-d12	9440	23.032				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products