

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P5284 SAS No.: P5284 SDG NO.: P5284

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 12/20/2024

Lab File ID: BN035749.D Time Analyzed: 19:33

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2885	7.272	7460	10.01	4592	13.93
UPPER LIMIT	5770	7.772	14920	10.509	9184	14.425
LOWER LIMIT	1442.5	6.772	3730	9.509	2296	13.425
EPA SAMPLE NO.						
01 RW5-SP201-20241212	2707	7.27	6635	10.02	4389	13.94
02 RW5-SP303-20241212	2594	7.27	6365	10.02	4185	13.94

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.