

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P5284 SAS No.: P5284 SDG NO.: P5284

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 12/23/2024

Lab File ID: BN035794.D Time Analyzed: 22:16

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	3234	7.264	7877	10.01	4482	13.92
UPPER LIMIT	6468	7.764	15754	10.509	8964	14.424
LOWER LIMIT	1617	6.764	3938.5	9.509	2241	13.424
EPA SAMPLE NO.						
01 RW5-SP100-20241212	3703	7.27	8793	10.01	5556	13.93
02 PB165667BL	4926	7.26	11442	10.01	7253	13.92
03 PB165667BS	4067	7.26	9754	10.01	5527	13.91

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.