

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P5288 SAS No.: P5288 SDG NO.: P5288

Continuing Calib Date: 12/17/2024 Initial Calibration Date(s): 11/21/2024 11/22/2024

Continuing Calib Time: 16:40 Initial Calibration Time(s): 16:13 01:47

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1248-1	(1)	4.85	4.86	4.79	4.93	0.01
Aroclor-1248-2	(2)	5.14	5.15	5.08	5.22	0.01
Aroclor-1248-3	(3)	5.36	5.36	5.29	5.43	0.00
Aroclor-1248-4	(4)	5.79	5.80	5.73	5.87	0.01
Aroclor-1248-5	(5)	5.83	5.83	5.76	5.90	0.00
TCX		3.63	3.64	3.59	3.69	0.01
DCB		9.52	9.53	9.43	9.63	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P5288 SAS No.: P5288 SDG NO.: P5288

Continuing Calib Date: 12/17/2024 Initial Calibration Date(s): 11/21/2024 11/22/2024

Continuing Calib Time: 16:40 Initial Calibration Time(s): 16:13 01:47

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1248-1	(1)	4.08	4.09	4.02	4.16	0.01
Aroclor-1248-2	(2)	4.33	4.34	4.27	4.41	0.01
Aroclor-1248-3	(3)	4.38	4.38	4.31	4.45	0.00
Aroclor-1248-4	(4)	4.55	4.56	4.49	4.63	0.01
Aroclor-1248-5	(5)	4.97	4.97	4.90	5.04	0.00
TCX		2.97	2.97	2.92	3.02	0.01
DCB		8.24	8.25	8.15	8.35	0.01