

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P5288 SAS No.: P5288 SDG NO.: P5288

Continuing Calib Date: 12/17/2024 Initial Calibration Date(s): 11/21/2024 11/22/2024

Continuing Calib Time: 16:55 Initial Calibration Time(s): 16:13 01:47

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	5.76	5.77	5.70	5.84	0.01
Aroclor-1254-2	(2)	6.00	6.00	5.93	6.07	0.00
Aroclor-1254-3	(3)	6.39	6.39	6.32	6.46	0.00
Aroclor-1254-4	(4)	6.70	6.71	6.64	6.78	0.01
Aroclor-1254-5	(5)	7.16	7.16	7.09	7.23	0.00
TCX		3.63	3.64	3.59	3.69	0.01
DCB		9.52	9.53	9.43	9.63	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P5288 SAS No.: P5288 SDG NO.: P5288

Continuing Calib Date: 12/17/2024 Initial Calibration Date(s): 11/21/2024 11/22/2024

Continuing Calib Time: 16:55 Initial Calibration Time(s): 16:13 01:47

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	4.92	4.93	4.86	5.00	0.01
Aroclor-1254-2	(2)	5.08	5.09	5.02	5.16	0.01
Aroclor-1254-3	(3)	5.51	5.52	5.45	5.59	0.01
Aroclor-1254-4	(4)	5.75	5.76	5.69	5.83	0.01
Aroclor-1254-5	(5)	6.20	6.21	6.14	6.28	0.01
TCX		2.97	2.97	2.92	3.02	0.01
DCB		8.24	8.25	8.15	8.35	0.01