

Instrument ID: ECD_R

Daily Analysis Runlog For Sequence/QC Batch ID # PR112124

Review By	yogesh	Review On	11/22/2024 8:46:45 AM
Supervise By	Ankita	Supervise On	11/22/2024 1:20:52 PM
SubDirectory	PR112124	HP Acquire Method	HP Processing Method PR112124CLP

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24020,PP24019,PP24018,PP24017,PP24016,PP24025,PP24024,PP24023,PP24022,PP24021,PP24030,PP24029,PP24028,PP24027,PP24026,P P24035,PP24034,PP24033,PP24032,PP24031,PP24040,PP24039,PP24038,PP24037,PP24036,PP24045,PP24044,PP24043,PP24042,PP24041,PP 24050,PP24049,PP24048,PP24047,PP24046,PP24055,PP24054,PP24053,PP24052,PP24051 PP24018,PP24023,PP24028,PP24033,PP24038,PP24043,PP24048,PP24053

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PR069291.D	21 Nov 2024 15:44		AJMA	Ok
2	AIBLK081	AIBLK103	PR069292.D	21 Nov 2024 15:59		AJMA	Ok
3	AR1660ICC100	AR16601244	PR069293.D	21 Nov 2024 16:13		AJMA	Ok,M
4	AR1660ICC200	AR16602245	PR069294.D	21 Nov 2024 16:28		AJMA	Ok
5	AR1660ICC400	AR16603246	PR069295.D	21 Nov 2024 16:43		AJMA	Ok
6	AR1660ICC800	AR16604247	PR069296.D	21 Nov 2024 16:57		AJMA	Ok
7	AR1660ICC1600	AR16605248	PR069297.D	21 Nov 2024 17:12		AJMA	Ok
8	AR1221ICC100	AR12211249	PR069298.D	21 Nov 2024 17:27		AJMA	Ok,M
9	AR1221ICC200	AR12212250	PR069299.D	21 Nov 2024 17:41		AJMA	Ok,M
10	AR1221ICC400	AR12213201	PR069300.D	21 Nov 2024 17:56		AJMA	Ok,M
11	AR1221ICC800	AR12214202	PR069301.D	21 Nov 2024 18:11		AJMA	Ok
12	AR1221ICC1600	AR12215203	PR069302.D	21 Nov 2024 18:25		AJMA	Ok
13	AR1232ICC100	AR12321204	PR069303.D	21 Nov 2024 18:40		AJMA	Ok
14	AR1232ICC200	AR12322205	PR069304.D	21 Nov 2024 18:55		AJMA	Ok
15	AR1232ICC400	AR12323206	PR069305.D	21 Nov 2024 19:09		AJMA	Ok
16	AR1232ICC800	AR12324207	PR069306.D	21 Nov 2024 19:24		AJMA	Ok
17	AR1232ICC1600	AR12325208	PR069307.D	21 Nov 2024 19:39		AJMA	Ok
18	AR1242ICC100	AR12421209	PR069308.D	21 Nov 2024 19:54		AJMA	Ok

Instrument ID: ECD_R

Daily Analysis Runlog For Sequence/QC Batch ID # PR112124

Review By	yogesh	Review On	11/22/2024 8:46:45 AM
Supervise By	Ankita	Supervise On	11/22/2024 1:20:52 PM
SubDirectory	PR112124	HP Acquire Method	HP Processing Method PR112124CLP
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24020,PP24019,PP24018,PP24017,PP24016,PP24025,PP24024,PP24023,PP24022,PP24021,PP24030,PP24029,PP24028,PP24027,PP24026,PP24035,PP24034,PP24033,PP24032,PP24031,PP24040,PP24039,PP24038,PP24037,PP24036,PP24045,PP24044,PP24043,PP24042,PP24041,PP24050,PP24049,PP24048,PP24047,PP24046,PP24055,PP24054,PP24053,PP24052,PP24051 PP24018,PP24023,PP24028,PP24033,PP24038,PP24043,PP24048,PP24053		

19	AR1242ICC200	AR12422210	PR069309.D	21 Nov 2024 20:08		AJMA	Ok
20	AR1242ICC400	AR12423211	PR069310.D	21 Nov 2024 20:23		AJMA	Ok
21	AR1242ICC800	AR12424212	PR069311.D	21 Nov 2024 20:38		AJMA	Ok
22	AR1242ICC1600	AR12425213	PR069312.D	21 Nov 2024 20:52		AJMA	Ok
23	AR1248ICC100	AR12481214	PR069313.D	21 Nov 2024 21:07		AJMA	Ok,M
24	AR1248ICC200	AR12482215	PR069314.D	21 Nov 2024 21:22		AJMA	Ok
25	AR1248ICC400	AR12483216	PR069315.D	21 Nov 2024 21:37		AJMA	Ok
26	AR1248ICC800	AR12484217	PR069316.D	21 Nov 2024 21:51		AJMA	Ok
27	AR1248ICC1600	AR12485218	PR069317.D	21 Nov 2024 22:06		AJMA	Ok
28	AR1254ICC100	AR12541219	PR069318.D	21 Nov 2024 22:21		AJMA	Ok
29	AR1254ICC200	AR12542220	PR069319.D	21 Nov 2024 22:35		AJMA	Ok
30	AR1254ICC400	AR12543221	PR069320.D	21 Nov 2024 22:50		AJMA	Ok
31	AR1254ICC800	AR12544222	PR069321.D	21 Nov 2024 23:05		AJMA	Ok
32	AR1254ICC1600	AR12545223	PR069322.D	21 Nov 2024 23:20		AJMA	Ok
33	AR1262ICC100	AR12621224	PR069323.D	21 Nov 2024 23:34		AJMA	Ok,M
34	AR1262ICC200	AR12622225	PR069324.D	21 Nov 2024 23:49		AJMA	Ok,M
35	AR1262ICC400	AR12623226	PR069325.D	22 Nov 2024 00:04		AJMA	Ok,M
36	AR1262ICC800	AR12624227	PR069326.D	22 Nov 2024 00:18		AJMA	Ok,M
37	AR1262ICC1600	AR12625228	PR069327.D	22 Nov 2024 00:33		AJMA	Ok,M

Instrument ID: ECD_R

Daily Analysis Runlog For Sequence/QC Batch ID # PR112124

Review By	yogesh	Review On	11/22/2024 8:46:45 AM
Supervise By	Ankita	Supervise On	11/22/2024 1:20:52 PM
SubDirectory	PR112124	HP Acquire Method	HP Processing Method PR112124CLP
STD. NAME	STD REF.#		
Tune/Reschk	PP24020,PP24019,PP24018,PP24017,PP24016,PP24025,PP24024,PP24023,PP24022,PP24021,PP24030,PP24029,PP24028,PP24027,PP24026,P P24035,PP24034,PP24033,PP24032,PP24031,PP24040,PP24039,PP24038,PP24037,PP24036,PP24045,PP24044,PP24043,PP24042,PP24041,PP 24050,PP24049,PP24048,PP24047,PP24046,PP24055,PP24054,PP24053,PP24052,PP24051		
Initial Calibration Stds			
CCC			
Internal Standard/PEM			
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard	PP24018,PP24023,PP24028,PP24033,PP24038,PP24043,PP24048,PP24053		

38	AR1268ICC100	AR12681229	PR069328.D	22 Nov 2024 00:48		AJ\MA	Ok,M
39	AR1268ICC200	AR12682230	PR069329.D	22 Nov 2024 01:02		AJ\MA	Ok,M
40	AR1268ICC400	AR12683231	PR069330.D	22 Nov 2024 01:17		AJ\MA	Ok,M
41	AR1268ICC800	AR12684232	PR069331.D	22 Nov 2024 01:32		AJ\MA	Ok,M
42	AR1268ICC1600	AR12685233	PR069332.D	22 Nov 2024 01:47		AJ\MA	Ok
43	AIBLK082	AIBLK104	PR069333.D	22 Nov 2024 02:01		AJ\MA	Ok

M : Manual Integration

Instrument ID: ECD_R

Daily Analysis Runlog For Sequence/QC Batch ID # PR121724

Review By	yogesh	Review On	12/18/2024 8:25:31 AM
Supervise By	Ankita	Supervise On	12/18/2024 11:28:15 AM
SubDirectory	PR121724	HP Acquire Method	HP Processing Method PR112124

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24020,PP24019,PP24018,PP24017,PP24016,PP24025,PP24024,PP24023,PP24022,PP24021,PP24030,PP24029,PP24028,PP24027,PP24026,PP24035,PP24034,PP24033,PP24032,PP24031,PP24040,PP24039,PP24038,PP24037,PP24036,PP24045,PP24044,PP24043,PP24042,PP24041,PP24050,PP24049,PP24048,PP24047,PP24046,PP24055,PP24054,PP24053,PP24052,PP24051 PP24018,PP24023,PP24028,PP24033,PP24038,PP24043,PP24048,PP24053

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PR069698.D	17 Dec 2024 09:05		AJMA	Ok
2	AIBLK062	AIBLK133	PR069699.D	17 Dec 2024 09:20		AJMA	Ok
3	AR1660CCC400	AR16603321	PR069700.D	17 Dec 2024 09:34		AJMA	Ok
4	P5232-14DL	E28S8DL	PR069701.D	17 Dec 2024 09:49	AR1260 hits,	AJMA	Ok
5	P5264-06	C0AS6	PR069702.D	17 Dec 2024 10:04	AR1242 Hit, AR1254 Hit	AJMA	Ok,M
6	P5264-03	C0AS2	PR069703.D	17 Dec 2024 10:19	AR1242 Hit, AR1254 Hit, AR1260 Hit, DCB high in both column	AJMA	Ok,M
7	P5264-01DL	C0AG1DL	PR069704.D	17 Dec 2024 10:33	AR1242 Hit, AR1254 Hit, AR1260 Hit, Need Dilution, DCB high in both column, TCMX high in 1st column	AJMA	Dilution
8	P5264-01DL2	C0AG1DL2	PR069705.D	17 Dec 2024 10:48	AR1242 Hit, AR1254 Hit, AR1260 Hit, Need further Dilution	AJMA	Not Ok
9	P5264-02DL	C0AN0DL	PR069706.D	17 Dec 2024 11:17	AR1242 Hit, AR1254 Hit, AR1260 Hit, Need Dilution, DCB high in both column, TCMX high in 1st column	AJMA	Not Ok
10	P5264-02DL2	C0AN0DL2	PR069707.D	17 Dec 2024 11:47	AR1242 Hit, AR1254 Hit, AR1260 Hit, Need Further Dilution	AJMA	Not Ok
11	P5264-04DL	C0AS4DL	PR069708.D	17 Dec 2024 12:16	AR1242 Hit, AR1254 Hit, AR1260 Hit, Need Dilution, DCB high in both column, TCMX high in 1st column	AJMA	Not Ok
12	P5264-04DL2	C0AS4DL2	PR069709.D	17 Dec 2024 12:45	AR1242 Hit, AR1254 Hit, AR1260 Hit, Need Dilution,	AJMA	Not Ok

Instrument ID: ECD_R

Daily Analysis Runlog For Sequence/QC Batch ID # PR121724

Review By	yogesh	Review On	12/18/2024 8:25:31 AM
Supervise By	Ankita	Supervise On	12/18/2024 11:28:15 AM
SubDirectory	PR121724	HP Acquire Method	HP Processing Method PR112124
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24020,PP24019,PP24018,PP24017,PP24016,PP24025,PP24024,PP24023,PP24022,PP24021,PP24030,PP24029,PP24028,PP24027,PP24026,PP24035,PP24034,PP24033,PP24032,PP24031,PP24040,PP24039,PP24038,PP24037,PP24036,PP24045,PP24044,PP24043,PP24042,PP24041,PP24050,PP24049,PP24048,PP24047,PP24046,PP24055,PP24054,PP24053,PP24052,PP24051 PP24018,PP24023,PP24028,PP24033,PP24038,PP24043,PP24048,PP24053		

13	P5264-01DL2	C0AG1DL2	PR069710.D	17 Dec 2024 13:15	AR1242 Hit, AR1254 Hit, AR1260 Hit, Need further dilution	AJMA	Ok,M
14	P5264-02DL	C0AN0DL	PR069711.D	17 Dec 2024 13:44	AR1242 Hit, AR1254 Hit, AR1260 Hit, Need Dilution, DCB high in both column, TCMX high in 1st column	AJMA	Dilution
15	P5264-02DL2	C0AN0DL2	PR069712.D	17 Dec 2024 14:13	AR1242 Hit, AR1254 Hit, AR1260 Hit,	AJMA	Ok,M
16	P5264-04DL	C0AS4DL	PR069713.D	17 Dec 2024 14:43	AR1242 Hit, AR1254 Hit, AR1260 Hit, Need Dilution, DCB high in both column, TCMX high in 1st column	AJMA	Dilution
17	P5264-04DL2	C0AS4DL2	PR069714.D	17 Dec 2024 15:12	AR1242 Hit, AR1254 Hit, AR1260 Hit,	AJMA	Ok,M
18	P5288-13	PCB-GPC2-BLANK	PR069715.D	17 Dec 2024 15:27		AJMA	Ok
19	P5288-14	PCB-GPC2-BLANK-SP	PR069716.D	17 Dec 2024 15:42		AJMA	Ok
20	AIBLK063	AIBLK134	PR069717.D	17 Dec 2024 15:56		AJMA	Ok
21	AR1660CCC400	AR16603322	PR069718.D	17 Dec 2024 16:11		AJMA	Ok,M
22	AR1242CCC400	AR12423323	PR069719.D	17 Dec 2024 16:26		AJMA	Ok,M
23	AR1248CCC400	AR12483324	PR069720.D	17 Dec 2024 16:40		AJMA	Ok
24	AR1254CCC400	AR12543325	PR069721.D	17 Dec 2024 16:55		AJMA	Ok,M
25	PB165700BL	ABLK700	PR069722.D	17 Dec 2024 17:10	Will be reanalyzed with cleanup sample	AJMA	Not Ok
26	PB165700BS	ALCS700	PR069723.D	17 Dec 2024 17:24		AJMA	Ok
27	P5265-01	E28Y6	PR069724.D	17 Dec 2024 17:39		AJMA	Ok
28	P5265-02	E28Y7	PR069725.D	17 Dec 2024 17:54	AR1260 Hlt	AJMA	Ok

Instrument ID: ECD_R

Daily Analysis Runlog For Sequence/QC Batch ID # PR121724

Review By	yogesh	Review On	12/18/2024 8:25:31 AM
Supervise By	Ankita	Supervise On	12/18/2024 11:28:15 AM
SubDirectory	PR121724	HP Acquire Method	HP Processing Method PR112124
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24020,PP24019,PP24018,PP24017,PP24016,PP24025,PP24024,PP24023,PP24022,PP24021,PP24030,PP24029,PP24028,PP24027,PP24026,P P24035,PP24034,PP24033,PP24032,PP24031,PP24040,PP24039,PP24038,PP24037,PP24036,PP24045,PP24044,PP24043,PP24042,PP24041,PP 24050,PP24049,PP24048,PP24047,PP24046,PP24055,PP24054,PP24053,PP24052,PP24051 PP24018,PP24023,PP24028,PP24033,PP24038,PP24043,PP24048,PP24053		

29	P5265-03	E28Y8	PR069726.D	17 Dec 2024 18:08	AR1260 Hlt	AJMA	Ok,M
30	P5265-04	E28Y9	PR069727.D	17 Dec 2024 18:23	AR1260 Hlt, DCB low in both column	AJMA	Ok,M
31	P5265-05	E28Z0	PR069728.D	17 Dec 2024 18:38	AR1260 Hlt, DCB low in both column	AJMA	Ok,M
32	P5265-06	E28Z1	PR069729.D	17 Dec 2024 18:52	AR1242 Hit, AR1254 Hit, AR1260 Hlt, DCB low in both column	AJMA	Ok,M
33	P5265-07	E28Z2	PR069730.D	17 Dec 2024 19:07	AR1254 Hit, AR1260 Hit, Need dilution	AJMA	Dilution
34	P5265-08	E28Z3	PR069731.D	17 Dec 2024 19:22	AR1242 Hit, AR1254 Hit, AR1260 Hlt, DCB low in 2nd column	AJMA	Ok,M
35	P5265-09	E28Z4	PR069732.D	17 Dec 2024 19:36	AR1260 Hlt, DCB low in 2nd column	AJMA	Ok
36	P5265-10	E28Z5	PR069733.D	17 Dec 2024 19:51	AR1260 Hlt, DCB low in 2nd column	AJMA	Ok
37	P5265-11	E28Z6	PR069734.D	17 Dec 2024 20:06	AR1260 Hit	AJMA	Ok
38	P5265-12	E28Z7	PR069735.D	17 Dec 2024 20:20	AR1260 Hit	AJMA	Ok,M
39	P5265-13	E28Z8	PR069736.D	17 Dec 2024 20:35	need Cleanup	AJMA	Not Ok
40	P5265-14	E28Z9	PR069737.D	17 Dec 2024 20:50	need Cleanup	AJMA	Not Ok
41	P5265-15	E2900	PR069738.D	17 Dec 2024 21:04	AR1242 Hit, AR1254 Hit, AR1260 Hlt	AJMA	Ok,M
42	P5265-16	E2901	PR069739.D	17 Dec 2024 21:19		AJMA	Ok
43	P5265-17	E2902	PR069740.D	17 Dec 2024 21:34	AR1260 Hlt	AJMA	Ok
44	P5265-18	E2903	PR069741.D	17 Dec 2024 21:48	AR1242 Hit, AR1254 Hit, AR1260 Hit, Need dilution	AJMA	Dilution

Instrument ID: ECD_R

Daily Analysis Runlog For Sequence/QC Batch ID # PR121724

Review By	yogesh	Review On	12/18/2024 8:25:31 AM
Supervise By	Ankita	Supervise On	12/18/2024 11:28:15 AM
SubDirectory	PR121724	HP Acquire Method	HP Processing Method PR112124
STD. NAME	STD REF.#		
Tune/Reschk	PP24020,PP24019,PP24018,PP24017,PP24016,PP24025,PP24024,PP24023,PP24022,PP24021,PP24030,PP24029,PP24028,PP24027,PP24026,P P24035,PP24034,PP24033,PP24032,PP24031,PP24040,PP24039,PP24038,PP24037,PP24036,PP24045,PP24044,PP24043,PP24042,PP24041,PP 24050,PP24049,PP24048,PP24047,PP24046,PP24055,PP24054,PP24053,PP24052,PP24051 PP24018,PP24023,PP24028,PP24033,PP24038,PP24043,PP24048,PP24053		
Initial Calibration Stds			
CCC			
Internal Standard/PEM			
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

45	P5265-20MS	E28Y9MS	PR069742.D	17 Dec 2024 22:03	contamination from previous sample	AJ\MA	Not Ok
46	P5265-21MSD	E28Y9MSD	PR069743.D	17 Dec 2024 22:18	contamination from previous sample	AJ\MA	Not Ok
47	AIBLK064	AIBLK135	PR069744.D	17 Dec 2024 22:32		AJ\MA	Ok
48	AR1660CCC400	AR16603326	PR069745.D	17 Dec 2024 22:47		AJ\MA	Ok
49	AR1242CCC400	AR12423327	PR069746.D	17 Dec 2024 23:02		AJ\MA	Ok
50	AR1248CCC400	AR12483328	PR069747.D	17 Dec 2024 23:16		AJ\MA	Ok
51	AR1254CCC400	AR12543329	PR069748.D	17 Dec 2024 23:31		AJ\MA	Ok

M : Manual Integration