

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P5288 SAS No.: P5288 SDG NO.: P5288

EPA Sample No.: SSTD020573 Date Analyzed: 12/19/2024

Lab File ID: BG063745.D Time Analyzed: 04:11

Instrument ID: BNA_G GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	117864	7.802	491816	10.59	379534	14.44
UPPER LIMIT	235728	8.302	983632	11.087	759068	14.942
LOWER LIMIT	58932	7.302	245908	10.087	189767	13.942
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	85945	7.80	341897	10.59	278695	14.44

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.