

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P5288 SAS No.: P5288 SDG NO.: P5288

EPA Sample No.: SSTD020573 Date Analyzed: 12/19/2024

Lab File ID: BG063745.D Time Analyzed: 04:11

Instrument ID: BNA_G GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	977646	17.191	1035740	21.439	1083380	24.448
UPPER LIMIT	1955290	17.691	2071480	21.939	2166760	24.948
LOWER LIMIT	488823	16.691	517870	20.939	541690	23.948
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	763382	17.20	777954	21.44	838756	24.45

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.