

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CHEM02
 Lab Code: CHEM Case No.: P5288 SAS No.: P5288 SDG NO.: P5288
 Lab File ID: VX044308.D Date Analyzed: 12/16/2024
 Instrument ID: MSVOA_X Time Analyzed: 08:58
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	95969	12.018				
UPPER LIMIT	191938	12.518				
LOWER LIMIT	47984.5	11.518				
EPA SAMPLE NO.						
Storage-Blank-SOIL-REF-6	72853	12.02				
Storage-Blank-WATER-REF-5	81818	12.02				
Storage-Blank-WATER-REF-4	86519	12.02				
Storage-Blank-SAMPLE-MANAGEMENT	79808	12.02				
VX1216WBL01	94759	12.02				
VX1216WBS01	96021	12.02				
VX1216WBSD01	96145	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.