

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: CHEM02
Lab Code: CHEM Case No.: P5288 SAS No.: P5288 SDG NO.: P5288
Lab File ID: VX044218.D BFB Injection Date: 12/11/2024
Instrument ID: MSVOA_X BFB Injection Time: 10:13
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	21.7
75	30.0 - 60.0% of mass 95	56.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.7
173	Less than 2.0% of mass 174	0.9 (1.2) 1
174	50.0 - 100.0% of mass 95	71.1
175	5.0 - 9.0% of mass 174	5.2 (7.3) 1
176	95.0 - 101.0% of mass 174	68.7 (96.6) 1
177	5.0 - 9.0% of mass 176	4.6 (6.8) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC001	VSTDIC001	VX044219.D	12/11/2024	10:41
VSTDIC005	VSTDIC005	VX044220.D	12/11/2024	11:27
VSTDIC020	VSTDIC020	VX044221.D	12/11/2024	11:50
VSTDIC050	VSTDIC050	VX044222.D	12/11/2024	12:14
VSTDIC100	VSTDIC100	VX044223.D	12/11/2024	12:37
VSTDIC150	VSTDIC150	VX044224.D	12/11/2024	13:00