

Laboratory Control Sample Summary

Client:	Weston Solutions	SDG No.:	P5290
Project:	Joint Base McGuire Dix Lakehurst JBMDL - DoD	Run No.:	LB134031

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134031BSS							
TOC	mg/Kg	1000	978		98	1	90-110	12/20/2024

Laboratory Control Sample Summary

Client:	Weston Solutions	SDG No.:	P5290
Project:	Joint Base McGuire Dix Lakehurst JBMDL - DoD	Run No.:	LB134031

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134031BSS2							
TOC	mg/Kg	1000	965		96	1	90-110	12/23/2024