

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB133954

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3157

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	81.7	12/16/2024	10:00
2	P5286-01		1	11	100.00	>212.0	>212.0	12/16/2024	10:30
3	P5286-01DUP		1	11	100.00	>212.0	>212.0	12/16/2024	11:00
4	P5287-05		1	13	100.00	>212.0	>212.0	12/16/2024	11:30
5	P5291-13		1	14	100.00	>212.0	>212.0	12/16/2024	12:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6133954

WorkList Name : fp-12-16 WorkList ID : 186349 Department : Wet-Chemistry Date : 12-16-2024 08:32:07

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5286-01	120424-COMP	Water	Flash Point	Cool 4 deg C	PSEG03	L41	12/13/2024	1010B
P5287-05	220-2-TRANS-WATER	Water	Flash Point	Cool 4 deg C	PSEG03	L61	12/13/2024	1010B
P5291-13	WC-20241213	Water	Flash Point	Cool 4 deg C	PARS02	L51	12/13/2024	1010B

Date/Time 12/16/2024 09:50
Raw Sample Received by: R14 CWS
Raw Sample Relinquished by: R14 CWS

Date/Time 12/16/2024 12:40
Raw Sample Received by: R14 CWS
Raw Sample Relinquished by: R14 CWS