


Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: PARSONS Main of New York, Inc. **SDG No.:** P5291
Contract: PARS02 **Lab Code:** CHEM **Case No.:** P5291 **SAS No.:** P5291

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165720BS							
Arsenic	ug/L	4000	3970		99	80 - 120	P
Barium	ug/L	1000	880		88	80 - 120	P
Cadmium	ug/L	1000	954		95	80 - 120	P
Chromium	ug/L	2000	2040		102	80 - 120	P
Lead	ug/L	5000	4790		96	80 - 120	P
Selenium	ug/L	10000	9680		97	80 - 120	P
Silver	ug/L	380	371		98	80 - 120	P



- 7 -

Client:	<u>PARSONS Main of New York, Inc.</u>	SDG No.:	<u>P5291</u>				
Contract:	<u>PARS02</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>P5291</u>	SAS No.:	<u>P5291</u>

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165736BS Mercury	ug/L	4.0	4.20		105	80 - 120	CV