

### CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: P5291 SAS No.: P5291 SDG NO.: P5291

Continuing Calib Date: 12/16/2024 Initial Calibration Date(s): 12/06/2024 12/06/2024

Continuing Calib Time: 21:12 Initial Calibration Time(s): 14:19 22:34

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1242-1 (1)   | 4.80       | 4.81      | 4.71      | 4.91 | 0.01       |
| Aroclor-1242-2 (2)   | 4.82       | 4.83      | 4.73      | 4.93 | 0.01       |
| Aroclor-1242-3 (3)   | 4.88       | 4.88      | 4.78      | 4.98 | 0.00       |
| Aroclor-1242-4 (4)   | 5.00       | 5.01      | 4.91      | 5.11 | 0.01       |
| Aroclor-1242-5 (5)   | 5.66       | 5.66      | 5.56      | 5.76 | 0.00       |
| Tetrachloro-m-xylene | 3.71       | 3.71      | 3.61      | 3.81 | 0.00       |
| Decachlorobiphenyl   | 8.78       | 8.79      | 8.69      | 8.89 | 0.01       |

### CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: P5291 SAS No.: P5291 SDG NO.: P5291

Continuing Calib Date: 12/16/2024 Initial Calibration Date(s): 12/06/2024 12/06/2024

Continuing Calib Time: 21:12 Initial Calibration Time(s): 14:19 22:34

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1242-1 (1)   | 4.79       | 4.80      | 4.70      | 4.90 | 0.01       |
| Aroclor-1242-2 (2)   | 4.81       | 4.82      | 4.72      | 4.92 | 0.01       |
| Aroclor-1242-3 (3)   | 4.99       | 4.99      | 4.89      | 5.09 | 0.00       |
| Aroclor-1242-4 (4)   | 5.07       | 5.08      | 4.98      | 5.18 | 0.01       |
| Aroclor-1242-5 (5)   | 5.59       | 5.60      | 5.50      | 5.70 | 0.01       |
| Tetrachloro-m-xylene | 3.70       | 3.71      | 3.61      | 3.81 | 0.01       |
| Decachlorobiphenyl   | 8.73       | 8.74      | 8.64      | 8.84 | 0.01       |