

CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: P5291 SAS No.: P5291 SDG NO.: P5291

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL02 Date Analyzed: 12/16/2024

Lab Sample No.: AR1660CCC500 Data File : PO108568.D Time Analyzed: 20:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.804	4.709	4.909	475.410	500.000	-4.9
Aroclor-1016-2	4.824	4.729	4.929	477.160	500.000	-4.6
Aroclor-1016-3	4.879	4.785	4.985	478.370	500.000	-4.3
Aroclor-1016-4	5.001	4.907	5.107	478.960	500.000	-4.2
Aroclor-1016-5	5.259	5.165	5.365	481.360	500.000	-3.7
Aroclor-1260-1	6.301	6.210	6.410	473.430	500.000	-5.3
Aroclor-1260-2	6.489	6.398	6.598	471.620	500.000	-5.7
Aroclor-1260-3	6.859	6.769	6.969	472.370	500.000	-5.5
Aroclor-1260-4	7.119	7.029	7.229	463.630	500.000	-7.3
Aroclor-1260-5	7.360	7.270	7.470	470.330	500.000	-5.9
Decachlorobiphenyl	8.775	8.691	8.891	44.000	50.000	-12.0
Tetrachloro-m-xylene	3.707	3.610	3.810	47.810	50.000	-4.4

CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: P5291 SAS No.: P5291 SDG NO.: P5291

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL02 Date Analyzed: 12/16/2024

Lab Sample No.: AR1660CCC500 Data File : PO108568.D Time Analyzed: 20:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.792	4.698	4.898	535.410	500.000	7.1
Aroclor-1016-2	4.811	4.718	4.918	532.240	500.000	6.4
Aroclor-1016-3	4.987	4.894	5.094	526.070	500.000	5.2
Aroclor-1016-4	5.028	4.935	5.135	519.220	500.000	3.8
Aroclor-1016-5	5.242	5.150	5.350	543.930	500.000	8.8
Aroclor-1260-1	6.277	6.186	6.386	542.650	500.000	8.5
Aroclor-1260-2	6.463	6.373	6.573	543.530	500.000	8.7
Aroclor-1260-3	6.617	6.527	6.727	543.470	500.000	8.7
Aroclor-1260-4	7.089	7.000	7.200	563.980	500.000	12.8
Aroclor-1260-5	7.328	7.239	7.439	562.740	500.000	12.5
Decachlorobiphenyl	8.724	8.641	8.841	52.510	50.000	5.0
Tetrachloro-m-xylene	3.705	3.608	3.808	52.990	50.000	6.0